

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG081919\
 Data File : BG042348.D
 Acq On : 20 Aug 2019 6:07
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 21 03:19:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG081919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 13:40:37 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.426	0.414	2.8	130	0.00
3	Pyridine	1.090	1.108	-1.7	126	0.00
4	n-Nitrosodimethylamine	0.375	0.386	-2.9	130	0.00
5 S	2-Fluorophenol	0.987	1.042	-5.6	131	0.00
6	Aniline	1.769	1.729	2.3	125	0.00
7 S	Phenol-d6	1.341	1.428	-6.5	129	0.00
8	2-Chlorophenol	1.200	1.262	-5.2	130	0.00
9	Benzaldehyde	0.754	0.764	-1.3	132	0.00
10 C	Phenol	1.363	1.450	-6.4	130	0.00
11	bis(2-Chloroethyl)ether	1.077	1.087	-0.9	126	0.00
12	1,3-Dichlorobenzene	1.527	1.530	-0.2	126	0.00
13 C	1,4-Dichlorobenzene	1.538	1.535	0.2	128	0.00
14	1,2-Dichlorobenzene	1.468	1.459	0.6	127	0.00
15	Benzyl Alcohol	0.922	0.955	-3.6	128	0.00
16	2,2'-oxybis(1-Chloropropane	1.485	1.476	0.6	125	0.00
17	2-Methylphenol	1.008	1.015	-0.7	127	0.00
18	Hexachloroethane	0.529	0.513	3.0	124	0.00
19 P	n-Nitroso-di-n-propylamine	0.862	0.865	-0.3	122	0.00
20	3+4-Methylphenols	1.394	1.428	-2.4	126	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
22	Acetophenone	0.425	0.435	-2.4	122	0.00
23 S	Nitrobenzene-d5	0.284	0.296	-4.2	125	0.00
24	Nitrobenzene	0.290	0.306	-5.5	125	0.00
25	Isophorone	0.569	0.580	-1.9	120	0.00
26 C	2-Nitrophenol	0.180	0.194	-7.8	124	0.00
27	2,4-Dimethylphenol	0.250	0.262	-4.8	126	0.00
28	bis(2-Chloroethoxy)methane	0.366	0.376	-2.7	123	0.00
29 C	2,4-Dichlorophenol	0.332	0.350	-5.4	124	0.00
30	1,2,4-Trichlorobenzene	0.406	0.408	-0.5	121	0.00
31	Naphthalene	0.927	0.969	-4.5	123	0.00
32	Benzoic acid	0.166	0.170	-2.4	124	0.01
33	4-Chloroaniline	0.423	0.426	-0.7	122	0.00
34 C	Hexachlorobutadiene	0.272	0.268	1.5	121	0.00
35	Caprolactam	0.106	0.113	-6.6	123	0.00
36 C	4-Chloro-3-methylphenol	0.311	0.330	-6.1	122	0.00
37	2-Methylnaphthalene	0.743	0.764	-2.8	121	0.00
38	1-Methylnaphthalene	0.708	0.730	-3.1	120	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	0.645	0.661	-2.5	120	0.00
41 P	Hexachlorocyclopentadiene	0.326	0.247	24.2	93	0.00
42 S	2,4,6-Tribromophenol	0.272	0.274	-0.7	114	0.00
43 C	2,4,6-Trichlorophenol	0.431	0.429	0.5	117	0.00
44	2,4,5-Trichlorophenol	0.444	0.467	-5.2	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG081919\
 Data File : BG042348.D
 Acq On : 20 Aug 2019 6:07
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 21 03:19:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG081919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 13:40:37 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.182	1.222	-3.4	117	0.00
46	1,1'-Biphenyl	1.305	1.356	-3.9	119	0.00
47	2-Chloronaphthalene	1.112	1.153	-3.7	120	0.00
48	2-Nitroaniline	0.263	0.276	-4.9	122	0.00
49	Acenaphthylene	1.672	1.768	-5.7	118	0.00
50	Dimethylphthalate	1.421	1.455	-2.4	116	0.00
51	2,6-Dinitrotoluene	0.333	0.348	-4.5	118	0.00
52 C	Acenaphthene	1.020	1.060	-3.9	118	0.00
53	3-Nitroaniline	0.327	0.335	-2.4	119	0.00
54 P	2,4-Dinitrophenol	0.185	0.082	55.7#	50#	0.00
55	Dibenzofuran	1.662	1.751	-5.4	118	0.00
56 P	4-Nitrophenol	0.236	0.245	-3.8	113	0.00
57	2,4-Dinitrotoluene	0.465	0.491	-5.6	117	0.00
58	Fluorene	1.352	1.411	-4.4	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.432	0.422	2.3	116	0.00
60	Diethylphthalate	1.375	1.412	-2.7	117	0.00
61	4-Chlorophenyl-phenylether	0.812	0.832	-2.5	116	0.00
62	4-Nitroaniline	0.346	0.365	-5.5	119	0.00
63	Azobenzene	0.958	0.999	-4.3	118	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.085	30.3#	77	0.00
66 c	n-Nitrosodiphenylamine	0.560	0.587	-4.8	118	0.00
67	4-Bromophenyl-phenylether	0.255	0.261	-2.4	117	0.00
68	Hexachlorobenzene	0.273	0.277	-1.5	116	0.00
69	Atrazine	0.192	0.156	18.8	121	0.00
70 C	Pentachlorophenol	0.145	0.134	7.6	106	0.00
71	Phenanthrene	0.996	1.045	-4.9	116	0.00
72	Anthracene	0.993	1.040	-4.7	116	0.00
73	Carbazole	0.882	0.945	-7.1	119	0.00
74	Di-n-butylphthalate	1.029	1.084	-5.3	117	0.00
75 C	Fluoranthene	1.201	1.266	-5.4	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
77	Benzidine	0.530	0.535	-0.9	107	0.00
78	Pyrene	1.231	1.288	-4.6	116	0.00
79 S	Terphenyl-d14	0.913	0.911	0.2	115	0.00
80	Butylbenzylphthalate	0.488	0.512	-4.9	120	0.00
81	Benzo(a)anthracene	1.228	1.287	-4.8	118	0.00
82	3,3'-Dichlorobenzidine	0.489	0.491	-0.4	115	0.00
83	Chrysene	1.175	1.236	-5.2	117	0.00
84	Bis(2-ethylhexyl)phthalate	0.680	0.719	-5.7	120	0.00
85 c	Di-n-octyl phthalate	1.160	1.233	-6.3	120	0.00
86	Indeno(1,2,3-cd)pyrene	1.569	1.648	-5.0	120	0.00
87 I	Perylene-d12	1.000	1.000	0.0	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG081919\
Data File : BG042348.D
Acq On : 20 Aug 2019 6:07
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 21 03:19:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG081919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 13:40:37 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.099	1.144	-4.1	119	0.00
89	Benzo(k)fluoranthene	1.075	1.111	-3.3	118	0.00
90 C	Benzo(a)pyrene	1.053	1.094	-3.9	120	0.00
91	Dibenzo(a,h)anthracene	1.051	1.088	-3.5	119	0.00
92	Benzo(g,h,i)perylene	1.047	1.100	-5.1	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0