

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090320\
 Data File : BG046512.D
 Acq On : 3 Sep 2020 10:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 03 11:19:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 15:06:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.470	0.449	4.5	122	0.00
3	Pyridine	1.375	1.288	6.3	115	0.00
4	n-Nitrosodimethylamine	0.507	0.491	3.2	115	0.00
5 S	2-Fluorophenol	1.129	1.037	8.1	115	0.00
6	Aniline	1.801	1.631	9.4	111	0.00
7 S	Phenol-d6	1.601	1.459	8.9	112	0.00
8	2-Chlorophenol	1.193	1.074	10.0	114	0.00
9	Benzaldehyde	0.897	0.824	8.1	114	0.00
10 C	Phenol	1.474	1.353	8.2	113	0.00
11	bis(2-Chloroethyl)ether	1.185	1.069	9.8	113	0.00
12	1,3-Dichlorobenzene	1.515	1.384	8.6	115	0.00
13 C	1,4-Dichlorobenzene	1.517	1.378	9.2	115	0.00
14	1,2-Dichlorobenzene	1.455	1.302	10.5	114	0.00
15	Benzyl Alcohol	1.028	0.982	4.5	115	0.00
16	2,2'-oxybis(1-Chloropropane	1.838	1.739	5.4	117	0.00
17	2-Methylphenol	1.046	0.943	9.8	111	0.00
18	Hexachloroethane	0.514	0.482	6.2	118	0.00
19 P	n-Nitroso-di-n-propylamine	0.958	0.886	7.5	113	0.00
20	3+4-Methylphenols	1.443	1.320	8.5	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.487	0.445	8.6	111	0.00
23 S	Nitrobenzene-d5	0.350	0.327	6.6	113	0.00
24	Nitrobenzene	0.346	0.320	7.5	113	0.00
25	Isophorone	0.642	0.601	6.4	111	0.00
26 C	2-Nitrophenol	0.182	0.174	4.4	111	0.00
27	2,4-Dimethylphenol	0.251	0.234	6.8	111	0.00
28	bis(2-Chloroethoxy)methane	0.413	0.389	5.8	112	0.00
29 C	2,4-Dichlorophenol	0.334	0.311	6.9	110	0.00
30	1,2,4-Trichlorobenzene	0.397	0.360	9.3	112	0.00
31	Naphthalene	0.977	0.890	8.9	111	0.00
32	Benzoic acid	0.156	0.134	14.1	91	0.00
33	4-Chloroaniline	0.431	0.396	8.1	109	0.00
34 C	Hexachlorobutadiene	0.258	0.238	7.8	112	0.00
35	Caprolactam	0.125	0.106	15.2	99	0.00
36 C	4-Chloro-3-methylphenol	0.327	0.300	8.3	108	0.00
37	2-Methylnaphthalene	0.770	0.705	8.4	110	0.00
38	1-Methylnaphthalene	0.730	0.664	9.0	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.660	0.617	6.5	109	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.286	0.3	107	0.00
42 S	2,4,6-Tribromophenol	0.271	0.238	12.2	100	0.00
43 C	2,4,6-Trichlorophenol	0.445	0.418	6.1	105	0.00
44	2,4,5-Trichlorophenol	0.468	0.430	8.1	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090320\
 Data File : BG046512.D
 Acq On : 3 Sep 2020 10:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 03 11:19:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 15:06:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.310	1.197	8.6	108	0.00
46	1,1'-Biphenyl	1.391	1.293	7.0	108	0.00
47	2-Chloronaphthalene	1.180	1.085	8.1	107	0.00
48	2-Nitroaniline	0.328	0.311	5.2	107	0.00
49	Acenaphthylene	1.790	1.635	8.7	105	0.00
50	Dimethylphthalate	1.547	1.388	10.3	103	0.00
51	2,6-Dinitrotoluene	0.341	0.311	8.8	104	0.00
52 C	Acenaphthene	1.109	1.012	8.7	106	0.00
53	3-Nitroaniline	0.369	0.331	10.3	100	0.00
54 P	2,4-Dinitrophenol	0.199	0.185	7.0	96	0.00
55	Dibenzofuran	1.780	1.600	10.1	105	0.00
56 P	4-Nitrophenol	0.272	0.245	9.9	98	0.00
57	2,4-Dinitrotoluene	0.486	0.443	8.8	102	0.00
58	Fluorene	1.494	1.321	11.6	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.454	0.413	9.0	103	0.00
60	Diethylphthalate	1.509	1.338	11.3	103	0.00
61	4-Chlorophenyl-phenylether	0.823	0.738	10.3	103	0.00
62	4-Nitroaniline	0.390	0.341	12.6	99	0.00
63	Azobenzene	1.227	1.108	9.7	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.113	0.112	0.9	99	0.00
66 c	n-Nitrosodiphenylamine	0.540	0.510	5.6	102	0.00
67	4-Bromophenyl-phenylether	0.233	0.224	3.9	103	0.00
68	Hexachlorobenzene	0.258	0.240	7.0	100	0.00
69	Atrazine	0.220	0.202	8.2	97	0.00
70 C	Pentachlorophenol	0.135	0.132	2.2	96	0.00
71	Phenanthrene	1.014	0.921	9.2	100	0.00
72	Anthracene	1.000	0.914	8.6	100	0.00
73	Carbazole	0.944	0.849	10.1	98	0.00
74	Di-n-butylphthalate	1.086	1.001	7.8	100	0.00
75 C	Fluoranthene	1.305	1.143	12.4	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.465	0.376	19.1	74	0.00
78	Pyrene	1.274	1.171	8.1	97	0.00
79 S	Terphenyl-d14	0.940	0.833	11.4	97	0.00
80	Butylbenzylphthalate	0.497	0.474	4.6	98	0.00
81	Benzo(a)anthracene	1.247	1.138	8.7	98	0.00
82	3,3'-Dichlorobenzidine	0.463	0.436	5.8	98	0.00
83	Chrysene	1.199	1.096	8.6	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.678	0.648	4.4	100	0.00
85 c	Di-n-octyl phthalate	1.161	1.129	2.8	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.436	1.277	11.1	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090320\
Data File : BG046512.D
Acq On : 3 Sep 2020 10:15
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 03 11:19:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 02 15:06:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.249	1.087	13.0	100	0.00
89	Benzo(k)fluoranthene	1.207	1.067	11.6	101	0.00
90 C	Benzo(a)pyrene	1.165	1.039	10.8	102	0.00
91	Dibenzo(a,h)anthracene	1.159	1.026	11.5	101	0.00
92	Benzo(g,h,i)perylene	1.158	1.030	11.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0