


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG090320\
 Data File : BG046511.D
 Acq On : 3 Sep 2020 9:32
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG082520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 02 15:06:53 2020



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.0	16389	PASS
68	69	0.00	2	1.1	176	PASS
69	198	0.00	100	26.7	16192	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	37.1	22533	PASS
197	198	0.00	2	0.2	116	PASS
198	198	100	100	100.0	60757	PASS
199	198	5	9	6.9	4180	PASS
275	198	10	60	28.1	17073	PASS
365	198	1	100	3.6	2181	PASS
441	198	0.01	100	9.2	5581	PASS
442	198	50	100	56.1	34085	PASS
443	442	15	24	19.9	6779	PASS