

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
 Data File : BG051538.D
 Acq On : 15 Dec 2021 15:38
 Operator : CG/JU
 Sample : PB141276BL
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Quant Time: Dec 15 23:24:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 14 14:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.281	152	27890	20.000	ng/ul	0.00
20) Naphthalene-d8	11.119	136	113283	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.914	164	84080	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.657	188	214194	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.975	240	218090	20.000	ng/ul	#-0.01
88) Perylene-d12	25.470	264	210290	20.000	ng/ul	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.605	96	4343	6.467	ng/uL	0.00
4) Pyridine-d5	4.034	84	66101	32.577	ng/ul	-0.01
7) Phenol-d5	7.406	99	84090	33.768	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.582	67	50430	34.066	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.800	132	61645	35.593	ng/ul	-0.01
15) 4-Methylphenol-d8	8.969	113	68262	35.463	ng/ul	-0.01
21) Nitrobenzene-d5	9.450	128	30829	37.119	ng/ul	0.00
24) 2-Nitrophenol-d4	10.179	143	34566	38.071	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.719	165	71442	36.285	ng/ul	-0.01
31) 4-Chloroaniline-d4	11.236	131	89948	34.698	ng/ul	0.00
46) Dimethylphthalate-d6	14.303	166	244903	36.391	ng/ul	-0.01
49) Acenaphthylene-d8	14.608	160	293930	35.168	ng/ul	0.00
54) 4-Nitrophenol-d4	15.066	143	35958	32.857	ng/ul	-0.01
60) Fluorene-d10	15.901	176	212574	35.256	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.995	200	28869	25.869	ng/ul	-0.02
73) Anthracene-d10	17.757	188	372915	36.497	ng/ul	0.00
81) Pyrene-d10	20.030	212	455030	34.667	ng/ul	-0.01
92) Benzo(a)pyrene-d12	25.235	264	406081	36.701	ng/ul	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
Data File : BG051538.D
Acq On : 15 Dec 2021 15:38
Operator : CG/JU
Sample : PB141276BL
Misc :
ALS Vial : 36 Sample Multiplier: 1

Quant Time: Dec 15 23:24:28 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 14 14:19:57 2021
Response via : Initial Calibration

