

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121721\
 Data File : BG051604.D
 Acq On : 17 Dec 2021 20:23
 Operator : CG/JU
 Sample : PB141493BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK493

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/20/2021
 Supervised By :mohammad ahmed 12/24/2021

Quant Time: Dec 17 23:04:53 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 14 14:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.273	152	30104	20.000	ng/u1	-0.01
20) Naphthalene-d8	11.110	136	121658	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.911	164	90363	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.654	188	222775	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.972	240	226395	20.000	ng/u1	#-0.01
88) Perylene-d12	25.461	264	217517	20.000	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.603	96	4582m	6.321	ng/uL	0.00
4) Pyridine-d5	4.031	84	44123	20.146	ng/u1	-0.01
7) Phenol-d5	7.403	99	79902	29.727	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.574	67	49312	30.861	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.797	132	55428	29.650	ng/u1	-0.01
15) 4-Methylphenol-d8	8.966	113	62319	29.994	ng/u1	-0.01
21) Nitrobenzene-d5	9.442	128	30146	33.798	ng/u1	-0.01
24) 2-Nitrophenol-d4	10.176	143	31312	32.113	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.716	165	63063	29.825	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.233	131	54737	19.662	ng/u1	0.00
46) Dimethylphthalate-d6	14.300	166	223776	30.940	ng/u1	-0.01
49) Acenaphthylene-d8	14.605	160	269369	29.988	ng/u1	0.00
54) 4-Nitrophenol-d4	15.064	143	35223	29.947	ng/u1	-0.01
60) Fluorene-d10	15.898	176	193596	29.876	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.992	200	34095	29.375	ng/u1	-0.02
73) Anthracene-d10	17.754	188	337109	31.722	ng/u1	0.00
81) Pyrene-d10	20.027	212	415949	30.527	ng/u1	-0.01
92) Benzo(a)pyrene-d12	25.226	264	374864	32.754	ng/u1	-0.03

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121721\
Data File : BG051604.D
Acq On : 17 Dec 2021 20:23
Operator : CG/JU
Sample : PB141493BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK493

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/20/2021
Supervised By :mohammad ahmed 12/24/2021

Quant Time: Dec 17 23:04:53 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 14 14:19:57 2021
Response via : Initial Calibration

