

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
 Data File : BG060202.D
 Acq On : 2 Jan 2024 19:52
 Operator : MA/JU
 Sample : 05899-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GHL-SS-09

Quant Time: Jan 03 00:41:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 30 03:14:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.935	152	224846	20.000	ng	0.00
21) Naphthalene-d8	10.743	136	925125	20.000	ng	0.00
39) Acenaphthene-d10	14.574	164	692158	20.000	ng	0.00
64) Phenanthrene-d10	17.324	188	1700209	20.000	ng	0.00
76) Chrysene-d12	21.590	240	1597066	20.000	ng	0.00
86) Perylene-d12	24.762	264	1753371	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.508	112	1408193	94.529	ng	0.00
7) Phenol-d6	7.101	99	1848721	83.582	ng	0.00
23) Nitrobenzene-d5	9.098	82	1134790	56.658	ng	0.00
42) 2,4,6-Tribromophenol	16.067	330	759510	81.974	ng	0.00
45) 2-Fluorobiphenyl	13.199	172	2860296	59.392	ng	0.00
79) Terphenyl-d14	19.944	244	4120316	49.599	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
Data File : BG060202.D
Acq On : 2 Jan 2024 19:52
Operator : MA/JU
Sample : 05899-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GHL-SS-09

Quant Time: Jan 03 00:41:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 30 03:14:04 2023
Response via : Initial Calibration

