

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
 Data File : BG060211.D
 Acq On : 3 Jan 2024 3:16
 Operator : MA/JU
 Sample : 05899-24
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GHL-SS-19-2-4

Quant Time: Jan 03 05:06:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 30 03:14:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.935	152	253974	20.000	ng	0.00
21) Naphthalene-d8	10.737	136	1105198	20.000	ng	0.00
39) Acenaphthene-d10	14.574	164	878081	20.000	ng	0.00
64) Phenanthrene-d10	17.324	188	2186977	20.000	ng	0.00
76) Chrysene-d12	21.589	240	1796086	20.000	ng	0.00
86) Perylene-d12	24.768	264	2018012	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.508	112	1658082	98.538	ng	0.00
7) Phenol-d6	7.100	99	2322543	92.961	ng	0.00
23) Nitrobenzene-d5	9.098	82	1377671	57.577	ng	0.00
42) 2,4,6-Tribromophenol	16.060	330	664446	56.529	ng	0.00
45) 2-Fluorobiphenyl	13.199	172	3534410	57.850	ng	0.00
79) Terphenyl-d14	19.944	244	4948656	52.969	ng	0.00
Target Compounds						
						Qvalue
72) Anthracene	17.453	178	1280216	11.923	ng	96
75) Fluoranthene	19.380	202	596215	4.702	ng	91
77) Benzidine	19.562	184	378139	7.607	ng	98
78) Pyrene	19.738	202	638695	5.290	ng	92
81) Benzo(a)anthracene	21.572	228	354238	3.102	ng	94
83) Chrysene	21.636	228	508433	4.583	ng	94
88) Benzo(b)fluoranthene	23.751	252	376518	3.105	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
Data File : BG060211.D
Acq On : 3 Jan 2024 3:16
Operator : MA/JU
Sample : 05899-24
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GHL-SS-19-2-4

Quant Time: Jan 03 05:06:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 30 03:14:04 2023
Response via : Initial Calibration

