

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
 Data File : BG060219.D
 Acq On : 3 Jan 2024 8:40
 Operator : MA/JU
 Sample : 05909-13
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GHL-SS-51

Quant Time: Jan 03 14:41:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 30 03:14:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	220271	20.000	ng	0.00
21) Naphthalene-d8	10.740	136	1023101	20.000	ng	0.00
39) Acenaphthene-d10	14.576	164	825948	20.000	ng	0.00
64) Phenanthrene-d10	17.320	188	1989750	20.000	ng	0.00
76) Chrysene-d12	21.586	240	1609101	20.000	ng	0.00
86) Perylene-d12	24.759	264	1791469	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.511	112	1431013	98.056	ng	0.00
7) Phenol-d6	7.097	99	1996869	92.155	ng	0.00
23) Nitrobenzene-d5	9.101	82	1150888	51.959	ng	0.00
42) 2,4,6-Tribromophenol	16.063	330	887407	80.263	ng	0.00
45) 2-Fluorobiphenyl	13.196	172	2660749	46.299	ng	0.00
79) Terphenyl-d14	19.941	244	3137260	37.483	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010224\
Data File : BG060219.D
Acq On : 3 Jan 2024 8:40
Operator : MA/JU
Sample : 05909-13
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GHL-SS-51

Quant Time: Jan 03 14:41:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 30 03:14:04 2023
Response via : Initial Calibration

