

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG010318\
 Data File : BG031888.D
 Acq On : 3 Jan 2018 13:29
 Operator : SJ/JU
 Sample : I7102-07
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BISP-SB-05

Quant Time: Jan 03 15:31:55 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG122117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 29 15:51:46 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.79	152	41676	20.00	ng	0.00
21) Naphthalene-d8	11.67	136	179752	20.00	ng	0.00
38) Acenaphthene-d10	15.41	164	133610	20.00	ng	-0.01
63) Phenanthrene-d10	18.15	188	377138	20.00	ng	-0.01
75) Chrysene-d12	22.49	240	451718	20.00	ng	-0.02
86) Perylene-d12	26.22	264	457736	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	6.21	112	206982	90.46	ng	0.00
7) Phenol-d6	7.89	99	273030	91.91	ng	0.00
23) Nitrobenzene-d5	9.99	82	149629	62.86	ng	0.00
41) 2,4,6-Tribromophenol	16.89	330	195262	95.78	ng	0.00
44) 2-Fluorobiphenyl	14.05	172	513669	48.25	ng	0.00
78) Terphenyl-d14	20.68	244	1073110	54.27	ng	-0.01
Target Compounds						
10) Phenol	7.92	94	10571	3.524	ng	Qvalue 87
49) Dimethylphthalate	14.84	163	45541	3.869	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG010318\
Data File : BG031888.D
Acq On : 3 Jan 2018 13:29
Operator : SJ/JU
Sample : I7102-07
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BISP-SB-05

Quant Time: Jan 03 15:31:55 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG122117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 29 15:51:46 2017
Response via : Initial Calibration

