


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010319\
 Data File : BG038906.D
 Acq On : 3 Jan 2019 16:46
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 12 15:26:27 2018



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2501

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	16987	PASS
68	69	0.00	2	1.0	147	PASS
69	198	0.00	100	29.8	14912	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.0	20549	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	50093	PASS
199	198	5	9	6.8	3425	PASS
275	198	10	60	30.5	15282	PASS
365	198	1	100	4.1	2073	PASS
441	198	0.01	100	13.0	6515	PASS
442	198	50	100	86.0	43061	PASS
443	442	15	24	19.5	8389	PASS