


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010522\
 Data File : BG051874.D
 Acq On : 5 Jan 2022 10:11
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 07 06:40:51 2022






AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.2	24027	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.9	24907	PASS
70	69	0.00	2	0.4	89	PASS
127	198	10	80	48.9	25437	PASS
197	198	0.00	2	0.2	79	PASS
198	198	100	100	100.0	51984	PASS
199	198	5	9	6.3	3273	PASS
275	198	10	60	32.4	16818	PASS
365	198	1	100	3.7	1923	PASS
441	198	0.01	100	11.8	6135	PASS
442	442	50	100	100.0	38837	PASS
443	442	15	24	19.4	7535	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010522\
Data File : BG051874.D
Acq On : 5 Jan 2022 10:11
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 07 06:53:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 07 06:40:51 2022
Response via : Initial Calibration



TIC: BG051874.D\data.ms

(70) Pentachlorophenol (C)

17.288min (+ 0.001) 19727.97 ng

response 61769

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.20	60.74
264.00	65.20	61.03
0.00	0.00	0.00

Instrument :
BNA_G
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/5/2022	BNA_G	<u>BG051874.D</u>
Compound Name	Response	Retention Time
DDT	219787	21.1
DDD	1522	20.695
DDE	0	20.09
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1522	221309	0.69