

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010523\
 Data File : BG056215.D
 Acq On : 5 Jan 2023 16:22
 Operator : CG/JU
 Sample : PB149947BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK947

Quant Time: Jan 05 17:04:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 01:03:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.332	152	38403	20.000	ng/u1	0.00
20) Naphthalene-d8	11.164	136	153240	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.942	164	133241	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.680	188	361877	20.000	ng/u1	0.00
79) Chrysene-d12	21.975	240	362055	20.000	ng/u1	0.00
88) Perylene-d12	25.436	264	386300	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.626	96	5541	6.535	ng/uL	0.00
4) Pyridine-d5	4.061	84	82993	30.641	ng/u1	0.00
7) Phenol-d5	7.457	99	96713	28.546	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.633	67	41925	30.177	ng/u1	0.00
11) 2-Chlorophenol-d4	7.851	132	66962	29.538	ng/u1	0.00
15) 4-Methylphenol-d8	9.020	113	77166	28.584	ng/u1	0.00
21) Nitrobenzene-d5	9.502	128	35346	29.209	ng/u1	0.00
24) 2-Nitrophenol-d4	10.230	143	45193	28.558	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.771	165	86298	27.357	ng/u1	0.00
31) 4-Chloroaniline-d4	11.288	131	100913	28.427	ng/u1	0.00
46) Dimethylphthalate-d6	14.331	166	315278	29.821	ng/u1	0.00
49) Acenaphthylene-d8	14.643	160	332619	28.640	ng/u1	0.00
54) 4-Nitrophenol-d4	15.107	143	47486	28.453	ng/u1	0.00
60) Fluorene-d10	15.929	176	281156	28.471	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.029	200	58970	23.193	ng/u1	0.00
73) Anthracene-d10	17.780	188	484209	29.112	ng/u1	0.00
81) Pyrene-d10	20.042	212	591746	27.935	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.195	264	562640	28.402	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010523\
Data File : BG056215.D
Acq On : 5 Jan 2023 16:22
Operator : CG/JU
Sample : PB149947BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK947

Quant Time: Jan 05 17:04:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 01:03:13 2023
Response via : Initial Calibration

