

Data Path : S:\HPCHEM1\BNA_G\DATA\BG010615\
 Data File : BG015858.D
 Acq On : 6 Jan 2015 13:31
 Operator : TP/IZ
 Sample : G1001-05
 Misc : LOD-WATER-8PPM
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 LOD-WATER2015-01

Quant Time: Jan 07 17:50:23 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG010515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 06 08:46:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	29926	20.00	ng	0.00
21) Naphthalene-d8	10.51	136	109128	20.00	ng	0.00
38) Acenaphthene-d10	14.37	164	81282	20.00	ng	0.00
63) Phenanthrene-d10	17.11	188	194368	20.00	ng	0.00
75) Chrysene-d12	21.29	240	311438	20.00	ng	0.00
86) Perylene-d12	23.56	264	296902	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	224529	64.52	ng	0.00
7) Phenol-d6	6.91	99	308561	63.67	ng	0.00
23) Nitrobenzene-d5	8.89	82	220202	39.59	ng	0.00
41) 2,4,6-Tribromophenol	15.86	330	188874	61.13	ng	0.00
44) 2-Fluorobiphenyl	12.99	172	464388	40.81	ng	0.00
78) Terphenyl-d14	19.74	244	1108201	39.76	ng	0.00
Target Compounds						
53) 2,4-Dinitrophenol	14.49	184	1764	11.61	ng	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_G\DATA\BG010615\
Data File : BG015858.D
Acq On : 6 Jan 2015 13:31
Operator : TP/IZ
Sample : G1001-05
Misc : LOD-WATER-8PPM
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
LOD-WATER2015-01

Quant Time: Jan 07 17:50:23 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG010515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 06 08:46:19 2015
Response via : Initial Calibration

