


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG010620\
 Data File : BG043925.D
 Acq On : 6 Jan 2020 9:54
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG123019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 31 13:32:23 2019



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	34941	PASS
68	69	0.00	2	0.7	255	PASS
69	198	0.00	100	34.9	34968	PASS
70	69	0.00	2	0.7	245	PASS
127	198	10	80	48.1	48157	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	100066	PASS
199	198	5	9	7.0	7008	PASS
275	198	10	60	25.7	25754	PASS
365	198	1	100	3.3	3341	PASS
441	198	0.01	100	11.6	11656	PASS
442	198	50	100	72.2	72296	PASS
443	442	15	24	20.8	15019	PASS