

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010622\
 Data File : BG051888.D
 Acq On : 6 Jan 2022 11:30
 Operator : CG/JU
 Sample : SSTD02044
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020444

Quant Time: Jan 06 13:33:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 13:28:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.253	152	25532	20.000	ng/u1	0.00
20) Naphthalene-d8	11.090	136	110517	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.891	164	79263	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.640	188	206457	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.958	240	216128	20.000	ng/u1	# 0.00
88) Perylene-d12	25.441	264	227224	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.577	96	6543	10.643	ng/uL	0.00
4) Pyridine-d5	4.006	84	47206	25.413	ng/u1	0.00
7) Phenol-d5	7.389	99	49296	21.624	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.560	67	33291	24.565	ng/u1	0.00
11) 2-Chlorophenol-d4	7.777	132	35330	22.283	ng/u1	0.00
15) 4-Methylphenol-d8	8.952	113	39636	22.493	ng/u1	0.00
21) Nitrobenzene-d5	9.422	128	17950	22.153	ng/u1	0.00
24) 2-Nitrophenol-d4	10.150	143	19986	22.564	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.702	165	38505	20.046	ng/u1	0.00
31) 4-Chloroaniline-d4	11.214	131	53477	21.146	ng/u1	0.00
46) Dimethylphthalate-d6	14.280	166	137934	21.742	ng/u1	0.00
49) Acenaphthylene-d8	14.585	160	166114	21.083	ng/u1	0.00
54) 4-Nitrophenol-d4	15.067	143	23104	22.394	ng/u1	0.00
60) Fluorene-d10	15.878	176	118503	20.849	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.984	200	26501	24.637	ng/u1	0.00
73) Anthracene-d10	17.740	188	204312	20.745	ng/u1	0.00
81) Pyrene-d10	20.013	212	263821	20.282	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.195	264	244400	20.442	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.618	88	6566	10.233	ng/uL#	93
5) Pyridine	4.029	79	48336	25.938	ng/u1	93
6) Benzaldehyde	7.378	77	39049	27.455	ng/u1	98
8) Phenol	7.419	94	53332	23.528	ng/u1	89
10) Bis(2-Chloroethyl)ether	7.654	93	40008	23.135	ng/u1	98
12) 2-Chlorophenol	7.812	128	36015	22.144	ng/u1	99
13) 2-Methylphenol	8.688	108	38364	22.431	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.776	45	60975	25.994	ng/u1#	95
16) Acetophenone	9.075	105	62987	22.756	ng/u1	93
17) N-Nitroso-di-n-propyla...	9.058	70	39932	24.043	ng/u1#	88
18) 4-Methylphenol	9.016	108	41488	23.015	ng/u1	93
19) Hexachloroethane	9.357	117	15581	21.956	ng/u1#	84
22) Nitrobenzene	9.463	77	56866	24.129	ng/u1	91
23) Isophorone	9.992	82	107110	22.638	ng/u1	99
25) 2-Nitrophenol	10.191	139	21741	22.962	ng/u1	90
26) 2,4-Dimethylphenol	10.232	107	48052	21.064	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.473	93	54472	21.681	ng/u1	97
29) 2,4-Dichlorophenol	10.732	162	37669	20.611	ng/u1	96
30) Naphthalene	11.143	128	122178	20.520	ng/u1	99
32) 4-Chloroaniline	11.243	127	53774	20.823	ng/u1	97
33) Hexachlorobutadiene	11.431	225	30583	19.710	ng/u1	94
34) Caprolactam	11.983	113	8905	16.094	ng/u1	95
35) 4-Chloro-3-methylphenol	12.347	107	45988	22.457	ng/u1	93
36) 2-Methylnaphthalene	12.735	142	87134	20.513	ng/u1	89

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010622\
 Data File : BG051888.D
 Acq On : 6 Jan 2022 11:30
 Operator : CG/JU
 Sample : SST02044
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST020444


Quant Time: Jan 06 13:33:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 13:28:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.952	142	91891	20.834	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	13.099	216	57620	20.899	ng/ul	93
40) Hexachlorocyclopentadiene	13.082	237	38729	19.347	ng/ul	97
41) 2,4,6-Trichlorophenol	13.328	196	36825	22.064	ng/ul	88
42) 2,4,5-Trichlorophenol	13.399	196	40447	23.124	ng/ul	98
43) 1,1'-Biphenyl	13.728	154	128735	21.203	ng/ul	98
44) 2-Chloronaphthalene	13.775	162	98531	20.991	ng/ul	99
45) 2-Nitroaniline	13.969	65	37818	26.373	ng/ul	98
47) Dimethylphthalate	14.327	163	137110	22.261	ng/ul	99
48) 2,6-Dinitrotoluene	14.450	165	28117	23.880	ng/ul	96
50) Acenaphthylene	14.615	152	162750	21.112	ng/ul	97
51) 3-Nitroaniline	14.779	138	27707	24.609	ng/ul	90
52) Acenaphthene	14.955	153	107755	21.158	ng/ul	93
53) 2,4-Dinitrophenol	14.985	184	15453	22.704	ng/ul	90
55) 4-Nitrophenol	15.079	109	26047	23.532	ng/ul	92
56) Dibenzofuran	15.284	168	159576	21.276	ng/ul	98
57) 2,4-Dinitrotoluene	15.232	165	41530	24.886	ng/ul#	89
58) 2,3,4,6-Tetrachlorophenol	15.508	232	36925	22.212	ng/ul#	88
59) Diethylphthalate	15.684	149	138319	21.558	ng/ul	98
61) Fluorene	15.937	166	130079	21.049	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.919	204	73670	21.402	ng/ul	95
63) 4-Nitroaniline	15.937	138	30224	26.175	ng/ul	88
66) 4,6-Dinitro-2-methylph...	16.001	198	24589	23.277	ng/ul#	99
67) N-Nitrosodiphenylamine	16.130	169	115178	20.741	ng/ul	97
68) 4-Bromophenyl-phenylether	16.818	248	50475	24.006	ng/ul	90
69) Hexachlorobenzene	16.947	284	55642	21.059	ng/ul#	89
70) Atrazine	17.070	200	53772	21.579	ng/ul	98
71) Pentachlorophenol	17.288	266	29442	18.187	ng/ul	97
72) Phenanthrene	17.681	178	225599	21.245	ng/ul	99
74) Anthracene	17.775	178	222504	20.741	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.698	216	64539	20.320	ng/uL	97
76) Pentachlorobenzene	15.214	250	61800	20.354	ng/uL	98
77) Carbazole	18.034	167	207366	21.586	ng/ul	98
78) Di-n-butylphthalate	18.580	149	244611	21.159	ng/ul	98
80) Fluoranthene	19.679	202	300359	20.117	ng/ul#	91
82) Pyrene	20.043	202	295664	20.010	ng/ul#	92
83) Butylbenzylphthalate	20.912	149	104716	21.132	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.834	252	110130	21.621	ng/ul#	97
85) Benzo(a)anthracene	21.934	228	296108	21.058	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.817	149	148227	21.355	ng/ul#	99
87) Chrysene	22.005	228	289275	21.294	ng/ul	98
89) Di-n-octyl phthalate	23.121	149	254951	20.832	ng/ul	100
90) Benzo(b)fluoranthene	24.325	252	310764	20.418	ng/ul#	94
91) Benzo(k)fluoranthene	24.396	252	289186	20.290	ng/ul#	96
93) Benzo(a)pyrene	25.271	252	289420	20.149	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.448	276	334756	21.347	ng/ul#	91
95) Dibenzo(a,h)anthracene	29.512	278	283972	21.202	ng/ul#	93
96) Benzo(g,h,i)perylene	30.687	276	146835	11.049	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 06 13:33:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 06 13:28:50 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.233 ng/uL
RT: 3.618 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion: 88 Resp: 6566
Ion Ratio Lower Upper
88 100
43 41.2 23.0 34.4#
58 78.5 62.4 93.6



#5
Pyridine
Concen: 25.938 ng/uL
RT: 4.029 min Scan# 92
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion: 79 Resp: 48336
Ion Ratio Lower Upper
79 100
52 81.0 60.8 91.2
51 43.4 30.5 45.7









#6
Benzaldehyde
Concen: 27.455 ng/ul
RT: 7.378 min Scan# 662
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion: 77 Resp: 39049
Ion Ratio Lower Upper
77 100
105 84.6 70.4 105.6
106 76.2 61.2 91.8



#8
Phenol
Concen: 23.528 ng/ul
RT: 7.419 min Scan# 669
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion: 94 Resp: 53332
Ion Ratio Lower Upper
94 100
65 37.2 25.1 37.7
66 46.1 31.6 47.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 23.135 ng/ul
 RT: 7.654 min Scan# 709
 Delta R.T. 0.000 min
 Lab File: BG051888.D
 Acq: 6 Jan 2022 11:30




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020444


Tgt Ion: 93	Resp:	40008
Ion	Ratio	Lower Upper
93	100	
63	84.6	68.6 103.0
95	30.4	26.1 39.1



#12
 2-Chlorophenol
 Concen: 22.144 ng/ul
 RT: 7.812 min Scan# 736
 Delta R.T. 0.000 min
 Lab File: BG051888.D
 Acq: 6 Jan 2022 11:30

Tgt Ion: 128	Resp:	36015
Ion	Ratio	Lower Upper
128	100	
64	56.6	45.3 67.9
130	34.7	27.1 40.7









#13
2-Methylphenol
Concen: 22.431 ng/ul
RT: 8.688 min Scan# 885
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:108 Resp: 38364
Ion Ratio Lower Upper
108 100
107 88.1 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.994 ng/ul
RT: 8.776 min Scan# 900
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion: 45 Resp: 60975
Ion Ratio Lower Upper
45 100
77 14.8 10.7 16.1
79 12.9 8.5 12.7#









#16
Acetophenone
Concen: 22.756 ng/ul
RT: 9.075 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:105 Resp: 62987
Ion Ratio Lower Upper
105 100
77 91.1 67.3 100.9
51 32.1 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 24.043 ng/ul
RT: 9.058 min Scan# 948
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion: 70 Resp: 39932
Ion Ratio Lower Upper
70 100
42 69.2 46.0 69.0#
101 8.9 7.0 10.4
130 13.9 13.4 20.2









#18
4-Methylphenol
Concen: 23.015 ng/ul
RT: 9.016 min Scan# 941
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:108 Resp: 41488
Ion Ratio Lower Upper
108 100
107 110.8 95.0 142.4



#19
Hexachloroethane
Concen: 21.956 ng/ul
RT: 9.357 min Scan# 999
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:117 Resp: 15581
Ion Ratio Lower Upper
117 100
201 116.4 77.2 115.8#
199 69.3 49.9 74.9





#22

Nitrobenzene

Concen: 24.129 ng/ul

RT: 9.463 min Scan# 1017

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion: 77 Resp: 56866

Ion Ratio Lower Upper

77 100

123 31.6 30.1 45.1

65 14.1 9.6 14.4



#23

Isophorone

Concen: 22.638 ng/ul

RT: 9.992 min Scan# 1107

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion: 82 Resp: 107110


Ion Ratio Lower Upper

82 100

95 8.0 5.6 8.4

138 14.6 11.8 17.8









#25
2-Nitrophenol
Concen: 22.962 ng/ul
RT: 10.191 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:139 Resp: 21741
Ion Ratio Lower Upper
139 100
65 66.3 48.0 72.0
109 42.6 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 21.064 ng/ul
RT: 10.232 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:107 Resp: 48052
Ion Ratio Lower Upper
107 100
121 42.9 39.3 58.9
122 76.1 63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 21.681 ng/ul

RT: 10.473 min Scan# 1189

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion: 93 Resp: 54472

Ion Ratio Lower Upper

93 100

95 34.6 26.5 39.7

123 11.8 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 20.611 ng/ul

RT: 10.732 min Scan# 1233

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion:162 Resp: 37669


Ion Ratio Lower Upper

162 100

164 63.2 49.4 74.0

98 42.4 30.4 45.6









#30
Naphthalene
Concen: 20.520 ng/ul
RT: 11.143 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:128 Resp: 122178
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 14.1 11.8 17.6



#32
4-Chloroaniline
Concen: 20.823 ng/ul
RT: 11.243 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:127 Resp: 53774
Ion Ratio Lower Upper
127 100
129 31.1 26.2 39.2









#33
Hexachlorobutadiene
Concen: 19.710 ng/ul
RT: 11.431 min Scan# 11431
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:225 Resp: 30583
Ion Ratio Lower Upper
225 100
223 66.2 50.6 75.8
227 68.8 49.8 74.8



#34
Caprolactam
Concen: 16.094 ng/ul
RT: 11.983 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:113 Resp: 8905
Ion Ratio Lower Upper
113 100
55 194.2 147.0 220.6
56 134.0 109.2 163.8





#35

4-Chloro-3-methylphenol

Concen: 22.457 ng/ul

RT: 12.347 min Scan# 1508

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:107 Resp: 45988

Ion Ratio Lower Upper

107 100

144 26.7 19.8 29.6

142 76.7 56.2 84.4



#36

2-Methylnaphthalene

Concen: 20.513 ng/ul

RT: 12.735 min Scan# 1574

Delta R.T. 0.000 min

Lab File: BG051888.D




Acq: 6 Jan 2022 11:30


Tgt Ion:142 Resp: 87134

Ion Ratio Lower Upper

142 100

141 101.8 73.3 109.9





#37

1-Methylnaphthalene

Concen: 20.834 ng/ul

RT: 12.952 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:142 Resp: 91891

Ion Ratio Lower Upper

142 100

141 94.5 77.8 116.8

116 4.9 3.3 4.9#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.899 ng/ul

RT: 13.099 min Scan# 1636

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Tgt Ion:216 Resp: 57620




Ion Ratio Lower Upper


216 100

214 84.0 61.0 91.4

179 20.6 16.2 24.2

108 14.3 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 19.347 ng/ul
RT: 13.082 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:237 Resp: 38729
Ion Ratio Lower Upper
237 100
235 64.0 49.4 74.0
272 14.7 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 22.064 ng/ul
RT: 13.328 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:196 Resp: 36825
Ion Ratio Lower Upper
196 100
198 104.2 74.3 111.5
200 35.5 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 23.124 ng/ul
RT: 13.399 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:196 Resp: 40447
Ion Ratio Lower Upper
196 100
198 97.8 77.0 115.4
200 33.6 25.4 38.2



#43
1,1'-Biphenyl
Concen: 21.203 ng/ul
RT: 13.728 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:154 Resp: 128735
Ion Ratio Lower Upper
154 100
153 43.6 34.5 51.7
76 14.3 13.1 19.7









#44
2-Chloronaphthalene
Concen: 20.991 ng/ul
RT: 13.775 min Scan# 1751
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:162 Resp: 98531
Ion Ratio Lower Upper
162 100
164 34.0 26.1 39.1
127 38.7 31.0 46.4



#45
2-Nitroaniline
Concen: 26.373 ng/ul
RT: 13.969 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion: 65 Resp: 37818
Ion Ratio Lower Upper
65 100
92 65.8 53.0 79.4
138 87.2 67.4 101.0





#47

Dimethylphthalate

Concen: 22.261 ng/ul

RT: 14.327 min Scan# 1845

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:163 Resp: 137110

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.4

164 9.9 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 23.880 ng/ul

RT: 14.450 min Scan# 1866

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion:165 Resp: 28117


Ion Ratio Lower Upper

165 100

89 67.6 57.4 86.0

121 23.1 18.8 28.2









#50
Acenaphthylene
Concen: 21.112 ng/ul
RT: 14.615 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:152 Resp: 162750
Ion Ratio Lower Upper
152 100
151 19.5 17.4 26.0
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 24.609 ng/ul
RT: 14.779 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:138 Resp: 27707
Ion Ratio Lower Upper
138 100
108 11.8 8.8 13.2
92 134.8 98.0 147.0









#52
Acenaphthene
Concen: 21.158 ng/ul
RT: 14.955 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:153 Resp: 107755
Ion Ratio Lower Upper
153 100
152 49.3 36.5 54.7
154 88.5 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 22.704 ng/ul
RT: 14.985 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:184 Resp: 15453
Ion Ratio Lower Upper
184 100
63 74.4 66.2 99.2
154 58.3 53.6 80.4









#55
4-Nitrophenol
Concen: 23.532 ng/ul
RT: 15.079 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:109 Resp: 26047
Ion Ratio Lower Upper
109 100
139 98.7 88.7 133.1
65 136.2 113.6 170.4



#56
Dibenzofuran
Concen: 21.276 ng/ul
RT: 15.284 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:168 Resp: 159576
Ion Ratio Lower Upper
168 100
139 40.5 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 24.886 ng/ul
RT: 15.232 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:165 Resp: 41530
Ion Ratio Lower Upper
165 100
63 60.8 42.1 63.1
182 3.3 1.8 2.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.212 ng/ul
RT: 15.508 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:232 Resp: 36925
Ion Ratio Lower Upper
232 100
131 37.7 38.9 58.3#
130 3.2 2.4 3.6
166 28.2 25.8 38.8









#59
Diethylphthalate
Concen: 21.558 ng/ul
RT: 15.684 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:149 Resp: 138319
Ion Ratio Lower Upper
149 100
177 22.0 16.8 25.2
150 13.7 10.1 15.1



#61
Fluorene
Concen: 21.049 ng/ul
RT: 15.937 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:166 Resp: 130079
Ion Ratio Lower Upper
166 100
165 105.7 81.0 121.6
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.402 ng/ul
RT: 15.919 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:204 Resp: 73670
Ion Ratio Lower Upper
204 100
206 35.1 27.4 41.2
141 55.3 48.6 73.0



#63
4-Nitroaniline
Concen: 26.175 ng/ul
RT: 15.937 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:138 Resp: 30224
Ion Ratio Lower Upper
138 100
92 50.4 49.3 73.9
108 100.9 72.6 108.8









#66
4,6-Dinitro-2-methylphenol
Concen: 23.277 ng/ul
RT: 16.001 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:198 Resp: 24589
Ion Ratio Lower Upper
198 100
182 3.5 3.8 5.8#
77 28.8 23.2 34.8



#67
N-Nitrosodiphenylamine
Concen: 20.741 ng/ul
RT: 16.130 min Scan# 2152
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:169 Resp: 115178
Ion Ratio Lower Upper
169 100
168 69.1 58.5 87.7
167 39.3 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 24.006 ng/ul
RT: 16.818 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:248 Resp: 50475
Ion Ratio Lower Upper
248 100
250 97.5 75.5 113.3
141 67.8 67.6 101.4



#69
Hexachlorobenzene
Concen: 21.059 ng/ul
RT: 16.947 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:284 Resp: 55642
Ion Ratio Lower Upper
284 100
142 27.9 31.9 47.9#
249 33.0 25.8 38.6









#70
Atrazine
Concen: 21.579 ng/ul
RT: 17.070 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:200 Resp: 53772
Ion Ratio Lower Upper
200 100
173 22.4 20.2 30.2
215 47.5 38.1 57.1



#71
Pentachlorophenol
Concen: 18.187 ng/ul
RT: 17.288 min Scan# 2349
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:266 Resp: 29442
Ion Ratio Lower Upper
266 100
264 65.6 54.3 81.5
268 60.5 51.0 76.6





#72

Phenanthrene

Concen: 21.245 ng/ul

RT: 17.681 min Scan# 2416

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:178 Resp: 225599

Ion Ratio Lower Upper

178 100

179 16.0 12.3 18.5

176 19.6 15.5 23.3



#74

Anthracene

Concen: 20.741 ng/ul

RT: 17.775 min Scan# 2432

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion:178 Resp: 222504


Ion Ratio Lower Upper

178 100

179 16.4 13.0 19.6

176 19.3 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.320 ng/uL
RT: 13.698 min Scan# 1738
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:216 Resp: 64539
Ion Ratio Lower Upper
216 100
214 80.8 61.3 91.9
218 46.3 37.0 55.6



#76
Pentachlorobenzene
Concen: 20.354 ng/uL
RT: 15.214 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:250 Resp: 61800
Ion Ratio Lower Upper
250 100
248 64.2 50.5 75.7
252 62.3 51.2 76.8









#77
 Carbazole
 Concen: 21.586 ng/ul
 RT: 18.034 min Scan# 2476
 Delta R.T. 0.000 min
 Lab File: BG051888.D
 Acq: 6 Jan 2022 11:30




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020444


Tgt Ion:167 Resp: 207366
 Ion Ratio Lower Upper
 167 100
 166 23.2 17.3 25.9
 139 13.0 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 21.159 ng/ul
 RT: 18.580 min Scan# 2569
 Delta R.T. 0.000 min
 Lab File: BG051888.D
 Acq: 6 Jan 2022 11:30

Tgt Ion:149 Resp: 244611
 Ion Ratio Lower Upper
 149 100
 150 8.2 7.5 11.3
 104 6.6 5.0 7.6









#80
Fluoranthene
Concen: 20.117 ng/ul
RT: 19.679 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:202 Resp: 300359
Ion Ratio Lower Upper
202 100
101 7.8 9.3 13.9#
100 5.9 7.0 10.6#



#82
Pyrene
Concen: 20.010 ng/ul
RT: 20.043 min Scan# 2818
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:202 Resp: 295664
Ion Ratio Lower Upper
202 100
101 9.5 10.7 16.1#
100 8.9 9.0 13.6#





#83

Butylbenzylphthalate

Concen: 21.132 ng/ul

RT: 20.912 min Scan# 2966

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:149 Resp: 104716

Ion Ratio Lower Upper

149 100

91 81.1 65.2 97.8

206 22.2 16.2 24.2



#84

3,3'-Dichlorobenzidine

Concen: 21.621 ng/ul

RT: 21.834 min Scan# 3123

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion:252 Resp: 110130


Ion Ratio Lower Upper

252 100

254 64.8 50.8 76.2

126 8.3 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 21.058 ng/ul
RT: 21.934 min Scan# 3140
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:228 Resp: 296108
Ion Ratio Lower Upper
228 100
229 19.0 15.7 23.5
226 27.5 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.355 ng/ul
RT: 21.817 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:149 Resp: 148227
Ion Ratio Lower Upper
149 100
167 26.0 21.1 31.7
279 4.7 2.7 4.1#





#87

Chrysene

Concen: 21.294 ng/ul

RT: 22.005 min Scan# 3152

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:228 Resp: 289275

Ion Ratio Lower Upper

228 100

226 29.5 24.8 37.2

229 19.5 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 20.832 ng/ul




RT: 23.121 min Scan# 3342


Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Tgt Ion:149 Resp: 254951









#90
Benzo(b)fluoranthene
Concen: 20.418 ng/ul
RT: 24.325 min Scan# 3547
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30




Instrument :
BNA_G
ClientSampleId :
SSTD020444


Tgt Ion:252 Resp: 310764
Ion Ratio Lower Upper
252 100
253 20.4 17.8 26.6
125 6.2 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 20.290 ng/ul
RT: 24.396 min Scan# 3559
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:252 Resp: 289186
Ion Ratio Lower Upper
252 100
253 21.0 17.3 25.9
125 6.3 7.8 11.6#





#93

Benzo(a)pyrene

Concen: 20.149 ng/ul

RT: 25.271 min Scan# 3708

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30

Instrument :

BNA_G

ClientSampleId :

SSTD020444





Tgt Ion:252 Resp: 289420

Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.2

125 6.6 9.1 13.7#



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.347 ng/ul

RT: 29.448 min Scan# 4419

Delta R.T. 0.000 min

Lab File: BG051888.D

Acq: 6 Jan 2022 11:30




Tgt Ion:276 Resp: 334756


Ion Ratio Lower Upper

276 100

138 11.0 15.5 23.3#

277 26.4 20.5 30.7









#95
Dibenzo(a,h)anthracene
Concen: 21.202 ng/ul
RT: 29.512 min Scan# 4430
Delta R.T. 0.000 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Instrument :
BNA_G
ClientSampleId :
SSTD020444

Tgt Ion:278 Resp: 283972
Ion Ratio Lower Upper
278 100
139 8.5 12.2 18.4#
279 23.8 19.5 29.3



#96
Benzo(g,h,i)perylene
Concen: 11.049 ng/ul
RT: 30.687 min Scan# 4630
Delta R.T. -0.012 min
Lab File: BG051888.D
Acq: 6 Jan 2022 11:30

Tgt Ion:276 Resp: 146835
Ion Ratio Lower Upper
276 100
138 12.8 16.6 24.8#
277 24.5 17.6 26.4

