

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010719\
 Data File : BG038971.D
 Acq On : 7 Jan 2019 17:54
 Operator : JU/SJ
 Sample : K1061-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TRE-COMP-1

Manual Integrations
 APPROVED

Sohil
 1/8/2019 4:49:13 PM

Quant Time: Jan 08 03:26:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	28003	20.00	ng	-0.01
21) Naphthalene-d8	10.86	136	104357	20.00	ng	-0.02
39) Acenaphthene-d10	14.68	164	70146	20.00	ng	-0.02
64) Phenanthrene-d10	17.43	188	171310	20.00	ng	-0.01
76) Chrysene-d12	21.71	240	176244	20.00	ng	-0.02
87) Perylene-d12	25.00	264	190224	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	87582	55.47	ng	0.00
7) Phenol-d6	7.21	99	74242	34.25	ng	0.00
23) Nitrobenzene-d5	9.22	82	150085	73.47	ng	-0.02
42) 2,4,6-Tribromophenol	16.18	330	130196	134.68	ng	-0.01
45) 2-Fluorobiphenyl	13.30	172	395128	77.28	ng	-0.02
79) Terphenyl-d14	20.03	244	685924	84.70	ng	-0.01
Target Compounds						
10) Phenol	7.24	94	4659	2.023	ng	Qvalue # 72
31) Naphthalene	10.91	128	67319	13.093	ng	98
32) Benzoic acid	10.15	122	7319	16.156	ng	95
37) 2-Methylnaphthalene	12.51	142	22494	6.132	ng	95
38) 1-Methylnaphthalene	12.73	142	12017m	3.376	ng	
50) Dimethylphthalate	14.13	163	274833	47.978	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010719\
Data File : BG038971.D
Acq On : 7 Jan 2019 17:54
Operator : JU/SJ
Sample : K1061-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TRE-COMP-1

Manual Integrations
APPROVED

Sohil
1/8/2019 4:49:13 PM

Quant Time: Jan 08 03:26:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

