

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010720\
 Data File : BG043956.D
 Acq On : 7 Jan 2020 15:44
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 07 16:41:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG123019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 31 13:32:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	40.463	-1.2	94	0.00
3	Pyridine	40.000	41.747	-4.4	96	0.00
4	n-Nitrosodimethylamine	40.000	39.530	1.2	93	0.00
5 S	2-Fluorophenol	80.000	84.051	-5.1	95	0.00
6	Aniline	40.000	40.250	-0.6	94	0.00
7 S	Phenol-d6	80.000	85.961	-7.5	99	0.00
8	2-Chlorophenol	40.000	41.243	-3.1	96	0.00
9	Benzaldehyde	40.000	35.822	10.4	88	0.00
10 C	Phenol	40.000	42.058	-5.1	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.351	1.6	91	0.00
12	1,3-Dichlorobenzene	40.000	40.183	-0.5	93	0.00
13 C	1,4-Dichlorobenzene	40.000	40.907	-2.3	94	0.00
14	1,2-Dichlorobenzene	40.000	40.358	-0.9	94	0.00
15	Benzyl Alcohol	40.000	40.374	-0.9	94	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	38.163	4.6	90	0.00
17	2-Methylphenol	40.000	41.299	-3.2	97	0.00
18	Hexachloroethane	40.000	40.273	-0.7	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.359	1.6	92	0.00
20	3+4-Methylphenols	40.000	41.823	-4.6	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	39.023	2.4	93	0.00
23 S	Nitrobenzene-d5	80.000	76.116	4.9	94	0.00
24	Nitrobenzene	40.000	38.760	3.1	94	0.00
25	Isophorone	40.000	38.726	3.2	93	0.00
26 C	2-Nitrophenol	40.000	41.330	-3.3	102	0.00
27	2,4-Dimethylphenol	40.000	39.513	1.2	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.541	3.6	92	0.00
29 C	2,4-Dichlorophenol	40.000	41.052	-2.6	99	0.00
30	1,2,4-Trichlorobenzene	40.000	39.475	1.3	96	0.00
31	Naphthalene	40.000	40.419	-1.0	95	0.00
32	Benzoic acid	40.000	28.338	29.2#	75	0.01
33	4-Chloroaniline	40.000	40.233	-0.6	96	0.00
34 C	Hexachlorobutadiene	40.000	38.850	2.9	95	0.00
35	Caprolactam	40.000	40.934	-2.3	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.667	-4.2	100	0.00
37	2-Methylnaphthalene	40.000	40.752	-1.9	98	0.00
38	1-Methylnaphthalene	40.000	40.382	-1.0	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.491	1.3	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.659	5.9	93	0.00
42 S	2,4,6-Tribromophenol	80.000	84.763	-6.0	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.183	-0.5	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.569	-1.4	99	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010720\
 Data File : BG043956.D
 Acq On : 7 Jan 2020 15:44
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 07 16:41:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG123019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 31 13:32:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	82.971	-3.7	96	0.00
46	1,1'-Biphenyl	40.000	40.426	-1.1	96	0.00
47	2-Chloronaphthalene	40.000	40.250	-0.6	98	0.00
48	2-Nitroaniline	40.000	40.980	-2.4	101	0.00
49	Acenaphthylene	40.000	41.058	-2.6	98	0.00
50	Dimethylphthalate	40.000	40.862	-2.2	97	0.00
51	2,6-Dinitrotoluene	40.000	40.954	-2.4	101	0.00
52 C	Acenaphthene	40.000	40.777	-1.9	100	0.00
53	3-Nitroaniline	40.000	41.190	-3.0	100	0.00
54 P	2,4-Dinitrophenol	40.000	36.961	7.6	95	0.00
55	Dibenzofuran	40.000	41.321	-3.3	98	0.00
56 P	4-Nitrophenol	40.000	40.554	-1.4	96	0.00
57	2,4-Dinitrotoluene	40.000	41.825	-4.6	101	0.00
58	Fluorene	40.000	41.417	-3.5	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.030	-2.6	99	0.00
60	Diethylphthalate	40.000	41.617	-4.0	99	0.00
61	4-Chlorophenyl-phenylether	40.000	40.999	-2.5	101	0.00
62	4-Nitroaniline	40.000	42.055	-5.1	101	0.00
63	Azobenzene	40.000	41.190	-3.0	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.213	2.0	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.337	1.7	101	0.00
67	4-Bromophenyl-phenylether	40.000	39.013	2.5	102	0.00
68	Hexachlorobenzene	40.000	39.346	1.6	102	0.00
69	Atrazine	40.000	39.505	1.2	95	0.00
70 C	Pentachlorophenol	40.000	37.067	7.3	91	0.00
71	Phenanthrene	40.000	40.596	-1.5	100	0.00
72	Anthracene	40.000	40.554	-1.4	100	0.00
73	Carbazole	40.000	40.786	-2.0	100	0.00
74	Di-n-butylphthalate	40.000	39.837	0.4	99	0.00
75 C	Fluoranthene	40.000	39.350	1.6	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	50.269	-25.7#	110	0.00
78	Pyrene	40.000	39.498	1.3	99	0.00
79 S	Terphenyl-d14	80.000	80.520	-0.6	99	0.00
80	Butylbenzylphthalate	40.000	41.540	-3.8	101	0.00
81	Benzo(a)anthracene	40.000	40.603	-1.5	99	0.00
82	3,3'-Dichlorobenzidine	40.000	39.851	0.4	96	0.00
83	Chrysene	40.000	41.253	-3.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.813	-2.0	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.866	-4.7	100	-0.01
86	Indeno(1,2,3-cd)pyrene	40.000	39.347	1.6	98	-0.02
87 I	Perylene-d12	20.000	20.000	0.0	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010720\
Data File : BG043956.D
Acq On : 7 Jan 2020 15:44
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 07 16:41:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG123019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 31 13:32:23 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	39.063	2.3	100	0.00
89	Benzo(k)fluoranthene	40.000	40.215	-0.5	99	0.00
90 C	Benzo(a)pyrene	40.000	39.557	1.1	99	0.00
91	Dibenzo(a,h)anthracene	40.000	38.470	3.8	98	-0.01
92	Benzo(q,h,i)perylene	40.000	37.464	6.3	95	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0