

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG010915\
 Data File : BG015878.D
 Acq On : 9 Jan 2015 20:04
 Operator : TP/IZ
 Sample : G1044-05
 Misc : W
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP-9970

Quant Time: Jan 12 13:23:51 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG010515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 06 08:46:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	32328	20.00	ng	0.00
21) Naphthalene-d8	10.52	136	115485	20.00	ng	0.00
38) Acenaphthene-d10	14.37	164	86800	20.00	ng	0.00
63) Phenanthrene-d10	17.12	188	204767	20.00	ng	0.00
75) Chrysene-d12	21.30	240	269721	20.00	ng	0.00
86) Perylene-d12	23.57	264	214731	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.34	112	113940	30.31	ng	0.00
7) Phenol-d6	6.93	99	106365	20.32	ng	0.02
23) Nitrobenzene-d5	8.89	82	255670	43.44	ng	0.00
41) 2,4,6-Tribromophenol	15.87	330	234082	70.94	ng	0.00
44) 2-Fluorobiphenyl	12.99	172	571030	46.99	ng	0.00
78) Terphenyl-d14	19.75	244	965135	39.98	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.26	88	5738	6.66	ng	# 63
6) Aniline	7.07	93	33977	9.60	ng	95
8) 2-Chlorophenol	7.30	128	14577	7.52	ng	# 76
10) Phenol	6.95	94	44932	15.76	ng	79
11) bis(2-Chloroethyl)ether	7.16	93	20739	9.87	ng	93
12) 1,3-Dichlorobenzene	7.62	146	4827m	2.08	ng	
13) 1,4-Dichlorobenzene	7.77	146	29347	12.38	ng	96
14) 1,2-Dichlorobenzene	8.08	146	28595	12.71	ng	99
17) 2-Methylphenol	8.19	107	11743m	6.04	ng	
20) 3+4-Methylphenols	8.51	107	82547	31.84	ng	86
27) 2,4-Dimethylphenol	9.71	122	18380	9.67	ng	85
31) Naphthalene	10.57	128	267873	43.80	ng	99
32) Benzoic acid	9.89	122	43764	46.17	ng	# 67
37) 2-Methylnaphthalene	12.18	142	38133	8.33	ng	96
49) Dimethylphthalate	13.83	163	25618	4.00	ng	97
51) Acenaphthene	14.44	154	12934	2.61	ng	89
59) Diethylphthalate	15.19	149	26078	3.69	ng	# 93
65) n-Nitrosodiphenylamine	15.63	169	83891	14.25	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG010915\
Data File : BG015878.D
Acq On : 9 Jan 2015 20:04
Operator : TP/IZ
Sample : G1044-05
Misc : W
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SP-9970

Quant Time: Jan 12 13:23:51 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG010515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 06 08:46:19 2015
Response via : Initial Calibration

