


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG010918\
 Data File : BG031945.D
 Acq On : 9 Jan 2018 13:52
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG122117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 05 15:31:34 2018



AutoFind: Scans 2592, 2593, 2594; Background Corrected with Scan 2586

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	18.3	12657	PASS
68	69	0.00	2	0.6	88	PASS
69	198	0.00	100	20.4	14056	PASS
70	69	0.00	2	0.5	68	PASS
127	198	10	80	32.9	22685	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	69034	PASS
199	198	5	9	7.0	4841	PASS
275	198	10	60	31.4	21658	PASS
365	198	1	100	3.8	2602	PASS
441	442	0.01	24	16.5	8851	PASS
442	198	50	100	77.6	53570	PASS
443	442	15	24	19.2	10271	PASS