


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010923\
Data File : BG056226.D
Acq On : 9 Jan 2023 8:35
Operator : CG/JU
Sample : DFTPP489
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP489

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
Title : SVOA CALIBRATION
Last Update : Thu Jan 05 01:03:13 2023



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	14.2	29347	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	21.2	43957	PASS
70	69	0.00	2	0.8	340	PASS
127	198	10	80	34.4	71283	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	206976	PASS
199	198	5	9	6.9	14239	PASS
275	198	10	60	33.6	69595	PASS
365	198	1	100	4.4	9077	PASS
441	198	0.01	100	12.0	24741	PASS
442	198	50	100	72.4	149840	PASS
443	442	15	24	21.1	31544	PASS

m/z	Abundance
73.00	805.0
147.00	1701.0
148.00	306.0
206.90	1062.0
208.10	163.0
209.10	177.0
220.90	1958.0
222.10	821.0
222.90	176.0
267.00	185.0
267.80	164.0
281.00	1207.0
281.90	248.0
282.70	165.0
295.10	600.0
296.10	175.0
326.80	221.0
340.90	374.0
355.00	872.0
356.00	233.0
357.00	245.0
369.00	244.0
401.10	213.0
429.00	563.0
430.00	191.0
431.10	347.0
503.10	152.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

m/z	Abundance
37.10	180.0
38.00	566.0
39.00	2309.0
50.00	10827.0
51.00	36296.0
52.00	1971.0
55.00	238.0
56.00	1241.0
57.00	2713.0
58.10	223.0
60.90	1004.0
61.90	896.0
63.00	2583.0
64.10	329.0
65.00	1059.0
68.90	53136.0
70.00	298.0
73.00	1571.0
74.00	6720.0
75.00	9461.0
76.00	3173.0
77.00	69160.0
78.00	4471.0
79.00	4980.0
80.00	3891.0
80.90	5104.0
81.90	1512.0
83.00	1125.0
85.00	744.0
86.00	1623.0
87.00	810.0
87.90	345.0
88.70	174.0
88.90	180.0
91.00	1295.0
91.90	1403.0
93.00	10478.0
94.00	715.0
96.00	248.0
98.00	9388.0
98.90	5741.0
99.80	632.0
101.00	3549.0
101.90	196.0
103.00	942.0
103.90	2039.0
104.90	1988.0
106.00	807.0
107.00	28120.0
108.00	4182.0
109.90	42152.0
110.90	7198.0
111.90	1057.0
112.90	233.0
115.90	1791.0
117.00	26232.0
118.00	1845.0
120.00	323.0
122.00	1948.0
122.90	3155.0
123.90	1633.0
124.90	1194.0
125.70	405.0
127.00	84128.0
128.00	6516.0
129.00	35320.0
130.00	3386.0
131.00	555.0
132.00	264.0
133.90	926.0
135.00	2905.0
135.90	1234.0
136.90	1695.0
137.90	394.0
138.70	508.0
140.00	545.0
140.90	5441.0
141.90	1748.0
143.00	905.0
144.00	448.0
144.90	430.0
145.80	1414.0
147.00	4106.0
148.00	7412.0
148.90	1687.0
149.90	223.0
151.00	802.0
152.00	391.0
153.00	1397.0
153.90	1103.0
154.90	3394.0
156.00	3835.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

157.00	657.0
158.00	989.0
159.00	894.0
159.90	1628.0
160.90	2823.0
162.00	918.0
162.80	272.0
163.70	217.0
165.00	2680.0
166.00	1821.0
166.90	11088.0
167.90	5062.0
168.90	1041.0
169.80	386.0
170.40	206.0
171.10	393.0
171.90	1115.0
173.00	1611.0
174.00	1823.0
175.00	4220.0
176.00	1503.0
176.90	2483.0
178.00	1129.0
178.90	9746.0
180.00	5821.0
181.00	3115.0
182.00	842.0
183.00	244.0
184.00	784.0
184.90	4459.0
186.00	33016.0
187.00	10616.0
187.90	967.0
189.00	2247.0
189.80	512.0
190.90	1291.0
192.00	2946.0
193.00	3907.0
193.90	713.0
194.90	494.0
196.00	5990.0
197.90	228864.0
198.90	16276.0
200.00	1224.0
201.30	888.0
203.00	2076.0
204.00	10340.0
205.00	14863.0
206.00	67000.0
207.00	11433.0
208.00	2822.0
209.00	1032.0
210.00	1423.0
210.80	2816.0
212.00	522.0
212.90	168.0
214.00	189.0
214.80	784.0
216.00	1901.0
216.90	17608.0
218.00	2851.0
218.70	164.0
220.90	10758.0
221.90	3050.0
223.00	4247.0
224.00	39704.0
224.90	10166.0
225.90	1146.0
227.00	17896.0
227.90	2536.0
229.00	3671.0
230.00	387.0
231.00	1563.0
231.70	268.0
232.80	381.0
234.00	1024.0
234.90	1525.0
235.80	644.0
237.00	1295.0
238.00	233.0
239.00	883.0
240.00	525.0
241.00	1003.0
242.00	1751.0
243.00	2644.0
243.90	30712.0
245.00	4014.0
246.00	6826.0
246.90	1108.0
247.90	217.0
249.00	1094.0
249.80	301.0
250.90	392.0
251.70	347.0
252.90	706.0
255.00	157504.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

255.90	26488.0
257.00	1608.0
258.00	11121.0
259.00	2216.0
260.00	235.0
261.10	206.0
264.00	400.0
264.90	4882.0
265.90	811.0
267.10	380.0
267.90	163.0
270.00	171.0
270.90	540.0
273.00	4463.0
274.00	12550.0
275.00	70072.0
275.90	9607.0
276.90	6549.0
278.00	1021.0
278.90	501.0
280.00	164.0
281.00	2141.0
281.90	520.0
283.10	1151.0
283.90	768.0
284.90	1200.0
286.00	229.0
288.80	233.0
290.00	326.0
291.10	331.0
291.90	332.0
292.90	1509.0
294.00	481.0
296.00	21912.0
296.90	2874.0
298.00	290.0
300.90	364.0
302.10	302.0
302.90	1713.0
303.90	762.0
308.00	302.0
308.90	347.0
309.90	249.0
312.80	164.0
313.80	958.0
314.90	2812.0
315.90	1380.0
316.90	262.0
321.00	717.0
322.00	360.0
322.90	6122.0
324.00	988.0
324.80	308.0
326.90	1412.0
327.90	674.0
328.90	221.0
331.90	508.0
333.00	568.0
333.90	3299.0
334.90	755.0
341.00	843.0
342.00	359.0
345.80	1319.0
346.90	331.0
350.90	172.0
351.90	1669.0
353.00	893.0
354.10	1177.0
355.20	1618.0
356.10	828.0
357.10	329.0
359.10	265.0
364.90	8167.0
365.90	1244.0
369.10	430.0
369.90	304.0
371.10	498.0
372.00	2360.0
373.00	706.0
382.80	543.0
383.90	327.0
390.20	369.0
391.10	309.0
392.10	189.0
400.70	206.0
402.00	881.0
402.90	1303.0
403.90	304.0
420.70	667.0
421.00	617.0
421.90	1133.0
422.90	5955.0
423.90	1269.0
424.90	262.0
429.10	619.0
430.10	349.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

441.00	20952.0
442.00	128944.0
442.90	28152.0
444.00	2479.0
445.00	236.0
489.00	153.0
503.00	345.0
504.10	224.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

m/z	Abundance
38.00	496.0
39.00	2433.0
49.10	565.0
50.00	10145.0
51.00	33376.0
52.10	2097.0
54.80	230.0
56.00	1653.0
57.00	2431.0
58.00	195.0
60.90	537.0
61.90	666.0
63.00	1724.0
63.90	250.0
65.00	951.0
65.90	160.0
69.00	50336.0
70.10	493.0
73.00	1558.0
74.00	6324.0
75.00	9632.0
76.00	3708.0
77.00	69264.0
78.00	4500.0
79.00	5076.0
79.90	3751.0
81.00	5937.0
81.90	1270.0
82.90	1024.0
83.90	155.0
85.00	745.0
85.90	1984.0
87.00	927.0
88.10	182.0
89.00	269.0
91.00	1521.0
92.00	1613.0
93.00	10647.0
94.00	608.0
95.90	438.0
97.10	452.0
98.00	8768.0
99.00	5567.0
100.00	607.0
100.90	3819.0
102.10	342.0
102.90	819.0
104.00	2482.0
105.00	2533.0
106.10	932.0
107.00	28472.0
108.00	3301.0
109.90	41416.0
111.00	6904.0
111.90	1000.0
112.90	350.0
114.30	169.0
116.00	1631.0
117.00	27480.0
118.00	2039.0
120.00	433.0
121.90	2057.0
122.90	2799.0
123.90	1201.0
124.90	1168.0
126.90	82568.0
128.00	6454.0
129.00	37272.0
129.90	2562.0
131.00	824.0
132.00	257.0
133.10	300.0
134.00	1643.0
134.90	3261.0
136.00	1258.0
137.00	1574.0
137.90	338.0
139.00	186.0
140.00	619.0
140.90	6131.0
142.00	1919.0
143.00	1126.0
143.80	280.0
144.90	415.0
146.00	1318.0
147.00	5449.0
148.00	6609.0
149.00	1533.0
149.90	375.0
151.10	1005.0
153.00	1964.0
154.00	1007.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

154.90	2754.0
156.00	4575.0
157.10	862.0
157.80	996.0
159.00	751.0
160.00	2149.0
160.90	2504.0
161.90	906.0
162.90	232.0
163.80	343.0
165.00	2463.0
165.90	2514.0
167.00	11156.0
168.00	4707.0
169.00	945.0
170.00	456.0
170.90	703.0
171.80	1030.0
172.90	1552.0
174.00	2849.0
175.00	5155.0
176.00	1244.0
177.00	2578.0
177.90	864.0
178.90	10018.0
179.90	6934.0
180.90	3398.0
182.10	513.0
182.80	270.0
184.00	915.0
184.90	4979.0
186.00	35344.0
187.00	9816.0
188.00	1112.0
189.00	2289.0
189.90	397.0
190.90	1417.0
191.90	3173.0
193.00	4200.0
193.90	760.0
194.90	840.0
195.90	7272.0
197.90	241792.0
198.90	15933.0
200.00	1702.0
200.80	419.0
201.50	738.0
203.00	2035.0
204.00	10779.0
205.00	18464.0
206.00	72632.0
206.90	10151.0
207.90	3603.0
208.90	1062.0
210.10	1489.0
211.10	2835.0
211.80	571.0
212.90	288.0
213.10	273.0
215.00	505.0
216.00	1782.0
216.90	20864.0
218.00	3108.0
219.20	195.0
221.00	12808.0
221.80	3378.0
222.90	4521.0
224.00	46336.0
225.00	10613.0
225.90	891.0
226.90	21264.0
227.90	2554.0
228.90	4168.0
229.90	659.0
230.90	1723.0
231.90	240.0
233.00	325.0
234.00	1225.0
234.90	1437.0
235.90	969.0
237.00	1561.0
237.70	297.0
238.80	665.0
239.80	382.0
240.90	1675.0
241.90	2628.0
242.90	2017.0
244.00	36296.0
245.00	4148.0
245.90	7631.0
247.00	1527.0
248.00	511.0
248.80	1214.0
250.00	227.0
250.90	483.0
251.90	560.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

252.90	1117.0
254.90	188160.0
255.90	27496.0
257.00	2375.0
258.00	11710.0
259.00	2001.0
259.80	210.0
260.90	509.0
263.00	209.0
263.90	463.0
264.90	5766.0
265.80	766.0
266.90	384.0
268.90	182.0
269.90	336.0
271.00	452.0
271.90	615.0
272.90	5035.0
274.00	15036.0
274.90	84888.0
275.90	10572.0
277.00	7290.0
277.90	1411.0
279.10	229.0
281.00	1779.0
281.90	751.0
283.00	1169.0
284.00	756.0
285.00	1302.0
286.00	345.0
288.80	358.0
289.90	332.0
291.00	153.0
292.00	406.0
292.90	1566.0
293.90	574.0
295.90	27312.0
296.90	4189.0
301.00	310.0
301.80	322.0
303.00	3113.0
303.90	796.0
305.10	174.0
307.90	512.0
308.80	202.0
309.90	257.0
312.80	305.0
314.00	1300.0
314.90	2790.0
315.90	1217.0
317.00	189.0
321.10	744.0
321.90	524.0
323.00	7048.0
323.90	1222.0
325.10	250.0
326.00	257.0
326.90	1510.0
327.90	955.0
328.70	185.0
332.10	672.0
334.00	4561.0
334.90	892.0
338.90	173.0
339.70	164.0
340.90	970.0
342.00	393.0
345.90	1854.0
347.00	420.0
352.00	1868.0
352.90	1308.0
354.00	2223.0
355.10	1607.0
356.10	505.0
357.10	234.0
358.00	260.0
359.10	199.0
364.90	11332.0
365.90	1874.0
369.10	369.0
370.00	401.0
371.10	586.0
372.00	2817.0
373.00	691.0
373.80	150.0
383.00	897.0
384.00	168.0
384.90	274.0
389.70	376.0
391.00	386.0
392.00	419.0
401.10	283.0
402.00	896.0
403.00	1469.0
403.90	550.0
415.30	160.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

420.90	1568.0
423.00	7733.0
424.00	1624.0
429.10	719.0
429.80	171.0
430.30	259.0
431.00	336.0
441.00	30288.0
442.00	178048.0
443.00	36896.0
444.00	3050.0
503.20	246.0
505.10	233.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

m/z	Abundance
37.90	254.0
39.00	1383.0
50.00	5549.0
51.00	18368.0
52.00	920.0
55.90	640.0
57.00	1713.0
61.00	247.0
62.10	667.0
63.00	1725.0
63.80	180.0
65.00	388.0
68.90	28400.0
70.00	229.0
73.00	1325.0
74.00	3756.0
74.90	5540.0
76.00	2041.0
77.00	39064.0
78.10	2574.0
79.00	2618.0
79.90	1970.0
81.00	3384.0
81.90	710.0
83.00	837.0
84.90	471.0
86.00	852.0
87.00	538.0
88.00	164.0
90.90	940.0
91.90	927.0
93.00	5708.0
93.90	344.0
95.00	267.0
96.10	255.0
97.10	216.0
98.00	3804.0
99.00	3149.0
100.00	280.0
101.00	1759.0
103.10	608.0
103.90	1345.0
104.90	1291.0
107.00	16512.0
108.00	2687.0
109.90	24512.0
111.00	3543.0
112.00	585.0
112.90	152.0
116.10	1210.0
117.00	15307.0
117.90	977.0
120.00	253.0
121.00	222.0
122.10	1337.0
123.00	1913.0
124.00	791.0
124.90	599.0
126.90	47152.0
127.90	3531.0
129.00	21392.0
129.90	1885.0
130.90	436.0
133.90	632.0
135.00	1967.0
135.90	617.0
137.00	1036.0
138.90	161.0
140.00	401.0
141.00	3861.0
142.00	921.0
142.90	625.0
144.00	253.0
144.90	232.0
145.90	620.0
147.00	2940.0
148.00	4132.0
149.00	1057.0
149.70	239.0
150.10	228.0
151.00	276.0
152.00	219.0
153.00	1024.0
154.00	718.0
155.00	1841.0
156.00	3266.0
157.10	453.0
157.90	483.0
159.00	651.0
160.00	1682.0
161.00	1618.0
161.90	423.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

162.80	192.0
164.00	296.0
164.90	1228.0
166.10	1413.0
166.90	7080.0
168.00	3304.0
169.00	701.0
169.80	195.0
171.00	379.0
171.90	843.0
173.00	774.0
174.00	2293.0
175.00	3491.0
176.00	899.0
176.90	1438.0
177.90	561.0
178.90	6088.0
180.00	3732.0
180.90	1965.0
182.10	347.0
182.90	252.0
183.90	453.0
185.00	3155.0
186.00	21520.0
187.00	6960.0
188.00	626.0
188.90	1294.0
190.00	292.0
191.10	1038.0
191.90	2024.0
193.00	2897.0
193.90	421.0
194.80	335.0
196.00	5548.0
197.90	150272.0
198.90	10509.0
199.90	1090.0
201.30	843.0
202.90	1114.0
204.00	6075.0
205.00	12064.0
206.00	48656.0
207.00	7363.0
208.00	1574.0
209.00	594.0
210.90	1755.0
211.80	375.0
213.00	273.0
214.90	706.0
215.90	929.0
216.90	13226.0
218.00	1798.0
218.80	245.0
221.00	9244.0
221.90	1951.0
222.90	3480.0
224.00	29272.0
225.00	7541.0
226.00	776.0
226.90	13296.0
228.00	1531.0
229.00	1980.0
229.70	280.0
230.20	311.0
230.90	1091.0
231.90	200.0
232.90	157.0
233.90	787.0
234.90	751.0
235.80	655.0
236.90	1076.0
238.80	353.0
239.90	536.0
240.90	762.0
241.90	1509.0
244.00	24568.0
244.90	2441.0
246.00	5450.0
246.90	1203.0
247.90	301.0
248.90	736.0
249.90	260.0
250.70	206.0
251.80	159.0
253.00	544.0
254.90	124808.0
255.90	18944.0
256.90	1266.0
257.90	8183.0
259.00	1264.0
260.00	180.0
261.00	380.0
263.30	176.0
263.70	175.0
264.00	186.0
264.90	3141.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489

266.00	409.0
267.00	445.0
268.10	228.0
270.10	180.0
271.10	250.0
271.80	508.0
272.90	3804.0
273.90	10929.0
274.90	53824.0
276.00	7181.0
276.90	4959.0
277.80	1064.0
279.00	266.0
281.00	1341.0
282.00	557.0
282.90	754.0
284.10	498.0
284.90	1026.0
286.00	283.0
288.90	155.0
289.70	153.0
290.80	209.0
291.90	254.0
292.90	1032.0
293.80	235.0
295.90	19144.0
296.90	2529.0
301.00	165.0
301.80	304.0
302.90	2006.0
304.00	306.0
308.00	368.0
308.90	341.0
309.80	309.0
313.80	967.0
314.20	770.0
315.00	1955.0
315.90	1140.0
316.80	151.0
321.00	514.0
321.90	244.0
323.00	4449.0
323.90	845.0
324.90	189.0
326.00	218.0
326.90	1178.0
328.00	457.0
332.00	358.0
332.90	482.0
334.00	2920.0
334.90	692.0
338.70	167.0
339.80	158.0
341.10	1096.0
342.10	348.0
345.90	1782.0
347.00	266.0
351.90	1403.0
352.90	1060.0
353.90	1528.0
355.00	1462.0
355.90	422.0
357.00	338.0
364.90	7732.0
365.90	911.0
369.00	539.0
370.00	450.0
371.00	552.0
371.90	2257.0
372.90	504.0
382.90	512.0
384.00	286.0
390.00	281.0
390.90	214.0
391.80	288.0
400.90	238.0
402.10	654.0
402.90	1257.0
404.10	302.0
414.80	210.0
415.10	206.0
420.80	1309.0
421.90	1060.0
422.90	6410.0
423.90	1138.0
429.00	585.0
430.10	230.0
431.00	262.0
441.00	22984.0
442.00	142528.0
443.00	29584.0
443.90	2305.0
445.00	357.0
476.30	150.0

Instrument :
BNA_G
ClientSampleId :
DFTPP489