

Data Path : Z:\HPCHEM1\BNA G\DATA\BG011017\
 Data File : BG025451.D
 Acq On : 10 Jan 2017 10:35
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 10 13:05:25 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 11:26:22 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.469	0.488	-4.1	110	0.00
3	Pyridine	1.361	1.467	-7.8	105	0.00
4	n-Nitrosodimethylamine	0.808	0.821	-1.6	100	0.00
5 S	2-Fluorophenol	1.100	1.113	-1.2	105	0.00
6	Aniline	2.100	2.207	-5.1	107	0.00
7 S	Phenol-d6	1.550	1.624	-4.8	104	0.00
8	2-Chlorophenol	1.241	1.234	0.6	102	0.00
9	Benzaldehyde	1.134	1.028	9.3	96	0.00
10 C	Phenol	1.718	1.802	-4.9	105	0.00
11	bis(2-Chloroethyl)ether	1.324	1.371	-3.5	107	0.00
12	1,3-Dichlorobenzene	1.502	1.553	-3.4	105	0.00
13 C	1,4-Dichlorobenzene	1.557	1.580	-1.5	104	0.00
14	1,2-Dichlorobenzene	1.486	1.462	1.6	102	0.00
15	Benzyl Alcohol	1.479	1.417	4.2	92	0.00
16	2,2'-oxybis(1-Chloropropane	2.451	2.581	-5.3	107	0.00
17	2-Methylphenol	1.128	1.205	-6.8	108	0.00
18	Hexachloroethane	0.598	0.604	-1.0	99	0.00
19 P	n-Nitroso-di-n-propylamine	1.309	1.344	-2.7	104	0.00
20	3+4-Methylphenols	1.561	1.670	-7.0	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.556	0.589	-5.9	102	0.00
23 S	Nitrobenzene-d5	0.453	0.495	-9.3	105	0.00
24	Nitrobenzene	0.497	0.537	-8.0	105	0.00
25	Isophorone	0.820	0.895	-9.1	104	0.00
26 C	2-Nitrophenol	0.190	0.197	-3.7	99	0.00
27	2,4-Dimethylphenol	0.312	0.337	-8.0	103	0.00
28	bis(2-Chloroethoxy)methane	0.426	0.443	-4.0	106	0.00
29 C	2,4-Dichlorophenol	0.334	0.368	-10.2	106	0.00
30	1,2,4-Trichlorobenzene	0.404	0.415	-2.7	102	0.00
31	Naphthalene	0.999	1.060	-6.1	103	0.00
32	Benzoic acid	0.251	0.244	2.8	92	0.00
33	4-Chloroaniline	0.420	0.455	-8.3	101	0.00
34 C	Hexachlorobutadiene	0.329	0.358	-8.8	107	0.00
35	Caprolactam	0.126	0.124	1.6	96	0.00
36 C	4-Chloro-3-methylphenol	0.412	0.432	-4.9	100	0.00
37	2-Methylnaphthalene	0.740	0.808	-9.2	105	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
39	1,2,4,5-Tetrachlorobenzene	0.660	0.669	-1.4	101	0.00
40 P	Hexachlorocyclopentadiene	0.192	0.219	-14.1	105	0.00
41 S	2,4,6-Tribromophenol	0.322	0.307	4.7	93	0.00
42 C	2,4,6-Trichlorophenol	0.416	0.420	-1.0	100	0.00
43	2,4,5-Trichlorophenol	0.435	0.411	5.5	94	0.00
44 S	2-Fluorobiphenyl	1.235	1.264	-2.3	104	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG011017\
 Data File : BG025451.D
 Acq On : 10 Jan 2017 10:35
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 10 13:05:25 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 11:26:22 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.345	1.380	-2.6	103	0.00
46	2-Chloronaphthalene	1.051	1.078	-2.6	104	0.00
47	2-Nitroaniline	0.444	0.486	-9.5	106	0.00
48	Acenaphthylene	1.629	1.645	-1.0	102	0.00
49	Dimethylphthalate	1.458	1.470	-0.8	99	0.00
50	2,6-Dinitrotoluene	0.319	0.320	-0.3	100	0.00
51 C	Acenaphthene	1.065	1.083	-1.7	102	0.00
52	3-Nitroaniline	0.306	0.300	2.0	95	0.00
53 P	2,4-Dinitrophenol	0.183	0.185	-1.1	94	0.00
54	Dibenzofuran	1.684	1.753	-4.1	103	0.00
55 P	4-Nitrophenol	0.229	0.219	4.4	91	0.00
56	2,4-Dinitrotoluene	0.464	0.464	0.0	98	0.00
57	Fluorene	1.410	1.421	-0.8	101	0.00
58	2,3,4,6-Tetrachlorophenol	0.466	0.448	3.9	96	0.00
59	Diethylphthalate	1.529	1.544	-1.0	101	0.00
60	4-Chlorophenyl-phenylether	0.799	0.806	-0.9	101	0.00
61	4-Nitroaniline	0.308	0.316	-2.6	92	0.00
62	Azobenzene	1.475	1.535	-4.1	102	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
64	4,6-Dinitro-2-methylphenol	0.139	0.147	-5.8	92	0.00
65 c	n-Nitrosodiphenylamine	0.541	0.552	-2.0	97	0.00
66	4-Bromophenyl-phenylether	0.247	0.259	-4.9	99	0.00
67	Hexachlorobenzene	0.283	0.286	-1.1	98	0.00
68	Atrazine	0.237	0.210	11.4	84	0.00
69 C	Pentachlorophenol	0.147	0.137	6.8	85	0.00
70	Phenanthrene	1.031	1.063	-3.1	99	0.00
71	Anthracene	1.047	1.084	-3.5	100	0.00
72	Carbazole	0.937	0.962	-2.7	98	0.00
73	Di-n-butylphthalate	1.152	1.169	-1.5	98	0.00
74 C	Fluoranthene	1.361	1.403	-3.1	100	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
76	Benzidine	0.499	0.422	15.4	87	0.00
77	Pyrene	1.045	1.057	-1.1	102	0.00
78 S	Terphenyl-d14	0.754	0.736	2.4	100	0.00
79	Butylbenzylphthalate	0.420	0.436	-3.8	109	0.00
80	Benzo(a)anthracene	1.116	1.160	-3.9	106	0.00
81	3,3'-Dichlorobenzidine	0.433	0.449	-3.7	106	0.00
82	Chrysene	1.070	1.103	-3.1	107	0.00
83	Bis(2-ethylhexyl)phthalate	0.584	0.623	-6.7	112	0.00
84 c	Di-n-octyl phthalate	0.970	1.052	-8.5	111	0.00
85	Indeno(1,2,3-cd)pyrene	1.360	1.491	-9.6	112	0.00
86 I	Perylene-d12	1.000	1.000	0.0	114	0.00
87	Benzo(b)fluoranthene	1.091	1.130	-3.6	112	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG011017\
Data File : BG025451.D
Acq On : 10 Jan 2017 10:35
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 10 13:05:25 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 10 11:26:22 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.054	1.081	-2.6	109	0.00
89 C	Benzo(a)pyrene	1.054	1.099	-4.3	112	0.00
90	Dibenzo(a,h)anthracene	1.117	1.173	-5.0	112	0.00
91	Benzo(a,h,i)perylene	1.110	1.168	-5.2	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0