


Data Path : Z:\HPCHEM1\BNA G\DATA\BG011117\  
 Data File : BG025470.D  
 Acq On : 11 Jan 2017 12:42  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG122116.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Jan 10 11:26:22 2017



AutoFind: Scans 2572, 2573, 2574; Background Corrected with Scan 2565

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 40.4      | 52144   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.5       | 255     | PASS             |
| 69          | 198          | 0.00         | 100          | 38.5      | 49679   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 41.3      | 53304   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 128957  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 8652    | PASS             |
| 275         | 198          | 10           | 60           | 32.4      | 41794   | PASS             |
| 365         | 198          | 1            | 100          | 4.6       | 5951    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.1      | 16713   | PASS             |
| 442         | 198          | 50           | 100          | 85.8      | 110586  | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 21662   | PASS             |