

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011123\
 Data File : BG056259.D
 Acq On : 11 Jan 2023 20:46
 Operator : CG/JU
 Sample : 01073-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 1524

Quant Time: Jan 12 04:23:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 12 03:38:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.317	152	42750	20.000	ng	0.00
21) Naphthalene-d8	11.143	136	156025	20.000	ng	0.00
39) Acenaphthene-d10	14.927	164	130438	20.000	ng	0.00
64) Phenanthrene-d10	17.659	188	339982	20.000	ng	0.00
76) Chrysene-d12	21.954	240	331174	20.000	ng	# 0.00
86) Perylene-d12	25.391	264	363339	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.831	112	157571	65.811	ng	0.00
7) Phenol-d6	7.447	99	295953	80.489	ng	0.00
23) Nitrobenzene-d5	9.486	82	184684	60.547	ng	0.00
42) 2,4,6-Tribromophenol	16.401	330	68981	41.757	ng	0.00
45) 2-Fluorobiphenyl	13.552	172	596590	63.159	ng	0.00
79) Terphenyl-d14	20.238	244	1220205	65.991	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011123\
Data File : BG056259.D
Acq On : 11 Jan 2023 20:46
Operator : CG/JU
Sample : 01073-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
1524

Quant Time: Jan 12 04:23:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 03:38:40 2023
Response via : Initial Calibration

