

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011216\
 Data File : BG020691.D
 Acq On : 13 Jan 2016 00:22
 Operator : SJ/IZ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02022

Quant Time: Jan 13 03:58:09 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 13 02:39:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.495	0.460	7.1	88	0.00
3 S	1,4-Dioxane-d8	0.409	0.380	7.1	94	0.00
4	Benzaldehyde	1.173	1.174	-0.1	107	0.00
5 S	Phenol-d5	1.714	1.575	8.1	98	0.00
6	Phenol	1.843	1.732	6.0	101	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.031	0.985	4.5	99	0.00
8	Bis(2-Chloroethyl)ether	1.356	1.320	2.7	103	0.00
9 S	2-Chlorophenol-d4	1.344	1.292	3.9	99	0.00
10	2-Chlorophenol	1.410	1.340	5.0	96	0.00
11	2-Methylphenol	1.419	1.338	5.7	101	0.00
12	2,2'-oxybis(1-Chloropropane	1.757	1.731	1.5	102	0.00
13 S	4-Methylphenol-d8	1.446	1.374	5.0	103	0.00
14	Acetophenone	2.385	2.376	0.4	104	0.00
15 P	N-Nitroso-di-n-propylamine	1.210	1.194	1.3	104	0.00
16	4-Methylphenol	1.574	1.511	4.0	101	0.00
17	Hexachloroethane	0.599	0.587	2.0	104	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
19 S	Nitrobenzene-d5	0.158	0.158	0.0	100	0.00
20	Nitrobenzene	0.410	0.413	-0.7	102	0.00
21	Isophorone	0.802	0.812	-1.2	104	0.00
22 S	2-Nitrophenol-d4	0.183	0.180	1.6	105	0.00
23 C	2-Nitrophenol	0.203	0.201	1.0	101	0.00
24	2,4-Dimethylphenol	0.427	0.440	-3.0	103	0.00
25	Bis(2-Chloroethoxy)methane	0.449	0.456	-1.6	106	0.00
26 S	2,4-Dichlorophenol-d3	0.349	0.359	-2.9	102	0.00
27 C	2,4-Dichlorophenol	0.370	0.372	-0.5	101	0.00
28	Naphthalene	1.130	1.129	0.1	104	0.00
29 S	4-Chloroaniline-d4	0.345	0.352	-2.0	109	0.00
30	4-Chloroaniline	0.371	0.373	-0.5	102	0.00
31 C	Hexachlorobutadiene	0.315	0.312	1.0	103	0.00
32	Caprolactam	0.126	0.129	-2.4	105	0.00
33 C	4-Chloro-3-methylphenol	0.400	0.413	-3.2	103	0.00
34	2-Methylnaphthalene	0.843	0.855	-1.4	106	0.00
35	1-Methylnaphthalene	0.796	0.802	-0.8	104	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
37	1,2,4,5-Tetrachlorobenzene	0.818	0.846	-3.4	109	0.00
38	Hexachlorocyclopentadiene	0.176	0.091	48.3#	104	0.00
39 C	2,4,6-Trichlorophenol	0.446	0.442	0.9	103	0.00
40	2,4,5-Trichlorophenol	0.490	0.516	-5.3	114	0.00
41	1,1'-Biphenyl	1.600	1.640	-2.5	106	0.00
42	2-Chloronaphthalene	1.303	1.318	-1.2	103	0.00
43	2-Nitroaniline	0.375	0.381	-1.6	101	0.00
44 S	Dimethylphthalate-d6	1.745	1.776	-1.8	104	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011216\
 Data File : BG020691.D
 Acq On : 13 Jan 2016 00:22
 Operator : SJ/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02022

Quant Time: Jan 13 03:58:09 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 13 02:39:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.875	1.883	-0.4	102	0.00
46	2,6-Dinitrotoluene	0.367	0.369	-0.5	105	0.00
47 S	Acenaphthylene-d8	1.960	1.985	-1.3	105	0.00
48	Acenaphthylene	2.069	2.110	-2.0	105	0.00
49	3-Nitroaniline	0.325	0.334	-2.8	105	0.00
50 C	Acenaphthene	1.428	1.449	-1.5	105	0.00
51	2,4-Dinitrophenol	0.095	0.059	37.9#	136	0.00
52 S	4-Nitrophenol-d4	0.234	0.243	-3.8	109	0.00
53	4-Nitrophenol	0.216	0.203	6.0	100	0.00
54	Dibenzofuran	2.149	2.215	-3.1	106	0.00
55	2,4-Dinitrotoluene	0.552	0.560	-1.4	102	0.00
56	2,3,4,6-Tetrachlorophenol	0.436	0.443	-1.6	103	0.00
57	Diethylphthalate	1.903	1.909	-0.3	101	0.00
58 S	Fluorene-d10	1.556	1.573	-1.1	103	0.00
59	Fluorene	1.753	1.803	-2.9	107	0.00
60	4-Chlorophenyl-phenylether	1.018	1.030	-1.2	105	0.00
61	4-Nitroaniline	0.377	0.405	-7.4	103	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.109	0.104	4.6	110	0.00
64	4,6-Dinitro-2-methylphenol	0.117	0.109	6.8	106	0.00
65	N-Nitrosodiphenylamine	0.579	0.577	0.3	104	0.00
66	4-Bromophenyl-phenylether	0.253	0.250	1.2	107	0.00
67	Hexachlorobenzene	0.280	0.273	2.5	105	0.00
68	Atrazine	0.257	0.257	0.0	102	0.00
69 C	Pentachlorophenol	0.077	0.065	15.6	108	0.00
70	Phenanthrene	1.190	1.186	0.3	105	0.00
71 S	Anthracene-d10	0.980	0.987	-0.7	106	0.00
72	Anthracene	1.201	1.202	-0.1	103	0.00
73	1,2,3,4-Tetrachlorobenzene	0.288	0.279	3.1	105	0.00
74	Pentachlorobenzene	0.300	0.300	0.0	107	0.00
75	Carbazole	0.984	1.030	-4.7	106	0.00
76	Di-n-butylphthalate	1.204	1.213	-0.7	102	0.00
77 C	Fluoranthene	1.447	1.531	-5.8	105	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
79 S	Pyrene-d10	0.951	0.956	-0.5	103	0.00
80	Pyrene	1.268	1.286	-1.4	103	0.00
81	Butylbenzylphthalate	0.426	0.432	-1.4	104	0.00
82	3,3'-Dichlorobenzidine	0.395	0.393	0.5	100	0.00
83	Benzo(a)anthracene	1.248	1.270	-1.8	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.629	0.633	-0.6	103	0.00
85	Chrysene	1.180	1.186	-0.5	105	0.00
86 I	Perylene-d12	1.000	1.000	0.0	108	0.00
87	Di-n-octyl phthalate	1.133	1.149	-1.4	103	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011216\
Data File : BG020691.D
Acq On : 13 Jan 2016 00:22
Operator : SJ/IZ
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02022

Quant Time: Jan 13 03:58:09 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 13 02:39:33 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.288	1.255	2.6	101	0.00
89	Benzo(k)fluoranthene	1.272	1.303	-2.4	106	0.00
90 S	Benzo(a)pyrene-d12	1.066	1.060	0.6	103	0.00
91 C	Benzo(a)pyrene	1.240	1.252	-1.0	105	0.00
92	Indeno(1,2,3-cd)pyrene	1.463	1.456	0.5	104	0.00
93	Dibenzo(a,h)anthracene	1.205	1.217	-1.0	105	0.00
94	Benzo(g,h,i)perylene	1.207	1.204	0.2	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0