

Data Path : Z:\HPCHEM1\BNA G\DATA\BG011217\
 Data File : BG025480.D
 Acq On : 12 Jan 2017 13:35
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

sohil
 1/13/2017 3:06:28 PM

Quant Time: Jan 13 00:57:51 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 11:26:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	70640	20.00	ng	0.00
21) Naphthalene-d8	11.03	136	313095	20.00	ng	0.00
38) Acenaphthene-d10	14.83	164	249152	20.00	ng	0.00
63) Phenanthrene-d10	17.57	188	568584	20.00	ng	0.00
75) Chrysene-d12	21.87	240	815391	20.00	ng	0.00
86) Perylene-d12	25.27	264	899173	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.74	112	342824	88.21	ng	0.00
7) Phenol-d6	7.36	99	484621	88.54	ng	0.00
23) Nitrobenzene-d5	9.39	82	580773	81.95	ng	0.00
41) 2,4,6-Tribromophenol	16.32	330	322771	80.47	ng	0.00
44) 2-Fluorobiphenyl	13.45	172	1244299	80.85	ng	0.00
78) Terphenyl-d14	20.15	244	2347523	76.33	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.58	88	74200	44.75	ng	92
3) Pyridine	4.00	79	213704	44.46	ng	94
4) n-Nitrosodimethylamine	3.91	42	134001	46.97	ng	# 96
6) Aniline	7.53	93	324656	43.78	ng	100
8) 2-Chlorophenol	7.77	128	186780	42.60	ng	99
9) Benzaldehyde	7.34	77	153024	38.21	ng	96
10) Phenol	7.39	94	264745	43.63	ng	98
11) bis(2-Chloroethyl)ether	7.61	93	204877	43.82	ng	88
12) 1,3-Dichlorobenzene	8.09	146	229463	43.25	ng	98
13) 1,4-Dichlorobenzene	8.24	146	230835	41.98	ng	100
14) 1,2-Dichlorobenzene	8.55	146	219145	41.77	ng	95
15) Benzyl Alcohol	8.45	79	224289	42.92	ng	92
16) 2,2'-oxybis(1-Chloropropan	8.72	45	389545	44.99	ng	97
17) 2-Methylphenol	8.65	107	173156	43.46	ng	91
18) Hexachloroethane	9.28	117	92955	44.02	ng	# 86
19) n-Nitroso-di-n-propylamine	9.01	70	205907	44.54	ng	97
20) 3+4-Methylphenols	8.98	107	241505	43.80	ng	94
22) Acetophenone	9.04	105	350920	40.34	ng	# 96
24) Nitrobenzene	9.43	77	325150	41.83	ng	98
25) Isophorone	9.94	82	534191	41.63	ng	97
26) 2-Nitrophenol	10.15	139	123339	41.49	ng	93
27) 2,4-Dimethylphenol	10.19	122	199655	40.85	ng	98
28) bis(2-Chloroethoxy)methane	10.42	93	260732	39.13	ng	97
29) 2,4-Dichlorophenol	10.68	162	220720	42.20	ng	96
30) 1,2,4-Trichlorobenzene	10.89	180	251413	39.79	ng	99
31) Naphthalene	11.08	128	632382	40.45	ng	97
32) Benzoic acid	10.36	122	155933	39.67	ng	89
33) 4-Chloroaniline	11.20	127	279501	42.50	ng	96
34) Hexachlorobutadiene	11.33	225	207363	40.20	ng	98
35) Caprolactam	11.99	113	80099	40.76	ng	95
36) 4-Chloro-3-methylphenol	12.31	107	259668	40.22	ng	88
37) 2-Methylnaphthalene	12.67	142	477911	41.24	ng	97
39) 1,2,4,5-Tetrachlorobenzene	13.03	216	344113	41.83	ng	97
40) Hexachlorocyclopentadiene	13.00	237	110335	41.33	ng	91

Data Path : Z:\HPCHEM1\BNA G\DATA\BG011217\
 Data File : BG025480.D
 Acq On : 12 Jan 2017 13:35
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTDCCC040

Manual Integrations
 APPROVED

sohil
 1/13/2017 3:06:28 PM

Quant Time: Jan 13 00:57:51 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 11:26:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.28	196	211183	40.77	nq	98
43) 2,4,5-Trichlorophenol	13.36	196	219945	40.57	nq	94
45) 1,1'-Biphenyl	13.66	154	692042	41.30	nq	98
46) 2-Chloronaphthalene	13.71	162	528867	40.39	nq	99
47) 2-Nitroaniline	13.93	65	245027	44.34	nq	91
48) Acenaphthylene	14.56	152	822310	40.52	nq	98
49) Dimethylphthalate	14.27	163	743213	40.92	nq	100
50) 2,6-Dinitrotoluene	14.41	165	163939	41.31	nq	98
51) Acenaphthene	14.89	154	534736	40.29	nq	97
52) 3-Nitroaniline	14.76	138	154525	40.51	nq	# 92
53) 2,4-Dinitrophenol	14.98	184	97351	40.12	nq	90
54) Dibenzofuran	15.22	168	868663	41.40	nq	98
55) 4-Nitrophenol	15.08	139	117486	41.23	nq	# 88
56) 2,4-Dinitrotoluene	15.21	165	237534	41.11	nq	# 89
57) Fluorene	15.87	166	723972	41.21	nq	97
58) 2,3,4,6-Tetrachlorophenol	15.46	232	230756	39.73	nq	93
59) Diethylphthalate	15.62	149	770426	40.45	nq	99
60) 4-Chlorophenyl-phenylether	15.85	204	415623	41.75	nq	94
61) 4-Nitroaniline	15.92	138	160780	41.89	nq	# 80
62) Azobenzene	16.15	77	778793	42.40	nq	99
64) 4,6-Dinitro-2-methylphenol	15.98	198	176218	44.75	nq	88
65) n-Nitrosodiphenylamine	16.08	169	647091	42.10	nq	97
66) 4-Bromophenyl-phenylether	16.75	248	303800	43.31	nq	98
67) Hexachlorobenzene	16.87	284	329551	40.94	nq	93
68) Atrazine	17.01	200	223137	33.14	nq	97
69) Pentachlorophenol	17.23	266	159605	38.25	nq	95
70) Phenanthrene	17.61	178	1227439	41.89	nq	96
71) Anthracene	17.71	178	1241768	41.71	nq	99
72) Carbazole	17.99	167	1122528	42.15	nq	98
73) Di-n-butylphthalate	18.50	149	1356352	41.40	nq	# 96
74) Fluoranthene	19.62	202	1635936	42.27	nq	94
76) Benzidine	19.79	184	718619	35.33	nq	96
77) Pyrene	19.98	202	1699105	39.87	nq	98
79) Butylbenzylphthalate	20.83	149	692396	40.47	nq	98
80) Benzo(a)anthracene	21.85	228	1855455	40.80	nq	98
81) 3,3'-Dichlorobenzidine	21.75	252	752464	42.60	nq	97
82) Chrysene	21.92	228	1781525	40.85	nq	99
83) Bis(2-ethylhexyl)phthalate	21.69	149	1003051	42.12	nq	99
84) Di-n-octyl phthalate	22.94	149	1744268	44.12	nq	96
85) Indeno(1,2,3-cd)pyrene	29.20	276	2462074	44.41	nq	99
87) Benzo(b)fluoranthene	24.18	252	2044190	41.67	nq	99
88) Benzo(k)fluoranthene	24.25	252	1949436m	41.15	nq	
89) Benzo(a)pyrene	25.11	252	1970784	41.59	nq	# 98
90) Dibenzo(a,h)anthracene	29.24	278	2110592	42.03	nq	# 98
91) Benzo(g,h,i)perylene	30.44	276	2092469	41.93	nq	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

