

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG011218\
 Data File : BG031996.D
 Acq On : 12 Jan 2018 11:11
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 1/15/2018 4:23:35 PM

Quant Time: Jan 12 17:00:07 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 12 16:45:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.78	152	49981	20.00	ng	0.00
21) Naphthalene-d8	11.66	136	199403	20.00	ng	0.00
38) Acenaphthene-d10	15.40	164	135489	20.00	ng	0.00
63) Phenanthrene-d10	18.14	188	336284	20.00	ng	0.00
75) Chrysene-d12	22.48	240	389825	20.00	ng	0.00
86) Perylene-d12	26.20	264	403503	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	6.20	112	14590	5.34	ng	0.00
7) Phenol-d6	7.87	99	17422	5.06	ng	0.00
23) Nitrobenzene-d5	9.96	82	15011	5.09	ng	0.00
41) 2,4,6-Tribromophenol	16.88	330	10983	4.90	ng	0.00
44) 2-Fluorobiphenyl	14.03	172	62364	5.71	ng	0.00
78) Terphenyl-d14	20.67	244	108299	6.13	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.86	88	3048	2.903	ng	# 58
3) Pyridine	4.32	79	6932	2.474	ng	# 70
4) n-Nitrosodimethylamine	4.22	42	2414	2.452	ng	# 79
6) Aniline	8.06	93	10840	2.497	ng	96
8) 2-Chlorophenol	8.31	128	8442	2.761	ng	# 70
9) Benzaldehyde	7.88	77	5281	2.648	ng	81
10) Phenol	7.90	94	8923	2.632	ng	95
11) bis(2-Chloroethyl)ether	8.15	93	7271	2.677	ng	# 71
12) 1,3-Dichlorobenzene	8.65	146	10524m	2.630	ng	
13) 1,4-Dichlorobenzene	8.81	146	10026	2.488	ng	94
14) 1,2-Dichlorobenzene	9.14	146	10826	2.778	ng	96
15) Benzyl Alcohol	9.01	79	5624	2.249	ng	# 74
16) 2,2'-oxybis(1-Chloropropan	9.31	45	7101	2.572	ng	93
17) 2-Methylphenol	9.20	107	6807	2.627	ng	# 87
18) Hexachloroethane	9.91	117	3274	2.644	ng	96
19) n-Nitroso-di-n-propylamine	9.58	70	5399	2.528	ng	# 77
20) 3+4-Methylphenols	9.54	107	9385	2.615	ng	92
22) Acetophenone	9.61	105	11335	2.503	ng	# 83
24) Nitrobenzene	10.01	77	7654	2.603	ng	# 68
25) Isophorone	10.54	82	14048	2.560	ng	# 85
26) 2-Nitrophenol	10.74	139	3946	2.265	ng	# 77
27) 2,4-Dimethylphenol	10.78	122	7105	2.570	ng	# 72
28) bis(2-Chloroethoxy)methane	11.02	93	8276	2.503	ng	93
29) 2,4-Dichlorophenol	11.28	162	8726	2.565	ng	95
30) 1,2,4-Trichlorobenzene	11.51	180	11186	2.546	ng	97
31) Naphthalene	11.71	128	25979	2.630	ng	# 93
33) 4-Chloroaniline	11.80	127	10750	2.538	ng	# 85
34) Hexachlorobutadiene	11.98	225	8602	2.726	ng	96
35) Caprolactam	12.51	113	2744	2.659	ng	# 44
36) 4-Chloro-3-methylphenol	12.86	107	8237	2.646	ng	# 96
37) 2-Methylnaphthalene	13.27	142	20442	2.607	ng	96
39) 1,2,4,5-Tetrachlorobenzene	13.62	216	16267	2.754	ng	# 97
42) 2,4,6-Trichlorophenol	13.85	196	8185	2.467	ng	96
43) 2,4,5-Trichlorophenol	13.92	196	9431	2.455	ng	# 88

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG011218\
 Data File : BG031996.D
 Acq On : 12 Jan 2018 11:11
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5


Manual Integrations
 APPROVED

Sohil
 1/15/2018 4:23:35 PM

Quant Time: Jan 12 17:00:07 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 12 16:45:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 1,1'-Biphenyl	14.25	154	31675	2.727	ng	93
46) 2-Chloronaphthalene	14.29	162	24193	2.677	ng	93
47) 2-Nitroaniline	14.48	65	3965	2.292	ng #	77
48) Acenaphthylene	15.13	152	36452	2.649	ng	95
49) Dimethylphthalate	14.83	163	31536	2.660	ng	99
50) 2,6-Dinitrotoluene	14.95	165	5026	2.042	ng #	63
51) Acenaphthene	15.46	154	23388	2.571	ng	92
52) 3-Nitroaniline	15.29	138	4813	2.165	ng #	73
54) Dibenzofuran	15.80	168	37489	2.762	ng	97
57) Fluorene	16.44	166	30798	2.767	ng	94
58) 2,3,4,6-Tetrachlorophenol	16.01	232	8449	2.378	ng #	84
59) Diethylphthalate	16.17	149	31130	2.708	ng	96
60) 4-Chlorophenyl-phenylether	16.42	204	18948	2.775	ng	90
62) Azobenzene	16.71	77	18876	2.624	ng	83
65) n-Nitrosodiphenylamine	16.63	169	26462	2.593	ng	93
66) 4-Bromophenyl-phenylether	17.31	248	12416	2.595	ng	94
67) Hexachlorobenzene	17.44	284	14625	2.763	ng #	88
68) Atrazine	17.55	200	11073	2.580	ng	98
70) Phenanthrene	18.18	178	52259	2.791	ng	97
71) Anthracene	18.27	178	51859m	2.777	ng	
72) Carbazole	18.52	167	42537	2.626	ng #	96
73) Di-n-butylphthalate	19.04	149	51188	2.674	ng	99
74) Fluoranthene	20.14	202	68001	2.901	ng	97
76) Benzidine	20.30	184	22549	2.313	ng #	93
77) Pyrene	20.50	202	71669	2.925	ng	97
79) Butylbenzylphthalate	21.36	149	23762	2.706	ng #	71
80) Benzo(a)anthracene	22.46	228	68098m	2.809	ng	
81) 3,3'-Dichlorobenzidine	22.34	252	21758	2.403	ng #	99
82) Chrysene	22.53	228	62731	2.694	ng	96
83) Bis(2-ethylhexyl)phthalate	22.30	149	35794	2.780	ng #	92
84) Di-n-octyl phthalate	23.69	149	54710	2.675	ng #	88
85) Indeno(1,2,3-cd)pyrene	30.52	276	72514m	2.545	ng	
87) Benzo(b)fluoranthene	25.00	252	67359m	2.762	ng	
88) Benzo(k)fluoranthene	25.08	252	62481	2.622	ng	98
89) Benzo(a)pyrene	26.03	252	61686	2.674	ng #	91
90) Dibenzo(a,h)anthracene	30.59	278	58756	2.643	ng #	90
91) Benzo(g,h,i)perylene	31.86	276	60017	2.203	ng #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.78 min Scan# 934
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
152	100		
150	143.1	126.6	189.8
115	49.6	53.0	79.4

Manual Integrations
APPROVED




Sohil
1/15/2018 4:23:35 PM




#2

1,4-Dioxane
Concen: 2.903 ng
RT: 3.86 min Scan# 97
Delta R.T. 0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
88	100		
58	53.5	79.6	119.4
43	17.8	27.4	41.0





#3

Pyridine

Concen: 2.474 ng

RT: 4.32 min Scan# 175

Delta R.T. 0.01 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	79	Resp:	6932
Ion	Ratio	Lower	Upper
79	100		
52	53.5	69.9	104.9#
51	32.7	34.0	51.0#

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#4

n-Nitrosodimethylamine

Concen: 2.452 ng




RT: 4.22 min Scan# 159




Delta R.T. 0.01 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion:	42	Resp:	2414
Ion	Ratio	Lower	Upper
42	100		
74	102.5	102.5	153.7#
44	17.8	18.9	28.3#





#5

2-Fluorophenol

Concen: 5.339 ng

RT: 6.20 min Scan# 495

Delta R.T. 0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion:	112	Resp:	14590
Ion Ratio	Lower	Upper	
112	100		
64	48.5	56.3	84.5#
63	33.0	30.1	45.1

Instrument :





BNA_G

ClientSampleId :

SSTDICC2.5

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#6

Aniline

Concen: 2.497 ng


RT: 8.06 min Scan# 812


Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion:	93	Resp:	10840
Ion Ratio	Lower	Upper	
93	100		
66	43.1	33.7	50.5
65	23.9	16.1	24.1





#7

Phenol-d6

Concen: 5.062 ng

RT: 7.87 min Scan# 780

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	99	Resp:	17422
Ion Ratio	Lower	Upper	
99	100		
42	15.7	14.1	21.1
71	33.8	26.1	39.1

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#8

2-Chlorophenol

Concen: 2.761 ng




RT: 8.31 min Scan# 855


Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion:	128	Resp:	8442
Ion Ratio	Lower	Upper	
128	100		
130	25.6	14.0	54.0
64	29.1	37.7	77.7#





#9

Benzaldehyde

Concen: 2.648 ng

RT: 7.88 min Scan# 781

Delta R.T. 0.01 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	77	Resp:	5281
Ion	Ratio	Lower	Upper
77	100		
105	110.0	71.3	111.3
106	100.9	64.2	104.2

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#10

Phenol

Concen: 2.632 ng




RT: 7.90 min Scan# 785

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion:	94	Resp:	8923
Ion	Ratio	Lower	Upper
94	100		
65	32.2	11.9	51.9
66	47.2	22.2	62.2









#11
bis(2-Chloroethyl)ether
Concen: 2.677 ng
RT: 8.15 min Scan# 828
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
93	100		
63	52.5	67.1	107.1#
95	35.0	11.5	51.5




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#12
1,3-Dichlorobenzene
Concen: 2.630 ng m
RT: 8.65 min Scan# 913
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
146	100		
148	70.5	51.4	77.2
75	23.2	26.6	39.8#






#13
 1,4-Dichlorobenzene
 Concen: 2.488 ng
 RT: 8.81 min Scan# 940
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5




Tgt Ion	Ratio	Lower	Upper
146	100		
148	60.6	53.7	80.5
111	41.8	35.2	52.8

Manual Integrations
 APPROVED

Sohil
 1/15/2018 4:23:35 PM




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#14
 1,2-Dichlorobenzene
 Concen: 2.778 ng
 RT: 9.14 min Scan# 996
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
146	100		
148	58.4	49.3	73.9
111	40.6	34.3	51.5



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 2.249 ng

RT: 9.01 min Scan# 973

Delta R.T. 0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
79	100		
108	97.6	57.2	85.8#
77	72.1	46.1	69.1#

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 2.572 ng




RT: 9.31 min Scan# 1024


Delta R.T. 0.01 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
45	100		
77	9.3	0.0	30.2
79	12.6	0.0	27.9









#17
2-Methylphenol
Concen: 2.627 ng
RT: 9.20 min Scan# 1006
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	6807		
108	126.9	83.9	125.9#	
77	48.2	37.2	55.8	
79	42.8	36.2	54.2	




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#18
Hexachloroethane
Concen: 2.644 ng
RT: 9.91 min Scan# 1126
Delta R.T. 0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	3274		
119	95.5	76.7	115.1	
201	127.3	95.7	143.5	









#19
n-Nitroso-di-n-propylamine
Concen: 2.528 ng
RT: 9.58 min Scan# 1071
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
70	100		
42	43.3	49.1	73.7#
101	7.5	8.5	12.7#
130	28.7	13.4	20.0#




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#20
3+4-Methylphenols
Concen: 2.615 ng
RT: 9.54 min Scan# 1064
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
107	100		
108	72.4	61.6	101.6
77	31.0	13.9	53.9
79	32.0	8.8	48.8





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 11.66 min Scan# 1424

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt	Ion	Ratio	Resp	Lower	Upper
136	100		199403		
137	11.8	9.2	13.8		
54	5.2	10.3	15.5#		
68	3.9	4.5	6.7#		

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#22

Acetophenone

Concen: 2.503 ng




RT: 9.61 min Scan# 1075


Delta R.T. -0.01 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt	Ion	Ratio	Resp	Lower	Upper
105	100		11335		
71	0.0	3.0	4.4#		
51	16.7	26.1	39.1#		
120	20.8	17.4	26.2		









#23
 Nitrobenzene-d5
 Concen: 5.093 ng
 RT: 9.96 min Scan# 1136
 Delta R.T. -0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampled :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	15011		
128	49.9	29.7	44.5#	
54	37.0	43.1	64.7#	




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#24
 Nitrobenzene
 Concen: 2.603 ng
 RT: 10.01 min Scan# 1144
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	7654		
123	67.2	33.0	49.6#	
65	12.0	13.0	19.6#	









#25
Isophorone
Concen: 2.560 ng
RT: 10.54 min Scan# 1234
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
82	100		
95	9.0	6.2	9.2
138	23.4	12.1	18.1




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#26
2-Nitrophenol
Concen: 2.265 ng
RT: 10.74 min Scan# 1268
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
139	100		
109	34.2	24.2	36.2
65	35.9	47.4	71.0









#27
 2,4-Dimethylphenol
 Concen: 2.570 ng
 RT: 10.78 min Scan# 1275
 Delta R.T. 0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
122	100		
107	83.2	100.2	150.2#
121	50.2	44.4	66.6




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#28
 bis(2-Chloroethoxy)methane
 Concen: 2.503 ng
 RT: 11.02 min Scan# 1315
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
93	100		
95	35.1	24.3	36.5
123	9.1	8.4	12.6









#29
 2,4-Dichlorophenol
 Concen: 2.565 ng
 RT: 11.28 min Scan# 1360
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
162	100		
164	67.9	48.0	88.0
98	33.1	21.1	61.1




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#30
 1,2,4-Trichlorobenzene
 Concen: 2.546 ng
 RT: 11.51 min Scan# 1399
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
180	100		
182	98.6	77.5	116.3
145	32.8	23.4	35.2









#31
Naphthalene
Concen: 2.630 ng
RT: 11.71 min Scan# 1433
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
128	100		
129	13.6	8.6	12.8#
127	17.3	11.6	17.4




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#33
4-Chloroaniline
Concen: 2.538 ng
RT: 11.80 min Scan# 1449
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
127	100		
129	36.0	24.0	36.0#
65	20.5	27.0	40.4#
92	24.8	16.6	25.0









#34
Hexachlorobutadiene
Concen: 2.726 ng
RT: 11.98 min Scan# 1479
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
225	100		
223	60.2	49.7	74.5
227	64.9	48.1	72.1




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#35
Caprolactam
Concen: 2.659 ng
RT: 12.51 min Scan# 1569
Delta R.T. -0.04 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
113	100		
55	115.3	186.9	226.9#
56	84.6	130.9	170.9#









#36
 4-Chloro-3-methylphenol
 Concen: 2.646 ng
 RT: 12.86 min Scan# 1629
 Delta R.T. -0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	8237		
144	28.5	18.2	27.4#	
142	71.8	59.0	88.4	




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#37
 2-Methylnaphthalene
 Concen: 2.607 ng
 RT: 13.27 min Scan# 1698
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	20442		
141	81.6	67.1	100.7	
115	33.7	30.7	46.1	









#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 15.40 min Scan# 2062
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.7	78.8	118.2
160	44.3	35.4	53.0




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 2.754 ng
RT: 13.62 min Scan# 1759
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
216	100		
214	77.3	63.0	94.4
179	21.3	17.0	25.6
108	8.8	12.8	19.2#









#41
 2,4,6-Tribromophenol
 Concen: 4.899 ng
 RT: 16.88 min Scan# 2313
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	10983		
332	91.7	77.0	115.6	
141	27.5	25.2	37.8	




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#42
 2,4,6-Trichlorophenol
 Concen: 2.467 ng
 RT: 13.85 min Scan# 1797
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	8185		
198	94.8	78.2	117.2	
200	27.7	24.6	36.8	









#43
 2,4,5-Trichlorophenol
 Concen: 2.455 ng
 RT: 13.92 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	9431		
198	88.8	76.2	114.4	
97	30.3	38.7	58.1	
132	22.1	20.2	30.2	




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#44
 2-Fluorobiphenyl
 Concen: 5.707 ng
 RT: 14.03 min Scan# 1828
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	62364		
171	35.1	30.0	45.0	
170	22.1	19.5	29.3	






#45
 1,1'-Biphenyl
 Concen: 2.727 ng
 RT: 14.25 min Scan# 1865
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampled :
 SSTDICC2.5




Tgt Ion	Ratio	Lower	Upper
154	100		
153	35.9	19.8	59.8
76	7.9	0.0	31.9

Manual Integrations
 APPROVED

Sohil
 1/15/2018 4:23:35 PM




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): B






#46
 2-Chloronaphthalene
 Concen: 2.677 ng
 RT: 14.29 min Scan# 1873
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
162	100		
127	33.7	30.7	46.1
164	35.5	26.0	39.0



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 2.292 ng
 RT: 14.48 min Scan# 1904
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion: 65 Resp: 3965
 Ion Ratio Lower Upper
 65 100
 92 51.9 54.6 82.0#
 138 114.6 72.9 109.3#




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#48
 Acenaphthylene
 Concen: 2.649 ng
 RT: 15.13 min Scan# 2015
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion: 152 Resp: 36452
 Ion Ratio Lower Upper
 152 100
 151 18.9 17.2 25.8
 153 11.4 10.2 15.4









#49
Dimethylphthalate
Concen: 2.660 ng
RT: 14.83 min Scan# 1964
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	31536		
194	4.5		3.4	5.2
164	9.7		8.2	12.4




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#50
2,6-Dinitrotoluene
Concen: 2.042 ng
RT: 14.95 min Scan# 1985
Delta R.T. -0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	5026		
63	39.7		52.7	79.1#
89	30.0		49.2	73.8#









#51
Acenaphthene
Concen: 2.571 ng
RT: 15.46 min Scan# 2072
Delta R.T. -0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	23388		
153	104.1	90.4	135.6	
152	47.8	41.6	62.4	




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#52
3-Nitroaniline
Concen: 2.165 ng
RT: 15.29 min Scan# 2043
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	4813		
108	14.5	17.5	26.3#	
92	98.3	105.8	158.8#	









#54
Dibenzenofuran
Concen: 2.762 ng
RT: 15.80 min Scan# 2129
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
168	100		
139	33.9	28.6	43.0
169	12.4	11.2	16.8




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#57
Fluorene
Concen: 2.767 ng
RT: 16.44 min Scan# 2238
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
166	100		
165	94.5	80.6	121.0
167	12.7	10.4	15.6









#58
2,3,4,6-Tetrachlorophenol
Concen: 2.378 ng
RT: 16.01 min Scan# 2165
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
232	100		
131	31.0	33.5	50.3#
130	0.7	2.2	3.2#
166	22.3	24.8	37.2#




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#59
Diethylphthalate
Concen: 2.708 ng
RT: 16.17 min Scan# 2192
Delta R.T. -0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
149	100		
177	26.0	18.5	27.7
150	12.3	10.2	15.2









#60
 4-Chlorophenyl-phenylether
 Concen: 2.775 ng
 RT: 16.42 min Scan# 2234
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
204	100		
206	34.3	26.2	39.2
141	46.9	45.8	68.6




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#62
 Azobenzene
 Concen: 2.624 ng
 RT: 16.71 min Scan# 2284
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
77	100		
182	43.7	7.4	47.4
105	24.4	0.3	40.3
51	25.8	11.6	51.6





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 18.14 min Scan# 2527

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	336284		
94	5.9	6.9	10.3#	
80	5.4	7.7	11.5#	

Manual Integrations
APPROVED




Sohil
1/15/2018 4:23:35 PM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#65

n-Nitrosodiphenylamine

Concen: 2.593 ng


RT: 16.63 min Scan# 2270

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11



Tgt Ion	Ratio	Resp	Lower	Upper
169	100	26462		
168	75.1	55.7	83.5	
167	32.6	30.1	45.1	




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#66
4-Bromophenyl-phenylether
Concen: 2.595 ng
RT: 17.31 min Scan# 2387
Delta R.T. 0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
248	100		
250	90.5	77.1	115.7
141	59.0	50.8	76.2




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#67
Hexachlorobenzene
Concen: 2.763 ng
RT: 17.44 min Scan# 2408
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
284	100		
142	20.3	26.8	40.2#
249	32.0	26.3	39.5









#68
Atrazine
Concen: 2.580 ng
RT: 17.55 min Scan# 2427
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.3	5.2	45.2
215	52.0	30.4	70.4




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#70
Phenanthrene
Concen: 2.791 ng
RT: 18.18 min Scan# 2534
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
178	100		
176	17.0	15.7	23.5
179	15.7	12.5	18.7









#71
 Anthracene
 Concen: 2.777 ng m
 RT: 18.27 min Scan# 2550
 Delta R.T. 0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
178	100		
176	15.5	16.0	24.0#
179	16.2	12.5	18.7




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#72
 Carbazole
 Concen: 2.626 ng
 RT: 18.52 min Scan# 2593
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.7	18.2	27.4
139	9.1	9.4	14.2#









#73
Di-n-butylphthalate
Concen: 2.674 ng
RT: 19.04 min Scan# 2680
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	51188		
150	9.2	7.8	11.6	
104	5.0	3.9	5.9	




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#74
Fluoranthene
Concen: 2.901 ng
RT: 20.14 min Scan# 2868
Delta R.T. -0.00 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	68001		
101	6.7	0.0	28.9	
203	16.6	0.0	37.5	





#75

Chrysene-d12

Concen: 20.000 ng

RT: 22.48 min Scan# 3266

Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	389825		
120	5.6	6.8	10.2#	
236	25.6	20.9	31.3	

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#76

Benzidine

Concen: 2.313 ng




RT: 20.30 min Scan# 2895


Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	22549		
185	16.9	11.1	16.7#	
183	10.5	10.6	16.0#	









#77
 Pyrene
 Concen: 2.925 ng
 RT: 20.50 min Scan# 2929
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
202	100		
200	19.6	16.9	25.3
203	18.2	13.9	20.9




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#78
 Terphenyl-d14
 Concen: 6.134 ng
 RT: 20.67 min Scan# 2958
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.8	8.8
122	6.5	7.0	10.4#





#79

Butylbenzylphthalate

Concen: 2.706 ng

RT: 21.36 min Scan# 3076

Delta R.T. 0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt	Ion	Ratio	Lower	Upper
149	100			
91	49.7	62.2	93.2#	
206	31.0	18.6	27.8#	

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#80

Benzo(a)anthracene

Concen: 2.809 ng m




RT: 22.46 min Scan# 3262


Delta R.T. -0.00 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt	Ion	Ratio	Lower	Upper
228	100			
226	22.4	22.7	34.1#	
229	21.3	16.2	24.2	









#81
 3,3'-Dichlorobenzidine
 Concen: 2.403 ng
 RT: 22.34 min Scan# 3243
 Delta R.T. -0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.7	50.8	76.2
126	7.0	7.4	11.2#




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#82
 Chrysene
 Concen: 2.694 ng
 RT: 22.53 min Scan# 3275
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.1	24.9	37.3
229	17.6	15.0	22.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 2.780 ng
 RT: 22.30 min Scan# 3236
 Delta R.T. -0.00 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
149	100		
167	34.5	24.3	36.5
279	8.1	4.0	6.0




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#84
 Di-n-octyl phthalate
 Concen: 2.675 ng
 RT: 23.69 min Scan# 3472
 Delta R.T. -0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.8	1.4	2.0
43	5.8	8.9	13.3









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 2.545 ng m
 RT: 30.52 min Scan# 4635
 Delta R.T. -0.02 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
276	100		
138	8.3	16.0	24.0#
277	0.0	20.0	30.0#




Manual Integrations
 APPROVED


Sohil
 1/15/2018 4:23:35 PM



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 26.20 min Scan# 3900
 Delta R.T. -0.01 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.4	17.4	26.0
265	23.5	17.7	26.5





#87

Benzo(b)fluoranthene

Concen: 2.762 ng m

RT: 25.00 min Scan# 3695

Delta R.T. -0.02 min

Lab File: BG031996.D





Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
252	100		
253	18.6	18.2	27.4
125	5.6	7.2	10.8

Manual Integrations
APPROVED

Sohil
1/15/2018 4:23:35 PM



#88

Benzo(k)fluoranthene

Concen: 2.622 ng




RT: 25.08 min Scan# 3708


Delta R.T. -0.02 min

Lab File: BG031996.D

Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.9	17.4	26.2
125	7.2	6.6	9.8









#89
Benzo(a)pyrene
Concen: 2.674 ng
RT: 26.03 min Scan# 3870
Delta R.T. -0.01 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
252	100		
253	18.1	18.3	27.5#
125	6.6	7.3	10.9#




Manual Integrations
APPROVED


Sohil
1/15/2018 4:23:35 PM



#90
Dibenzo(a,h)anthracene
Concen: 2.643 ng
RT: 30.59 min Scan# 4646
Delta R.T. -0.03 min
Lab File: BG031996.D
Acq: 12 Jan 2018 11:11

Tgt Ion	Ratio	Lower	Upper
278	100		
139	7.1	9.8	14.6#
279	27.4	18.4	27.6





#91
 Benzo(g,h,i)perylene
 Concen: 2.203 ng
 RT: 31.86 min Scan# 4863
 Delta R.T. -0.04 min
 Lab File: BG031996.D
 Acq: 12 Jan 2018 11:11

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
276	100		
277	17.9	18.3	27.5#
138	10.1	13.9	20.9#

Manual Integrations
 APPROVED

Sohil
 1/15/2018 4:23:35 PM

