


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG011218\
 Data File : BG032023.D
 Acq On : 13 Jan 2018 4:11
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG011218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 15 10:05:15 2018



AutoFind: Scans 2590, 2591, 2592; Background Corrected with Scan 2583

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	17.9	13728	PASS
68	69	0.00	2	1.9	313	PASS
69	198	0.00	100	21.1	16133	PASS
70	69	0.00	2	0.4	60	PASS
127	198	10	80	32.1	24594	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	76541	PASS
199	198	5	9	7.1	5401	PASS
275	198	10	60	32.2	24610	PASS
365	198	1	100	4.1	3164	PASS
441	442	0.01	24	15.2	9455	PASS
442	198	50	100	81.1	62056	PASS
443	442	15	24	19.7	12207	PASS