

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011222\
 Data File : BG051941.D
 Acq On : 12 Jan 2022 12:11
 Operator : CG/JU
 Sample : N1100-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BGLN3

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
 Supervised By :mohammad ahmed 01/18/2022

Quant Time: Jan 12 23:50:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 14:43:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.246	152	23443	20.000	ng/ul	0.00
20) Naphthalene-d8	11.089	136	100355	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.890	164	75983	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.645	188	267106	20.000	ng/ul	0.00
79) Chrysene-d12	21.968	240	294616	20.000	ng/ul	# 0.01
88) Perylene-d12	25.469	264	303387	20.000	ng/ul	0.03

01/13/2022
 Supervised By :mohammad
 ahmed

01/18/2022

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.576	96	2155	3.285	ng/uL	0.00
4) Pyridine-d5	4.004	84	11515	5.646	ng/ul	0.00
7) Phenol-d5	7.382	99	48351	21.073	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.547	67	41521	28.507	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.770	132	34092	22.442	ng/ul	0.00
15) 4-Methylphenol-d8	8.945	113	54355	30.475	ng/ul	0.00
21) Nitrobenzene-d5	9.421	128	16882	21.289	ng/ul	0.00
24) 2-Nitrophenol-d4	10.149	143	19920	22.713	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.695	165	40400	23.567	ng/ul	0.00
31) 4-Chloroaniline-d4	11.206	131	52920	22.167	ng/ul	0.00
46) Dimethylphthalate-d6	14.279	166	183807	29.266	ng/ul	0.00
49) Acenaphthylene-d8	14.584	160	184510	24.321	ng/ul	0.00
54) 4-Nitrophenol-d4	15.066	143	40209	38.292	ng/ul	0.00
60) Fluorene-d10	15.882	176	148148	27.055	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.988	200	41011	24.338	ng/ul	0.00
73) Anthracene-d10	17.745	188	359389	28.363	ng/ul	0.00
81) Pyrene-d10	20.024	212	418311	24.671	ng/ul	0.01
92) Benzo(a)pyrene-d12	25.223	264	421955	26.419	ng/ul	0.03

Target Compounds				Qvalue		
26) 2,4-Dimethylphenol	10.290	107	15689m	7.502	ng/ul	
30) Naphthalene	11.136	128	232479	42.750	ng/ul#	98
36) 2-Methylnaphthalene	12.734	142	241020	62.221	ng/ul	96
37) 1-Methylnaphthalene	12.951	142	115353	29.016	ng/ul	97
43) 1,1'-Biphenyl	13.727	154	49476	8.435	ng/ul#	95
47) Dimethylphthalate	14.326	163	7790	1.258	ng/ul#	53
52) Acenaphthene	14.954	153	33087	6.661	ng/ul#	91
56) Dibenzofuran	15.283	168	48807	6.765	ng/ul#	81
59) Diethylphthalate	15.683	149	16475	2.579	ng/ul#	80
61) Fluorene	15.935	166	41238	6.860	ng/ul#	92
69) Hexachlorobenzene	16.952	284	5988	1.727	ng/ul#	59
71) Pentachlorophenol	17.292	266	4758	2.457	ng/ul#	71
72) Phenanthrene	17.686	178	125903	8.996	ng/ul#	94
74) Anthracene	17.780	178	28780	2.054	ng/ul#	83
77) Carbazole	18.038	167	25782	2.042	ng/ul#	74
78) Di-n-butylphthalate	18.585	149	31997	2.144	ng/ul#	81
80) Fluoranthene	19.689	202	55462	2.831	ng/ul#	59
82) Pyrene	20.047	202	60772	3.150	ng/ul#	93
83) Butylbenzylphthalate	20.923	149	48989	7.297	ng/ul#	88
84) 3,3'-Dichlorobenzidine	21.851	252	9526	1.401	ng/ul#	75
85) Benzo(a)anthracene	21.945	228	42794	2.209	ng/ul#	91
86) Bis(2-ethylhexyl)phtha...	21.863	149	6683120	680.650	ng/ul#	46
87) Chrysene	22.015	228	42439	2.275	ng/ul#	95
89) Di-n-octyl phthalate	23.143	149	857601	51.172	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011222\
 Data File : BG051941.D
 Acq On : 12 Jan 2022 12:11
 Operator : CG/JU
 Sample : N1100-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BGLN3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
 Supervised By :mohammad ahmed 01/18/2022

Quant Time: Jan 12 23:50:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 14:43:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
90) Benzo(b)fluoranthene	24.342	252	40549	2.017	ng/ul#	83	
91) Benzo(k)fluoranthene	24.418	252	33713	1.814	ng/ul#	85	
93) Benzo(a)pyrene	25.299	252	34667	1.822	ng/ul#	89	
94) Indeno(1,2,3-cd)pyrene	29.476	276	43643m	2.004	ng/ul		
95) Dibenzo(a,h)anthracene	29.552	278	36606m	1.987	ng/ul		
96) Benzo(g,h,i)perylene	30.715	276	35651m	1.947	ng/ul		

01/13/2022
 Supervised By :mohammad
 ahmed
 01/18/2022

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011222\
 Data File : BG051941.D
 Acq On : 12 Jan 2022 12:11
 Operator : CG/JU
 Sample : N1100-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

BGLN3

Manual Integrations


APPROVED


Reviewed By :Jagrut Upadhyay 01/13/2022
 Supervised By :mohammad ahmed 01/18/2022

01/13/2022

Supervised By :mohammad
 ahmed

01/18/2022







#26
2,4-Dimethylphenol
Concen: 7.502 ng/ul m
RT: 10.290 min Scan# 1158
Delta R.T. 0.058 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3

Tgt Ion:107 Resp: 15689
Ion Ratio Lower Upper
107 100
121 32.1 39.3 58.9
122 20.1 63.7 95.5



Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022






01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#30
Naphthalene
Concen: 42.750 ng/ul
RT: 11.136 min Scan# 1302
Delta R.T. -0.007 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11


Tgt Ion:128 Resp: 232479
Ion Ratio Lower Upper
128 100
129 12.8 8.5 12.7#
127 14.7 11.8 17.6





#36
2-Methylnaphthalene
Concen: 62.221 ng/ul
RT: 12.734 min Scan# 1574
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11


Instrument :
BNA_G
ClientSampleId :
BGLN3



Tgt Ion:142 Resp: 241020
Ion Ratio Lower Upper
142 100
141 87.5 73.3 109.9



Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022






01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#37
1-Methylnaphthalene
Concen: 29.016 ng/ul
RT: 12.951 min Scan# 1611
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Tgt Ion:142 Resp: 115353
Ion Ratio Lower Upper
142 100
141 94.8 77.8 116.8
116 4.4 3.3 4.9





#43
1,1'-Biphenyl
Concen: 8.435 ng/ul
RT: 13.727 min Scan# 1742
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :

BNA_G

ClientSampleId :

BGLN3

Tgt Ion:154 Resp: 49470
Ion Ratio Lower Upper
154 100
153 41.1 34.5 51.7
76 12.0 13.1 19.7

Manual Integrations





APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022

01/13/2022




Supervised By :mohammad
ahmed


01/18/2022



#47
Dimethylphthalate
Concen: 1.258 ng/ul
RT: 14.326 min Scan# 1845
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Tgt Ion:163 Resp: 7790
Ion Ratio Lower Upper
163 100
194 50.7 3.0 4.4#
164 15.2 8.2 12.4#









#52
Acenaphthene
Concen: 6.661 ng/u1
RT: 14.954 min Scan# 1952
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3

Tgt Ion:153 Resp: 3308
Ion Ratio Lower Upper
153 100
152 55.0 36.5 54.7
154 76.9 65.8 98.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022



#56
Dibenzofuran
Concen: 6.765 ng/u1
RT: 15.283 min Scan# 2008
Delta R.T. -0.001 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Tgt Ion:168 Resp: 48807
Ion Ratio Lower Upper
168 100
139 50.7 31.2 46.8#





#59

Diethylphthalate

Concen: 2.579 ng/ul

RT: 15.683 min Scan# 2077

Delta R.T. -0.001 min

Lab File: BG051941.D

Acq: 12 Jan 2022 12:11

Instrument :

BNA_G

ClientSampleId :

BGLN3

Tgt Ion:149 Resp: 16479

Ion	Ratio	Lower	Upper
149	100		
177	32.2	16.8	25.2
150	17.7	10.1	15.1

Manual Integrations





APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022

01/13/2022

Supervised By :mohammad
ahmed

01/18/2022



#61

Fluorene

Concen: 6.860 ng/ul

RT: 15.935 min Scan# 2119




Delta R.T. -0.001 min


Lab File: BG051941.D

Acq: 12 Jan 2022 12:11

Tgt Ion:166 Resp: 41238


Ion	Ratio	Lower	Upper
166	100		
165	109.0	81.0	121.6
167	18.6	11.0	16.4





#69
Hexachlorobenzene
Concen: 1.727 ng/u1
RT: 16.952 min Scan# 21
Delta R.T. 0.005 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3






Tgt Ion:284 Resp: 5988
Ion Ratio Lower Upper
284 100
142 76.5 31.9 47.9#
249 22.1 25.8 38.6#

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022




01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#71
Pentachlorophenol
Concen: 2.457 ng/u1
RT: 17.292 min Scan# 2350
Delta R.T. 0.005 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Tgt Ion:266 Resp: 4758
Ion Ratio Lower Upper
266 100
264 40.1 54.3 81.5#
268 45.4 51.0 76.6#







#72
Phenanthrene
Concen: 8.996 ng/u1
RT: 17.686 min Scan# 2416
Delta R.T. 0.005 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3

Tgt Ion:178 Resp: 125900
Ion Ratio Lower Upper
178 100
179 19.9 12.3 18.5
176 20.1 15.5 23.3



Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022






01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#74
Anthracene
Concen: 2.054 ng/u1
RT: 17.780 min Scan# 2433
Delta R.T. 0.005 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11


Tgt Ion:178 Resp: 28780
Ion Ratio Lower Upper
178 100
179 27.7 13.0 19.6#
176 23.5 15.6 23.4#





#77
 Carbazole
 Concen: 2.042 ng/ul
 RT: 18.038 min Scan# 2477
 Delta R.T. 0.005 min
 Lab File: BG051941.D
 Acq: 12 Jan 2022 12:11

Instrument :
 BNA_G
 ClientSampleId :
 BGLN3






Tgt Ion:167 Resp: 2578
 Ion Ratio Lower Upper
 167 100
 166 30.8 17.3 25.9
 139 27.6 10.6 16.0

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
 Supervised By :mohammad ahmed 01/18/2022




01/13/2022
 Supervised By :mohammad
 ahmed


01/18/2022



#78
 Di-n-butylphthalate
 Concen: 2.144 ng/ul
 RT: 18.585 min Scan# 2570
 Delta R.T. 0.005 min
 Lab File: BG051941.D
 Acq: 12 Jan 2022 12:11


Tgt Ion:149 Resp: 31997
 Ion Ratio Lower Upper
 149 100
 150 16.8 7.5 11.3#
 104 12.0 5.0 7.6#





#80
Fluoranthene
Concen: 2.831 ng/u1
RT: 19.689 min Scan# 21
Delta R.T. 0.011 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3






Tgt Ion:202 Resp: 5546
Ion Ratio Lower Upper
202 100
101 9.7 9.3 13.9
100 41.1 7.0 10.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022




01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#82
Pyrene
Concen: 3.150 ng/u1
RT: 20.047 min Scan# 2819
Delta R.T. 0.005 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Tgt Ion:202 Resp: 60772
Ion Ratio Lower Upper
202 100
101 9.9 10.7 16.1#
100 9.0 9.0 13.6#






#83
Butylbenzylphthalate
Concen: 7.297 ng/ul
RT: 20.923 min Scan# 2968
Delta R.T. 0.011 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :

BNA_G

ClientSampleId :

BGLN3



Tgt Ion:149 Resp: 4898
Ion Ratio Lower Upper
149 100
91 92.0 65.2 97.8
206 25.6 16.2 24.2

Manual Integrations




APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022


01/13/2022

Supervised By :mohammad
ahmed



01/18/2022




#84
3,3'-Dichlorobenzidine
Concen: 1.401 ng/ul
RT: 21.851 min Scan# 3126
Delta R.T. 0.016 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11




Tgt Ion:252 Resp: 9526
Ion Ratio Lower Upper
252 100
254 57.6 50.8 76.2
126 59.5 9.0 13.4





#85
Benzo(a)anthracene
Concen: 2.209 ng/u1
RT: 21.945 min Scan# 3142
Delta R.T. 0.011 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11



Instrument :
BNA_G
ClientSampleId :
BGLN3



Tgt Ion:228 Resp: 42794
Ion Ratio Lower Upper
228 100
229 28.5 15.7 23.5
226 28.5 22.3 33.5


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022




01/13/2022
Supervised By :mohammad
ahmed



01/18/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 680.650 ng/u1
RT: 21.863 min Scan# 3128
Delta R.T. 0.046 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11




Tgt Ion:149 Resp: 6683120
Ion Ratio Lower Upper
149 100
167 55.8 21.1 31.7#
279 11.7 2.7 4.1#





#87
Chrysene
Concen: 2.275 ng/ul
RT: 22.015 min Scan# 3153
Delta R.T. 0.011 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11



Instrument :
BNA_G
ClientSampleId :
BGLN3



Tgt Ion:228 Resp: 42439
Ion Ratio Lower Upper
228 100
226 30.4 24.8 37.2
229 25.3 15.8 23.6


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022






01/13/2022
Supervised By :mohammad
ahmed


01/18/2022



#89
Di-n-octyl phthalate
Concen: 51.172 ng/ul
RT: 23.143 min Scan# 3346
Delta R.T. 0.022 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11


Tgt Ion:149 Resp: 857601





#90
Benzo(b)fluoranthene
Concen: 2.017 ng/ul
RT: 24.342 min Scan# 3550
Delta R.T. 0.016 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3






Tgt Ion:252 Resp: 40549
Ion Ratio Lower Upper
252 100
253 25.4 17.8 26.6
125 25.4 8.5 12.7

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022

01/13/2022
Supervised By :mohammad
ahmed



01/18/2022




#91
Benzo(k)fluoranthene
Concen: 1.814 ng/ul
RT: 24.418 min Scan# 3563
Delta R.T. 0.022 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11




Tgt Ion:252 Resp: 33713
Ion Ratio Lower Upper
252 100
253 26.8 17.3 25.9#
125 18.9 7.8 11.6#





#93
Benzo(a)pyrene
Concen: 1.822 ng/ul
RT: 25.299 min Scan# 31
Delta R.T. 0.028 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11



Instrument :
BNA_G
ClientSampleId :
BGLN3



Tgt Ion:252 Resp: 3466
Ion Ratio Lower Upper
252 100
253 23.4 17.4 26.2
125 20.9 9.1 13.7


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022




01/13/2022
Supervised By :mohammad
ahmed



01/18/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 2.004 ng/ul m
RT: 29.476 min Scan# 4424
Delta R.T. 0.028 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11




Tgt Ion:276 Resp: 43643
Ion Ratio Lower Upper
276 100
138 15.0 15.5 23.3#
277 22.3 20.5 30.7





#95
Dibenzo(a,h)anthracene
Concen: 1.987 ng/ul m
RT: 29.552 min Scan# 4435
Delta R.T. 0.040 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11

Instrument :
BNA_G
ClientSampleId :
BGLN3






Tgt Ion:278 Resp: 36600
Ion Ratio Lower Upper
278 100
139 13.2 12.2 18.4
279 25.0 19.5 29.3

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/13/2022
Supervised By :mohammad ahmed 01/18/2022

01/13/2022
Supervised By :mohammad
ahmed

01/18/2022



#96
Benzo(g,h,i)perylene
Concen: 1.947 ng/ul m
RT: 30.715 min Scan# 4635
Delta R.T. 0.016 min
Lab File: BG051941.D
Acq: 12 Jan 2022 12:11



Tgt Ion:276 Resp: 35651
Ion Ratio Lower Upper
276 100
138 22.6 16.6 24.8
277 27.6 17.6 26.4#

