


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011222\
 Data File : BG051936.D
 Acq On : 12 Jan 2022 8:59
 Operator : CG/JU
 Sample : DFTPP452
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP452

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jan 06 14:43:02 2022



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.3	22492	PASS
68	69	0.00	2	1.4	337	PASS
69	198	0.00	100	43.4	23667	PASS
70	69	0.00	2	0.3	67	PASS
127	198	10	80	47.6	25933	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	54493	PASS
199	198	5	9	6.6	3608	PASS
275	198	10	60	29.9	16278	PASS
365	198	1	100	4.2	2300	PASS
441	198	0.01	100	12.1	6613	PASS
442	198	50	100	80.9	44101	PASS
443	442	15	24	18.7	8259	PASS

m/z	Abundance
40.05	173.0
44.15	588.0
73.15	152.0
167.25	185.0
207.05	501.0
281.15	173.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
39.15	999.0
40.05	325.0
44.15	681.0
49.15	251.0
50.15	3104.0
51.15	13549.0
52.15	605.0
55.25	152.0
56.15	335.0
57.05	930.0
61.05	244.0
63.15	487.0
64.15	236.0
65.25	172.0
68.15	358.0
69.05	14362.0
73.15	335.0
74.15	1956.0
75.05	2877.0
76.15	717.0
77.15	17072.0
78.15	1227.0
79.15	1049.0
80.15	1159.0
81.05	1274.0
82.05	173.0
83.25	201.0
85.15	239.0
86.15	504.0
87.05	204.0
91.05	366.0
92.25	389.0
93.05	2357.0
94.15	365.0
95.95	167.0
98.15	1604.0
99.15	1072.0
101.05	609.0
102.95	191.0
104.05	529.0
105.05	549.0
107.05	4777.0
108.05	1061.0
109.05	162.0
110.05	8835.0
111.05	1757.0
112.05	185.0
116.05	370.0
117.05	4514.0
118.15	296.0
122.05	285.0
123.15	766.0
124.05	314.0
124.95	219.0
127.15	16184.0
128.15	1307.0
129.05	7567.0
130.05	588.0
131.15	202.0
134.05	342.0
135.05	723.0
136.15	207.0
137.15	192.0
137.95	155.0
141.05	1304.0
142.15	391.0
143.05	388.0
145.85	281.0
147.05	617.0
148.05	1412.0
149.05	333.0
151.15	249.0
153.05	337.0
154.15	197.0
155.25	607.0
156.15	1106.0
156.95	187.0
158.05	216.0
159.85	252.0
160.25	165.0
161.15	561.0
165.15	433.0
165.95	224.0
167.05	1962.0
168.05	1277.0
171.05	154.0
172.05	165.0
173.15	350.0
174.15	311.0
175.05	987.0
176.15	165.0
176.95	503.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

179.05	1683.0
180.05	1183.0
180.85	447.0
185.05	975.0
186.05	5482.0
187.05	1542.0
188.25	209.0
189.15	530.0
191.05	248.0
192.15	384.0
193.05	867.0
195.95	1096.0
198.05	38840.0
199.05	2681.0
200.05	222.0
201.55	163.0
201.85	220.0
203.05	476.0
204.15	1893.0
205.05	2584.0
206.05	11414.0
207.05	1960.0
208.05	559.0
209.05	254.0
210.05	185.0
211.15	356.0
216.15	388.0
217.05	3009.0
218.15	310.0
221.05	1603.0
221.95	584.0
223.05	943.0
224.05	7122.0
225.05	1306.0
225.95	286.0
227.05	3063.0
227.95	381.0
229.05	697.0
230.05	156.0
230.95	254.0
234.05	249.0
235.05	325.0
237.05	270.0
239.95	181.0
241.05	228.0
241.95	347.0
242.95	236.0
244.05	5143.0
245.05	756.0
246.15	1380.0
247.15	319.0
248.95	218.0
253.25	238.0
255.05	30816.0
256.05	3628.0
257.05	396.0
257.95	1714.0
259.05	469.0
264.95	770.0
266.05	180.0
271.95	315.0
273.05	851.0
274.05	2659.0
275.05	11841.0
275.95	1741.0
276.95	1485.0
278.05	415.0
281.05	243.0
284.05	217.0
292.95	252.0
296.05	4702.0
297.05	815.0
303.05	414.0
303.95	184.0
314.05	188.0
314.95	659.0
316.15	258.0
323.05	1340.0
324.05	212.0
326.95	394.0
328.05	211.0
333.95	947.0
335.05	387.0
345.85	156.0
352.05	520.0
353.15	317.0
354.05	422.0
365.05	2002.0
366.15	302.0
371.05	168.0
371.95	603.0
372.95	154.0
382.95	180.0
392.05	181.0
401.95	395.0
402.95	404.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

421.05	544.0
422.05	409.0
422.95	2936.0
424.15	509.0
441.05	6144.0
442.05	39352.0
443.05	7156.0
444.05	766.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
38.05	336.0
39.15	1796.0
40.15	377.0
44.05	703.0
45.15	194.0
50.15	6925.0
51.15	25856.0
52.15	1066.0
56.05	772.0
57.05	1651.0
61.15	291.0
62.15	456.0
63.15	839.0
65.15	762.0
69.05	27744.0
70.05	202.0
73.15	219.0
74.15	2764.0
75.15	4752.0
76.25	1653.0
77.15	29568.0
78.15	1798.0
79.15	2141.0
80.15	1319.0
81.05	2735.0
82.15	806.0
83.15	398.0
85.05	282.0
86.05	547.0
87.15	338.0
88.05	178.0
91.15	491.0
92.05	695.0
93.15	3674.0
94.05	497.0
96.15	212.0
98.05	2481.0
99.05	2333.0
100.15	289.0
101.05	1019.0
103.15	203.0
103.95	720.0
105.05	899.0
105.95	273.0
107.05	9576.0
108.05	1334.0
110.05	14501.0
111.05	2389.0
112.05	313.0
114.05	152.0
116.05	514.0
117.15	8101.0
117.95	638.0
121.95	770.0
123.05	1176.0
124.05	301.0
125.05	417.0
127.05	29088.0
128.15	2286.0
129.05	12416.0
130.05	1355.0
131.05	262.0
134.05	430.0
135.05	1178.0
136.15	524.0
136.95	401.0
137.15	425.0
139.95	197.0
141.05	1777.0
142.05	564.0
143.25	383.0
145.95	241.0
146.95	823.0
147.95	2107.0
149.05	493.0
149.95	175.0
151.45	247.0
153.05	522.0
154.15	464.0
155.05	1485.0
156.05	1530.0
157.15	233.0
158.15	325.0
158.95	203.0
160.15	667.0
161.05	536.0
162.15	241.0
163.95	172.0
165.15	459.0
166.05	640.0
167.15	2918.0
167.95	1417.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

168.95	167.0
169.95	167.0
171.15	397.0
171.85	235.0
172.15	237.0
173.25	214.0
174.05	720.0
175.05	1271.0
177.05	904.0
177.95	167.0
179.05	2525.0
180.05	1738.0
181.05	789.0
185.05	1508.0
186.05	9374.0
187.05	2106.0
188.15	425.0
188.95	618.0
190.95	262.0
191.25	262.0
191.95	891.0
192.95	1067.0
194.05	323.0
196.05	2026.0
198.05	63144.0
198.95	4171.0
200.15	304.0
201.45	346.0
203.05	629.0
204.05	2630.0
205.05	4857.0
206.05	18352.0
207.05	3051.0
207.95	651.0
210.25	521.0
210.85	677.0
215.15	231.0
216.05	816.0
217.05	5188.0
218.05	768.0
221.15	3055.0
222.05	669.0
223.15	1203.0
224.05	11278.0
225.15	2677.0
226.05	365.0
227.05	4092.0
228.05	794.0
229.05	932.0
230.95	318.0
232.05	177.0
234.05	223.0
234.95	380.0
235.95	214.0
237.05	398.0
240.15	224.0
241.05	162.0
242.05	657.0
243.05	722.0
244.05	7810.0
245.05	1421.0
246.05	1876.0
246.95	628.0
249.05	266.0
252.95	193.0
253.55	273.0
255.05	46352.0
256.05	7923.0
256.95	485.0
258.05	2391.0
258.95	272.0
265.05	1507.0
265.75	163.0
269.95	154.0
271.05	186.0
271.75	156.0
272.95	1312.0
274.05	3038.0
275.05	19656.0
276.05	1812.0
276.95	1820.0
283.05	349.0
284.05	201.0
285.05	201.0
288.95	204.0
292.95	340.0
294.95	175.0
296.05	6033.0
297.05	1015.0
303.05	567.0
304.15	283.0
313.05	155.0
313.95	356.0
314.95	897.0
316.05	438.0
321.05	217.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

323.05	1741.0
323.95	246.0
326.95	376.0
328.05	166.0
333.95	1113.0
334.95	301.0
341.15	167.0
345.95	541.0
346.85	222.0
352.05	735.0
353.05	298.0
353.95	646.0
364.95	2678.0
365.95	360.0
371.05	175.0
371.35	197.0
372.05	809.0
372.95	363.0
382.95	482.0
389.85	202.0
401.95	350.0
402.85	549.0
404.05	166.0
420.95	489.0
421.95	494.0
423.05	2799.0
424.05	778.0
441.05	8202.0
442.05	51728.0
443.05	10002.0
444.05	1003.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

m/z	Abundance
38.15	273.0
39.15	2391.0
40.05	410.0
44.05	464.0
50.15	7239.0
51.15	28072.0
52.15	1394.0
55.05	296.0
56.05	861.0
57.15	1893.0
61.15	512.0
62.15	648.0
63.05	937.0
64.15	165.0
65.05	548.0
68.05	654.0
69.05	28896.0
73.25	323.0
74.15	3400.0
75.15	4687.0
76.15	1823.0
77.15	31240.0
78.15	2062.0
79.05	2155.0
80.05	1697.0
81.05	2731.0
82.15	382.0
83.25	305.0
85.05	370.0
86.15	784.0
87.15	335.0
91.05	642.0
92.05	461.0
93.05	4024.0
94.15	212.0
95.95	176.0
98.05	3182.0
99.05	2617.0
99.95	254.0
101.05	1532.0
103.15	440.0
104.05	644.0
105.05	738.0
106.05	327.0
107.15	10039.0
108.05	1651.0
110.05	16416.0
111.05	2383.0
112.05	335.0
116.15	714.0
117.05	8469.0
118.05	706.0
122.05	792.0
123.05	1199.0
124.05	517.0
125.05	528.0
127.05	32528.0
128.05	2206.0
129.05	13085.0
129.95	1032.0
131.05	278.0
132.15	219.0
134.15	495.0
135.15	813.0
136.05	406.0
137.15	724.0
137.95	171.0
141.05	1887.0
142.05	446.0
143.15	519.0
144.05	231.0
145.05	224.0
146.15	294.0
147.05	627.0
148.05	2263.0
149.05	341.0
151.15	377.0
152.95	657.0
154.15	424.0
155.15	992.0
156.05	1407.0
157.05	166.0
157.95	174.0
159.05	390.0
160.05	738.0
161.05	838.0
162.05	392.0
165.05	717.0
165.95	437.0
167.05	3247.0
168.05	1638.0
169.15	364.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

172.05	295.0
173.05	457.0
173.95	757.0
175.15	1207.0
176.15	239.0
177.15	866.0
177.85	286.0
179.05	2755.0
180.05	1874.0
180.95	1030.0
181.95	174.0
182.95	165.0
183.95	259.0
185.15	1408.0
186.05	9163.0
187.05	3159.0
188.15	407.0
189.05	790.0
190.95	185.0
192.05	1038.0
193.05	1105.0
193.75	177.0
196.05	2059.0
198.05	61496.0
199.05	3973.0
200.05	309.0
201.35	282.0
202.95	567.0
204.15	2469.0
205.05	3638.0
206.05	18344.0
207.05	2805.0
208.05	765.0
209.15	161.0
210.95	548.0
214.95	157.0
216.05	435.0
217.05	5389.0
217.95	789.0
220.15	272.0
221.05	2140.0
223.15	1149.0
224.05	9469.0
225.05	2858.0
226.15	432.0
227.05	4712.0
228.05	736.0
228.95	996.0
230.15	187.0
231.05	369.0
234.05	236.0
235.05	368.0
236.05	222.0
237.05	270.0
239.95	209.0
241.15	243.0
242.05	531.0
243.05	621.0
244.15	6925.0
245.05	932.0
246.05	1634.0
246.85	220.0
247.15	216.0
249.05	280.0
253.15	247.0
255.05	41072.0
256.05	6435.0
256.95	549.0
257.95	2547.0
259.05	540.0
259.85	155.0
265.15	1262.0
265.85	172.0
273.05	848.0
274.05	3646.0
275.05	17336.0
276.05	2278.0
277.05	1579.0
278.05	330.0
283.05	429.0
283.85	194.0
284.95	410.0
293.05	465.0
296.05	5174.0
297.05	622.0
298.05	159.0
302.95	615.0
303.95	295.0
310.05	161.0
313.35	153.0
313.95	180.0
314.95	547.0
315.95	452.0
320.95	295.0
323.05	1533.0
323.95	406.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452

326.95	310.0
333.15	440.0
334.05	1118.0
335.05	339.0
340.95	234.0
345.85	457.0
352.05	509.0
353.15	434.0
354.05	332.0
364.95	2220.0
366.05	190.0
371.95	752.0
383.05	178.0
390.05	169.0
401.95	516.0
402.95	731.0
403.95	157.0
404.25	175.0
420.95	461.0
421.95	378.0
422.95	2244.0
423.95	489.0
425.25	164.0
441.05	5493.0
442.05	41224.0
443.05	7618.0
444.05	906.0

Instrument :
BNA_G
ClientSampleId :
DFTPP452