


Data Path : Z:\HPCHEM1\BNA_G\Data\BG011416\
Data File : BG020728.D
Acq On : 14 Jan 2016 17:31
Operator : SJ/IZ
Sample : DFTPP23
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP23

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
Title : SVOA CALIBRATION
Last Update : Thu Jan 14 23:48:02 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.8	6588	PASS
68	69	0.00	2	0.7	51	PASS
69	198	0.00	100	29.8	7341	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.0	9835	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	24594	PASS
199	198	5	9	7.3	1793	PASS
275	198	10	60	29.3	7211	PASS
365	198	1	100	3.8	933	PASS
441	442	0.01	100	16.3	3541	PASS
442	198	50	100	88.3	21720	PASS
443	442	15	24	19.0	4133	PASS

m/z	Abundance
39.00	490.0
44.00	319.0
50.00	2215.0
51.00	7446.0
52.00	331.0
55.90	292.0
57.00	544.0
63.00	356.0
65.00	180.0
69.00	8264.0
73.90	881.0
75.00	1306.0
76.00	500.0
77.00	9809.0
78.00	753.0
79.00	589.0
80.00	405.0
81.00	753.0
81.90	301.0
86.10	199.0
90.90	185.0
92.00	184.0
92.90	1150.0
98.00	891.0
98.90	672.0
100.90	369.0
103.90	211.0
105.00	291.0
107.00	3125.0
107.90	501.0
110.00	5103.0
111.00	669.0
115.90	216.0
117.00	2485.0
121.90	158.0
123.00	272.0
125.00	171.0
127.00	10733.0
128.00	1116.0
129.00	4501.0
130.00	360.0
134.90	384.0
137.10	255.0
140.90	627.0
142.00	190.0
146.00	177.0
146.90	304.0
147.90	618.0
152.90	206.0
153.90	150.0
155.00	418.0
156.00	370.0
157.90	154.0
159.90	219.0
161.00	305.0
164.90	251.0
166.00	368.0
166.90	1212.0
168.00	634.0
171.90	170.0
174.00	432.0
174.90	475.0
176.90	287.0
178.90	1041.0
180.00	672.0
180.90	289.0
185.00	495.0
186.00	3920.0
187.00	1106.0
188.80	212.0
191.90	328.0
193.00	354.0
195.90	871.0
197.90	25000.0
198.90	1777.0
199.90	159.0
202.90	180.0
204.00	876.0
205.00	1610.0
206.00	7497.0
207.00	982.0
207.90	298.0
211.00	237.0
217.00	1946.0
218.00	213.0
221.00	1007.0
221.80	280.0
223.00	401.0
224.00	4013.0
225.00	1096.0
226.90	1822.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

227.90	332.0
228.90	315.0
231.00	245.0
241.00	161.0
242.00	253.0
242.90	268.0
244.00	3126.0
244.90	414.0
245.90	694.0
255.00	16880.0
256.00	2289.0
256.90	201.0
257.90	1024.0
264.90	536.0
273.00	516.0
273.90	1374.0
275.00	6915.0
276.00	1090.0
277.00	788.0
296.00	1870.0
296.90	205.0
303.00	311.0
314.90	184.0
323.00	629.0
334.00	398.0
351.90	221.0
364.90	726.0
366.00	159.0
371.90	192.0
402.90	154.0
421.00	151.0
422.00	157.0
423.00	1039.0
424.00	213.0
441.00	2761.0
442.00	16904.0
443.00	3296.0
444.00	385.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

m/z	Abundance
39.00	567.0
39.90	261.0
44.00	298.0
50.00	1604.0
51.00	7326.0
52.00	487.0
55.90	219.0
57.00	456.0
63.00	354.0
64.90	153.0
69.00	8356.0
74.00	1022.0
75.00	1460.0
76.00	525.0
77.00	9304.0
78.00	525.0
78.90	642.0
79.90	505.0
81.00	720.0
81.90	187.0
83.00	183.0
85.90	202.0
91.00	235.0
92.00	169.0
93.00	1313.0
98.00	933.0
99.00	630.0
101.00	413.0
103.90	313.0
104.90	285.0
106.90	2920.0
108.00	598.0
110.00	5999.0
110.90	897.0
116.00	220.0
117.00	2556.0
117.90	234.0
122.00	205.0
122.90	383.0
124.90	210.0
127.00	11114.0
128.00	928.0
129.00	4517.0
129.90	412.0
134.00	216.0
134.90	456.0
136.00	171.0
136.90	215.0
140.90	659.0
141.90	206.0
145.90	192.0
146.90	510.0
147.90	731.0
152.90	187.0
153.90	202.0
155.00	310.0
156.00	713.0
159.90	191.0
160.90	281.0
165.00	295.0
166.00	255.0
167.00	1399.0
167.90	595.0
172.90	185.0
174.00	362.0
175.00	658.0
176.90	263.0
178.90	1093.0
179.90	535.0
181.00	351.0
184.90	521.0
186.00	3870.0
187.00	1333.0
187.90	157.0
188.90	398.0
192.00	295.0
193.00	461.0
196.00	707.0
197.90	28920.0
198.90	1968.0
202.90	220.0
203.90	1039.0
205.00	1973.0
206.00	8425.0
207.00	1141.0
208.00	287.0
209.90	164.0
210.90	315.0
215.90	204.0
216.90	2000.0
217.90	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

221.00	1416.0
221.90	269.0
222.90	446.0
223.90	4395.0
225.00	1205.0
226.00	218.0
227.00	1945.0
228.00	334.0
228.90	418.0
231.00	246.0
233.90	156.0
234.90	179.0
241.90	224.0
243.00	290.0
244.00	3608.0
245.00	498.0
245.90	677.0
255.00	19576.0
255.90	2861.0
256.90	309.0
257.90	1172.0
258.90	234.0
265.00	400.0
272.90	752.0
274.00	1339.0
275.00	7972.0
275.90	1072.0
276.90	815.0
284.90	154.0
292.90	158.0
295.90	2345.0
296.90	443.0
303.00	378.0
314.00	208.0
314.90	211.0
315.90	154.0
323.00	896.0
324.00	173.0
327.00	170.0
334.00	451.0
346.00	170.0
351.90	214.0
353.90	275.0
364.90	1115.0
372.00	417.0
402.90	218.0
420.90	185.0
422.10	236.0
423.00	1265.0
424.00	322.0
441.00	4318.0
442.00	24968.0
443.00	4759.0
444.00	438.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

m/z	Abundance
39.10	369.0
40.00	198.0
44.00	269.0
50.00	1146.0
51.00	4992.0
52.00	248.0
55.90	162.0
57.00	321.0
63.00	315.0
67.90	155.0
68.90	5405.0
74.00	531.0
75.00	953.0
75.90	381.0
77.00	6172.0
78.00	421.0
79.00	466.0
79.90	274.0
80.90	509.0
83.00	177.0
92.00	166.0
93.00	752.0
98.00	467.0
98.90	395.0
100.90	340.0
103.90	204.0
104.90	185.0
107.00	2286.0
108.00	497.0
110.00	3798.0
110.90	725.0
116.00	156.0
116.90	1587.0
123.00	284.0
127.00	7658.0
128.00	660.0
129.00	2902.0
129.90	224.0
135.00	308.0
136.00	197.0
136.90	248.0
140.90	392.0
147.00	322.0
148.00	612.0
152.90	198.0
154.90	207.0
155.90	265.0
157.90	159.0
159.90	176.0
160.90	182.0
164.90	294.0
165.90	237.0
167.00	1056.0
167.90	485.0
174.00	161.0
175.00	492.0
177.00	278.0
178.90	922.0
180.00	458.0
181.00	189.0
185.00	413.0
186.00	3107.0
187.00	917.0
188.90	243.0
191.90	331.0
193.00	355.0
195.90	586.0
197.90	19864.0
198.90	1635.0
202.90	251.0
204.00	879.0
205.00	1412.0
206.00	6149.0
207.00	657.0
207.90	190.0
210.90	274.0
216.00	196.0
216.90	1752.0
217.90	167.0
221.00	934.0
223.00	451.0
224.00	3372.0
225.00	939.0
226.90	1631.0
227.90	194.0
228.90	238.0
237.00	155.0
241.90	273.0
242.90	165.0
244.00	2674.0
245.00	289.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

245.90	582.0
253.90	166.0
255.00	15727.0
256.00	1901.0
257.00	176.0
257.90	988.0
264.90	391.0
272.90	550.0
274.00	1335.0
275.00	6748.0
276.00	892.0
276.90	460.0
296.00	1873.0
296.90	280.0
303.00	262.0
314.90	326.0
316.00	151.0
323.00	650.0
324.00	215.0
333.90	360.0
346.00	260.0
352.00	220.0
353.10	179.0
354.00	196.0
364.90	958.0
371.90	336.0
402.90	204.0
422.10	203.0
423.00	1027.0
424.00	333.0
441.00	3545.0
442.00	23288.0
443.00	4345.0
444.00	392.0

Instrument :

BNA_G

ClientSampleId :

DFTPP23

m/z	Abundance
43.90	248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23