

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011516\
 Data File : BG020744.D
 Acq On : 15 Jan 2016 9:30
 Operator : SJ/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02030

Quant Time: Jan 15 22:13:08 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 14 23:48:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.495	0.473	4.4	95	0.00
3 S	1,4-Dioxane-d8	0.409	0.450	-10.0	117	0.00
4	Benzaldehyde	1.173	1.207	-2.9	115	0.00
5 S	Phenol-d5	1.714	1.634	4.7	106	0.00
6	Phenol	1.843	1.855	-0.7	114	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.031	1.049	-1.7	111	0.00
8	Bis(2-Chloroethyl)ether	1.356	1.397	-3.0	115	0.00
9 S	2-Chlorophenol-d4	1.344	1.406	-4.6	113	0.00
10	2-Chlorophenol	1.410	1.459	-3.5	110	0.00
11	2-Methylphenol	1.419	1.410	0.6	112	0.00
12	2,2'-oxybis(1-Chloropropane	1.757	1.817	-3.4	113	0.00
13 S	4-Methylphenol-d8	1.446	1.377	4.8	108	0.00
14	Acetophenone	2.385	2.439	-2.3	112	0.00
15 P	N-Nitroso-di-n-propylamine	1.210	1.213	-0.2	111	0.00
16	4-Methylphenol	1.574	1.572	0.1	110	0.00
17	Hexachloroethane	0.599	0.611	-2.0	113	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
19 S	Nitrobenzene-d5	0.158	0.167	-5.7	114	0.00
20	Nitrobenzene	0.410	0.432	-5.4	114	0.00
21	Isophorone	0.802	0.768	4.2	105	0.00
22 S	2-Nitrophenol-d4	0.183	0.186	-1.6	116	0.00
23 C	2-Nitrophenol	0.203	0.205	-1.0	110	0.00
24	2,4-Dimethylphenol	0.427	0.428	-0.2	107	0.00
25	Bis(2-Chloroethoxy)methane	0.449	0.449	0.0	112	0.00
26 S	2,4-Dichlorophenol-d3	0.349	0.365	-4.6	111	0.00
27 C	2,4-Dichlorophenol	0.370	0.387	-4.6	113	0.00
28	Naphthalene	1.130	1.177	-4.2	115	0.00
29 S	4-Chloroaniline-d4	0.345	0.361	-4.6	119	0.00
30	4-Chloroaniline	0.371	0.380	-2.4	111	0.00
31 C	Hexachlorobutadiene	0.315	0.315	0.0	111	0.00
32	Caprolactam	0.126	0.117	7.1	102	0.00
33 C	4-Chloro-3-methylphenol	0.400	0.401	-0.3	107	0.00
34	2-Methylnaphthalene	0.843	0.859	-1.9	114	0.00
35	1-Methylnaphthalene	0.796	0.805	-1.1	111	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
37	1,2,4,5-Tetrachlorobenzene	0.818	0.869	-6.2	114	0.00
38	Hexachlorocyclopentadiene	0.176	0.038	78.4#	44	0.00
39 C	2,4,6-Trichlorophenol	0.446	0.462	-3.6	110	0.00
40	2,4,5-Trichlorophenol	0.490	0.545	-11.2	122	0.00
41	1,1'-Biphenyl	1.600	1.664	-4.0	110	0.00
42	2-Chloronaphthalene	1.303	1.390	-6.7	111	0.00
43	2-Nitroaniline	0.375	0.405	-8.0	110	0.00
44 S	Dimethylphthalate-d6	1.745	1.672	4.2	99	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011516\
 Data File : BG020744.D
 Acq On : 15 Jan 2016 9:30
 Operator : SJ/IZ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02030

Quant Time: Jan 15 22:13:08 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 14 23:48:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.875	1.759	6.2	98	0.00
46	2,6-Dinitrotoluene	0.367	0.373	-1.6	108	0.00
47 S	Acenaphthylene-d8	1.960	2.026	-3.4	109	0.00
48	Acenaphthylene	2.069	2.131	-3.0	108	0.00
49	3-Nitroaniline	0.325	0.329	-1.2	105	0.00
50 C	Acenaphthene	1.428	1.457	-2.0	107	0.00
51	2,4-Dinitrophenol	0.095	0.085	10.5	202#	0.00
52 S	4-Nitrophenol-d4	0.234	0.231	1.3	105	0.00
53	4-Nitrophenol	0.216	0.215	0.5	108	0.00
54	Dibenzofuran	2.149	2.191	-2.0	107	0.00
55	2,4-Dinitrotoluene	0.552	0.558	-1.1	104	0.00
56	2,3,4,6-Tetrachlorophenol	0.436	0.409	6.2	97	0.00
57	Diethylphthalate	1.903	1.793	5.8	97	0.00
58 S	Fluorene-d10	1.556	1.579	-1.5	105	0.00
59	Fluorene	1.753	1.759	-0.3	106	0.00
60	4-Chlorophenyl-phenylether	1.018	1.027	-0.9	106	0.00
61	4-Nitroaniline	0.377	0.403	-6.9	104	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.109	0.104	4.6	104	0.00
64	4,6-Dinitro-2-methylphenol	0.117	0.119	-1.7	109	0.00
65	N-Nitrosodiphenylamine	0.579	0.596	-2.9	101	0.00
66	4-Bromophenyl-phenylether	0.253	0.256	-1.2	103	0.00
67	Hexachlorobenzene	0.280	0.284	-1.4	103	0.00
68	Atrazine	0.257	0.256	0.4	96	0.00
69 C	Pentachlorophenol	0.077	0.074	3.9	116	0.00
70	Phenanthrene	1.190	1.216	-2.2	101	0.00
71 S	Anthracene-d10	0.980	1.028	-4.9	104	0.00
72	Anthracene	1.201	1.235	-2.8	100	0.00
73	1,2,3,4-Tetrachlorobenzene	0.288	0.317	-10.1	112	0.00
74	Pentachlorobenzene	0.300	0.323	-7.7	109	0.00
75	Carbazole	0.984	1.040	-5.7	101	0.00
76	Di-n-butylphthalate	1.204	1.217	-1.1	97	0.00
77 C	Fluoranthene	1.447	1.528	-5.6	99	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
79 S	Pyrene-d10	0.951	0.966	-1.6	99	0.00
80	Pyrene	1.268	1.283	-1.2	98	0.00
81	Butylbenzylphthalate	0.426	0.448	-5.2	102	0.00
82	3,3'-Dichlorobenzidine	0.395	0.426	-7.8	103	0.00
83	Benzo(a)anthracene	1.248	1.304	-4.5	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.629	0.650	-3.3	100	0.00
85	Chrysene	1.180	1.228	-4.1	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Di-n-octyl phthalate	1.133	1.243	-9.7	104	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011516\
Data File : BG020744.D
Acq On : 15 Jan 2016 9:30
Operator : SJ/IZ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02030

Quant Time: Jan 15 22:13:08 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 14 23:48:02 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.288	1.350	-4.8	101	0.00
89	Benzo(k)fluoranthene	1.272	1.297	-2.0	98	0.00
90 S	Benzo(a)pyrene-d12	1.066	1.114	-4.5	100	0.00
91 C	Benzo(a)pyrene	1.240	1.308	-5.5	102	0.00
92	Indeno(1,2,3-cd)pyrene	1.463	1.481	-1.2	98	0.00
93	Dibenzo(a,h)anthracene	1.205	1.261	-4.6	100	0.00
94	Benzo(g,h,i)perylene	1.207	1.206	0.1	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0