


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG011516\
 Data File : BG020743.D
 Acq On : 15 Jan 2016 8:51
 Operator : SJ/IZ
 Sample : DFTPP24
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP24

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG011216.M
 Title : SVOA CALIBRATION
 Last Update : Sat Jan 16 00:43:13 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.9	7135	PASS
68	69	0.00	2	0.7	61	PASS
69	198	0.00	100	29.5	8140	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	38.5	10609	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	27586	PASS
199	198	5	9	6.7	1843	PASS
275	198	10	60	28.1	7749	PASS
365	198	1	100	3.7	1017	PASS
441	442	0.01	100	15.2	3662	PASS
442	198	50	100	87.4	24098	PASS
443	442	15	24	18.5	4465	PASS

m/z	Abundance
44.00	290.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

m/z	Abundance
38.00	191.0
39.00	595.0
40.00	193.0
44.00	265.0
50.00	2283.0
51.00	9145.0
52.00	447.0
56.00	226.0
57.00	672.0
61.00	157.0
62.00	163.0
63.00	485.0
65.00	204.0
69.00	9566.0
74.00	840.0
75.00	1755.0
75.90	442.0
77.00	11107.0
78.00	692.0
79.00	834.0
79.90	468.0
81.00	802.0
82.00	180.0
83.00	196.0
85.00	183.0
86.00	304.0
90.90	192.0
91.90	272.0
93.00	1157.0
98.00	849.0
99.00	819.0
101.00	513.0
103.90	330.0
104.90	257.0
107.00	3642.0
108.00	629.0
110.00	6370.0
110.90	889.0
116.10	191.0
117.00	3039.0
118.00	194.0
121.90	318.0
123.00	517.0
125.00	195.0
127.00	11851.0
128.00	845.0
129.00	4995.0
129.90	409.0
133.90	160.0
135.00	484.0
135.90	173.0
136.90	241.0
140.90	680.0
141.90	242.0
142.80	188.0
146.00	177.0
146.90	231.0
147.90	792.0
149.00	172.0
153.00	186.0
153.90	214.0
155.00	474.0
156.00	715.0
159.90	187.0
161.00	367.0
164.90	313.0
166.00	349.0
166.90	1341.0
168.00	682.0
172.90	189.0
173.90	321.0
175.00	499.0
175.90	204.0
176.80	341.0
178.90	1158.0
179.90	829.0
181.00	445.0
185.00	627.0
186.00	4367.0
187.00	1358.0
187.90	154.0
188.90	291.0
191.90	297.0
192.90	420.0
196.00	955.0
197.90	30096.0
198.90	2100.0
199.90	211.0
201.30	166.0
202.90	241.0
204.00	1176.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

205.00	2227.0
206.00	9084.0
207.00	972.0
207.90	293.0
210.90	325.0
215.00	153.0
215.90	163.0
216.90	2297.0
217.90	239.0
220.90	1303.0
221.90	475.0
223.00	530.0
224.00	4528.0
225.00	1388.0
227.00	2076.0
227.90	338.0
229.00	431.0
231.00	232.0
233.90	154.0
234.90	184.0
236.90	165.0
241.90	207.0
243.10	280.0
244.00	3508.0
245.00	512.0
245.90	611.0
253.90	209.0
255.00	18568.0
255.90	2657.0
256.90	289.0
257.90	1140.0
259.00	185.0
264.90	529.0
272.90	696.0
274.00	1598.0
275.00	7926.0
276.00	1109.0
277.00	829.0
283.10	168.0
295.90	2188.0
296.90	384.0
303.10	272.0
314.00	196.0
314.90	225.0
316.00	210.0
323.00	666.0
326.90	193.0
333.90	469.0
345.90	181.0
352.00	311.0
353.00	193.0
354.00	258.0
364.90	935.0
365.80	162.0
372.00	327.0
402.90	155.0
420.90	207.0
421.90	287.0
423.00	1042.0
423.90	275.0
441.00	2765.0
442.00	19616.0
443.00	3742.0
444.00	441.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

m/z	Abundance
39.00	726.0
39.90	158.0
43.90	342.0
50.00	1904.0
51.00	7529.0
52.00	482.0
55.90	224.0
57.00	562.0
62.00	179.0
63.00	448.0
68.00	183.0
69.00	9576.0
74.00	934.0
75.00	1519.0
76.00	467.0
77.00	9935.0
78.00	673.0
79.00	821.0
80.00	488.0
81.00	683.0
81.90	184.0
83.00	232.0
85.00	169.0
86.00	321.0
90.90	198.0
93.00	1261.0
97.90	1057.0
99.00	754.0
100.90	470.0
102.90	187.0
104.00	327.0
105.00	313.0
107.00	3728.0
107.90	610.0
108.90	159.0
110.00	6165.0
110.90	1097.0
112.00	197.0
116.00	251.0
117.00	2791.0
118.00	252.0
121.90	300.0
123.00	321.0
123.90	257.0
124.90	167.0
127.00	12069.0
127.90	914.0
129.00	4978.0
129.90	467.0
133.90	239.0
135.00	292.0
136.00	181.0
137.00	161.0
140.90	819.0
142.10	162.0
142.90	152.0
147.00	375.0
147.90	681.0
148.90	252.0
153.00	263.0
154.00	210.0
154.90	468.0
156.00	739.0
160.00	331.0
161.00	427.0
165.00	317.0
166.00	274.0
167.00	1463.0
167.90	597.0
172.90	190.0
174.00	344.0
175.00	530.0
175.90	271.0
177.00	324.0
179.00	1273.0
180.00	926.0
180.90	248.0
185.00	491.0
186.00	4051.0
186.90	1318.0
187.90	155.0
188.90	351.0
190.90	165.0
192.00	300.0
193.00	528.0
195.90	1198.0
197.90	32040.0
198.90	2200.0
199.90	158.0
201.40	160.0
202.90	316.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

204.00	1300.0
205.00	2127.0
206.00	9068.0
206.90	1144.0
207.90	369.0
210.00	169.0
211.00	454.0
216.00	207.0
216.90	2744.0
217.90	292.0
220.90	1447.0
221.90	365.0
222.90	667.0
224.00	5287.0
225.00	1315.0
226.90	1916.0
227.90	303.0
229.00	524.0
235.00	190.0
235.90	161.0
242.00	252.0
243.00	372.0
244.00	4121.0
245.00	708.0
246.00	829.0
246.90	203.0
253.10	178.0
255.00	22008.0
256.00	3117.0
256.90	267.0
258.00	1352.0
258.90	218.0
264.90	486.0
272.90	538.0
273.90	1865.0
275.00	9043.0
276.00	1370.0
276.90	885.0
285.00	160.0
292.90	193.0
296.00	2406.0
296.90	443.0
302.90	244.0
314.00	200.0
315.00	332.0
316.00	294.0
323.00	741.0
324.00	183.0
326.90	159.0
334.00	747.0
345.90	198.0
352.00	172.0
352.90	214.0
354.00	192.0
364.90	1222.0
372.00	542.0
402.00	247.0
403.00	252.0
421.00	155.0
422.00	458.0
423.00	1649.0
424.00	354.0
441.00	4569.0
442.00	29824.0
443.00	5287.0
444.00	512.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

m/z	Abundance
39.00	434.0
44.00	406.0
50.00	1187.0
51.00	4731.0
52.00	197.0
56.00	177.0
56.90	332.0
63.00	174.0
69.00	5280.0
74.00	534.0
75.00	1058.0
76.00	328.0
77.00	6088.0
78.00	450.0
79.00	492.0
79.90	347.0
80.90	589.0
81.90	188.0
86.00	173.0
91.90	159.0
92.90	706.0
98.00	611.0
98.90	554.0
100.90	294.0
103.90	249.0
104.90	178.0
107.00	2301.0
108.00	362.0
110.00	3709.0
111.00	496.0
117.00	1636.0
127.00	7907.0
127.90	529.0
128.90	3263.0
129.90	213.0
134.00	163.0
134.90	365.0
140.90	422.0
141.90	160.0
146.90	260.0
147.90	500.0
154.90	375.0
155.90	375.0
161.00	182.0
166.00	158.0
166.90	881.0
168.00	364.0
173.90	240.0
175.00	321.0
176.90	206.0
178.90	692.0
180.00	393.0
180.90	176.0
184.90	525.0
186.00	2968.0
187.00	767.0
188.90	263.0
191.90	294.0
192.90	319.0
195.90	641.0
197.90	20624.0
198.90	1229.0
199.90	163.0
202.80	217.0
203.90	880.0
205.00	1574.0
206.00	5826.0
207.00	770.0
208.00	233.0
210.90	162.0
216.00	173.0
216.90	1552.0
217.80	184.0
220.90	748.0
221.80	176.0
222.90	204.0
224.00	3245.0
225.00	887.0
226.90	1440.0
228.00	260.0
228.90	363.0
234.00	161.0
236.90	169.0
243.00	224.0
244.00	2624.0
245.00	502.0
245.90	664.0
254.90	13786.0
255.90	2073.0
257.00	181.0
257.90	857.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24

259.00	172.0
264.90	353.0
272.90	572.0
274.00	1147.0
275.00	6279.0
275.90	786.0
276.90	576.0
295.90	1762.0
296.90	240.0
303.00	214.0
314.90	245.0
316.00	184.0
323.00	585.0
326.90	156.0
334.00	386.0
345.90	216.0
351.90	227.0
364.90	894.0
372.00	359.0
382.90	191.0
402.00	209.0
403.00	232.0
404.00	152.0
421.00	176.0
421.90	198.0
423.00	1123.0
424.00	211.0
441.00	3653.0
442.00	22856.0
443.00	4366.0
444.00	461.0

Instrument :
BNA_G
ClientSampleId :
DFTPP24