



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011624\
Data File : BG060332.D
Acq On : 16 Jan 2024 12:10
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jan 09 04:07:12 2024






AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2487

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	232384	PASS
68	69	0.00	2	1.2	3308	PASS
69	198	0.00	100	30.9	279061	PASS
70	69	0.00	2	0.4	987	PASS
127	198	10	80	31.8	287125	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	902464	PASS
199	198	5	9	7.0	63221	PASS
275	198	10	60	27.4	247467	PASS
365	198	1	100	3.6	32675	PASS
441	198	0.01	100	13.2	119088	PASS
442	442	50	100	100.0	717227	PASS
443	442	15	24	19.8	142349	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011624\
Data File : BG060332.D
Acq On : 16 Jan 2024 12:10
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 16 13:18:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration



TIC: BG060332.D\data.ms

(70) Pentachlorophenol (C)

16.966min (-0.000) 25318.90 ng




response 938300

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	71.80	63.09
264.00	62.30	63.97
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011624\
Data File : BG060332.D
Acq On : 16 Jan 2024 12:10
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 16 13:18:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration



TIC: BG060332.D\data.ms

(77) Benzidine

19.563min (+ 0.006) 1291351.18 ng m

response 5779303

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.30	22.89#
183.00	11.20	17.40#
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/16/2024	BNA_G	<u>BG060332.D</u>
Compound Name	Response	Retention Time
DDT	3071676	20.791
DDD	62231	20.344
DDE	3627	19.845
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
65858	3137534	2.10

Instrument :
BNA_G
ClientSampleId :
DFTPP