

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011724\
 Data File : BG060343.D
 Acq On : 17 Jan 2024 15:01
 Operator : MA/JU
 Sample : PB158472BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB158472BL

Quant Time: Jan 17 15:51:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.932	152	233644	20.000	ng	0.00
21) Naphthalene-d8	10.735	136	1026107	20.000	ng	0.00
39) Acenaphthene-d10	14.566	164	852946	20.000	ng	0.00
64) Phenanthrene-d10	17.315	188	2376940	20.000	ng	0.00
76) Chrysene-d12	21.575	240	2333840	20.000	ng	0.00
86) Perylene-d12	24.742	264	2519752	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.506	112	2075038	146.077	ng	0.00
7) Phenol-d6	7.098	99	3030142	144.056	ng	0.00
23) Nitrobenzene-d5	9.095	82	2048608	94.274	ng	0.00
42) 2,4,6-Tribromophenol	16.058	330	1699068	135.392	ng	0.00
45) 2-Fluorobiphenyl	13.191	172	4820801	78.580	ng	0.00
79) Terphenyl-d14	19.936	244	6903678	69.740	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011724\
Data File : BG060343.D
Acq On : 17 Jan 2024 15:01
Operator : MA/JU
Sample : PB158472BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB158472BL

Quant Time: Jan 17 15:51:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.932 min Scan# 81
Delta R.T. 0.002 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01




Instrument :
BNA_G
ClientSampleId :
PB158472BL


Tgt Ion:152 Resp: 233644
Ion Ratio Lower Upper
152 100
150 144.5 114.5 171.7
115 53.1 43.0 64.6



#5
2-Fluorophenol
Concen: 146.077 ng
RT: 5.506 min Scan# 411
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01

Tgt Ion:112 Resp: 2075038
Ion Ratio Lower Upper
112 100
64 64.4 49.8 74.6
63 30.7 24.4 36.6









#7
Phenol-d6
Concen: 144.056 ng
RT: 7.098 min Scan# 61
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01

Instrument :
BNA_G
ClientSampleId :
PB158472BL

Tgt Ion: 99 Resp: 3030142




Ion	Ratio	Lower	Upper
99	100		
42	14.3	10.9	16.3
71	30.1	21.5	32.3




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.735 min Scan# 1301
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01

Tgt Ion:136 Resp: 1026107

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.2	12.4
54	8.6	6.2	9.2
68	8.5	6.3	9.5









#23
Nitrobenzene-d5
Concen: 94.274 ng
RT: 9.095 min Scan# 1022
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01




Instrument :
BNA_G
ClientSampleId :
PB158472BL


Tgt Ion: 82 Resp: 2048608
Ion Ratio Lower Upper
82 100
128 37.0 29.6 44.4
54 41.9 33.6 50.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.566 min Scan# 1953
Delta R.T. -0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01

Tgt Ion:164 Resp: 852946
Ion Ratio Lower Upper
164 100
162 103.0 84.4 126.6
160 45.0 34.6 52.0









#42
2,4,6-Tribromophenol
Concen: 135.392 ng
RT: 16.058 min Scan# 21
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01




Instrument :
BNA_G
ClientSampleId :
PB158472BL


Tgt Ion:330 Resp: 1699068
Ion Ratio Lower Upper
330 100
332 96.4 76.7 115.1
141 24.6 18.6 28.0



#45
2-Fluorobiphenyl
Concen: 78.580 ng
RT: 13.191 min Scan# 1719
Delta R.T. 0.003 min
Lab File: BG060343.D
Acq: 17 Jan 2024 15:01

Tgt Ion:172 Resp: 4820801
Ion Ratio Lower Upper
172 100
171 45.7 34.6 51.8
170 31.9 24.4 36.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.315 min Scan# 24

Delta R.T. 0.003 min

Lab File: BG060343.D

Acq: 17 Jan 2024 15:01

Instrument :

BNA_G

ClientSampleId :

PB158472BL





Tgt Ion:188 Resp: 2376940

Ion Ratio Lower Upper

188 100

94 9.3 7.1 10.7

80 10.8 8.5 12.7



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.575 min Scan# 3146

Delta R.T. -0.003 min

Lab File: BG060343.D

Acq: 17 Jan 2024 15:01




Tgt Ion:240 Resp: 2333840


Ion Ratio Lower Upper

240 100

120 10.6 7.3 10.9

236 27.2 20.3 30.5





#79

Terphenyl-d14

Concen: 69.740 ng

RT: 19.936 min Scan# 2

Delta R.T. 0.003 min

Lab File: BG060343.D

Acq: 17 Jan 2024 15:01

Instrument :

BNA_G

ClientSampleId :

PB158472BL





Tgt Ion:244 Resp: 6903678

Ion Ratio Lower Upper

244 100

212 20.0 10.9 16.3#

122 21.9 12.6 18.8#



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.742 min Scan# 3685

Delta R.T. 0.003 min

Lab File: BG060343.D

Acq: 17 Jan 2024 15:01

Tgt Ion:264 Resp: 2519752

Ion Ratio Lower Upper

264 100

260 23.4 17.8 26.6

265 21.7 16.8 25.2

