

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011824\
 Data File : BG060350.D
 Acq On : 18 Jan 2024 13:22
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 13:56:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	53.174	-32.9#	113	0.00
3	Pyridine	40.000	46.874	-17.2	95	0.00
4	n-Nitrosodimethylamine	40.000	39.875	0.3	85	0.00
5 S	2-Fluorophenol	80.000	82.230	-2.8	106	0.00
6	Aniline	40.000	44.867	-12.2	101	0.00
7 S	Phenol-d6	80.000	88.695	-10.9	92	0.00
8	2-Chlorophenol	40.000	42.075	-5.2	100	0.00
9	Benzaldehyde	40.000	39.605	1.0	88	0.00
10 C	Phenol	40.000	44.022	-10.1	92	0.00
11	bis(2-Chloroethyl)ether	40.000	42.149	-5.4	101	0.00
12	1,3-Dichlorobenzene	40.000	40.572	-1.4	98	0.00
13 C	1,4-Dichlorobenzene	40.000	40.158	-0.4	99	0.00
14	1,2-Dichlorobenzene	40.000	38.813	3.0	85	0.00
15	Benzyl Alcohol	40.000	45.121	-12.8	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.277	1.8	85	0.00
17	2-Methylphenol	40.000	41.473	-3.7	87	0.00
18	Hexachloroethane	40.000	39.097	2.3	85	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.833	-9.6	89	0.00
20	3+4-Methylphenols	40.000	43.848	-9.6	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	41.044	-2.6	91	0.00
23 S	Nitrobenzene-d5	80.000	84.514	-5.6	95	0.00
24	Nitrobenzene	40.000	39.927	0.2	89	0.00
25	Isophorone	40.000	41.717	-4.3	93	0.00
26 C	2-Nitrophenol	40.000	43.164	-7.9	95	0.00
27	2,4-Dimethylphenol	40.000	41.058	-2.6	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.591	-1.5	92	0.00
29 C	2,4-Dichlorophenol	40.000	41.191	-3.0	91	0.00
30	1,2,4-Trichlorobenzene	40.000	39.023	2.4	88	0.00
31	Naphthalene	40.000	39.669	0.8	91	0.00
32	Benzoic acid	40.000	42.765	-6.9	104	-0.01
33	4-Chloroaniline	40.000	41.216	-3.0	94	0.00
34 C	Hexachlorobutadiene	40.000	37.141	7.1	90	0.00
35	Caprolactam	40.000	46.891	-17.2	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.713	-9.3	98	0.00
37	2-Methylnaphthalene	40.000	40.488	-1.2	94	0.00
38	1-Methylnaphthalene	40.000	40.513	-1.3	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.289	6.8	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.842	5.4	89	0.00
42 S	2,4,6-Tribromophenol	80.000	85.955	-7.4	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.483	-1.2	100	0.00
44	2,4,5-Trichlorophenol	40.000	40.905	-2.3	100	0.00
45 S	2-Fluorobiphenyl	80.000	76.623	4.2	100	0.00
46	1,1'-Biphenyl	40.000	37.796	5.5	96	0.00
47	2-Chloronaphthalene	40.000	38.057	4.9	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011824\
 Data File : BG060350.D
 Acq On : 18 Jan 2024 13:22
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 13:56:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.216	-10.5	105	0.00
49	Acenaphthylene	40.000	40.216	-0.5	99	0.00
50	Dimethylphthalate	40.000	40.513	-1.3	101	0.00
51	2,6-Dinitrotoluene	40.000	43.919	-9.8	103	0.00
52 C	Acenaphthene	40.000	40.368	-0.9	100	0.00
53	3-Nitroaniline	40.000	44.539	-11.3	106	0.00
54 P	2,4-Dinitrophenol	40.000	44.069	-10.2	116	0.00
55	Dibenzofuran	40.000	40.283	-0.7	100	0.00
56 P	4-Nitrophenol	40.000	50.402	-26.0#	111	0.00
57	2,4-Dinitrotoluene	40.000	45.810	-14.5	109	0.00
58	Fluorene	40.000	40.627	-1.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.515	-8.8	106	0.00
60	Diethylphthalate	40.000	41.837	-4.6	103	0.00
61	4-Chlorophenyl-phenylether	40.000	40.148	-0.4	102	0.00
62	4-Nitroaniline	40.000	45.129	-12.8	107	0.00
63	Azobenzene	40.000	40.464	-1.2	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.348	-10.9	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.703	0.7	103	0.00
67	4-Bromophenyl-phenylether	40.000	39.666	0.8	104	0.00
68	Hexachlorobenzene	40.000	38.681	3.3	102	0.00
69	Atrazine	40.000	37.026	7.4	98	0.00
70 C	Pentachlorophenol	40.000	46.014	-15.0	112	0.00
71	Phenanthrene	40.000	39.148	2.1	103	0.00
72	Anthracene	40.000	39.233	1.9	103	0.00
73	Carbazole	40.000	40.130	-0.3	104	0.00
74	Di-n-butylphthalate	40.000	40.704	-1.8	104	0.00
75 C	Fluoranthene	40.000	38.762	3.1	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	58.968	-47.4#	151	0.00
78	Pyrene	40.000	37.873	5.3	104	0.00
79 S	Terphenyl-d14	80.000	74.973	6.3	104	0.00
80	Butylbenzylphthalate	40.000	42.784	-7.0	110	0.00
81	Benzo(a)anthracene	40.000	38.699	3.3	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.255	-0.6	103	0.00
83	Chrysene	40.000	38.907	2.7	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.686	-4.2	107	0.00
85 c	Di-n-octyl phthalate	40.000	42.342	-5.9	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.825	0.4	104	0.00
88	Benzo(b)fluoranthene	40.000	39.736	0.7	102	0.00
89	Benzo(k)fluoranthene	40.000	39.822	0.4	102	0.00
90 C	Benzo(a)pyrene	40.000	39.844	0.4	104	0.00
91	Dibenzo(a,h)anthracene	40.000	39.681	0.8	102	0.00
92	Benzo(g,h,i)perylene	40.000	39.745	0.6	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011824\
Data File : BG060350.D
Acq On : 18 Jan 2024 13:22
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 18 13:56:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0