

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011824\
 Data File : BG060355.D
 Acq On : 18 Jan 2024 16:50
 Operator : MA/JU
 Sample : P1159-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-011724

Quant Time: Jan 18 17:23:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.926	152	267690	20.000	ng	0.00
21) Naphthalene-d8	10.729	136	930519	20.000	ng	0.00
39) Acenaphthene-d10	14.566	164	680440	20.000	ng	0.00
64) Phenanthrene-d10	17.310	188	1703139	20.000	ng	0.00
76) Chrysene-d12	21.569	240	1554065	20.000	ng	0.00
86) Perylene-d12	24.736	264	1801811	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.500	112	1708866	104.999	ng	0.00
7) Phenol-d6	7.092	99	2518942	104.522	ng	0.00
23) Nitrobenzene-d5	9.090	82	1330430	67.514	ng	0.00
42) 2,4,6-Tribromophenol	16.052	330	962580	96.150	ng	0.00
45) 2-Fluorobiphenyl	13.185	172	3255062	66.509	ng	0.00
79) Terphenyl-d14	19.930	244	4711347	72.288	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011824\
Data File : BG060355.D
Acq On : 18 Jan 2024 16:50
Operator : MA/JU
Sample : P1159-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-011724

Quant Time: Jan 18 17:23:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.926 min Scan# 81
Delta R.T. -0.004 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50




Instrument :
BNA_G
ClientSampleId :
TP-011724


Tgt Ion:152 Resp: 267690
Ion Ratio Lower Upper
152 100
150 146.9 114.5 171.7
115 53.7 43.0 64.6



#5
2-Fluorophenol
Concen: 104.999 ng
RT: 5.500 min Scan# 410
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50

Tgt Ion:112 Resp: 1708866
Ion Ratio Lower Upper
112 100
64 65.2 49.8 74.6
63 30.9 24.4 36.6









#7
Phenol-d6
Concen: 104.522 ng
RT: 7.092 min Scan# 61
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50

Instrument :
BNA_G
ClientSampleId :
TP-011724

Tgt Ion: 99 Resp: 2518942




Ion	Ratio	Lower	Upper
99	100		
42	12.7	10.9	16.3
71	29.3	21.5	32.3




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.729 min Scan# 1300
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50

Tgt Ion: 136 Resp: 930519

Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.2	12.4
54	7.8	6.2	9.2
68	7.2	6.3	9.5









#23
Nitrobenzene-d5
Concen: 67.514 ng
RT: 9.090 min Scan# 1022
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50




Instrument :
BNA_G
ClientSampleId :
TP-011724


Tgt Ion: 82 Resp: 1330430
Ion Ratio Lower Upper
82 100
128 37.4 29.6 44.4
54 41.1 33.6 50.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.566 min Scan# 1953
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50

Tgt Ion:164 Resp: 680440
Ion Ratio Lower Upper
164 100
162 102.6 84.4 126.6
160 43.3 34.6 52.0









#42
2,4,6-Tribromophenol
Concen: 96.150 ng
RT: 16.052 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50




Instrument :
BNA_G
ClientSampleId :
TP-011724


Tgt Ion:330 Resp: 962580
Ion Ratio Lower Upper
330 100
332 99.1 76.7 115.1
141 24.1 18.6 28.0



#45
2-Fluorobiphenyl
Concen: 66.509 ng
RT: 13.185 min Scan# 1718
Delta R.T. -0.003 min
Lab File: BG060355.D
Acq: 18 Jan 2024 16:50

Tgt Ion:172 Resp: 3255062
Ion Ratio Lower Upper
172 100
171 38.5 34.6 51.8
170 27.4 24.4 36.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.310 min Scan# 24

Delta R.T. -0.003 min

Lab File: BG060355.D

Acq: 18 Jan 2024 16:50

Instrument :

BNA_G

ClientSampleId :

TP-011724





Tgt Ion:188 Resp: 1703139

Ion Ratio Lower Upper

188 100

94 9.0 7.1 10.7

80 9.9 8.5 12.7



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.569 min Scan# 3145

Delta R.T. -0.009 min

Lab File: BG060355.D

Acq: 18 Jan 2024 16:50




Tgt Ion:240 Resp: 1554065


Ion Ratio Lower Upper

240 100

120 10.3 7.3 10.9

236 25.2 20.3 30.5









#79
 Terphenyl-d14
 Concen: 72.288 ng
 RT: 19.930 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG060355.D
 Acq: 18 Jan 2024 16:50

Instrument :
 BNA_G
 ClientSampleId :
 TP-011724

Tgt Ion:244 Resp: 4711347
 Ion Ratio Lower Upper
 244 100
 212 11.1 10.9 16.3
 122 13.3 12.6 18.8



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.736 min Scan# 3684
 Delta R.T. -0.003 min
 Lab File: BG060355.D
 Acq: 18 Jan 2024 16:50

Tgt Ion:264 Resp: 1801811
 Ion Ratio Lower Upper
 264 100
 260 23.0 17.8 26.6
 265 21.3 16.8 25.2

