

Data Path : Z:\HPCHEM1\BNA G\Data\BG011918\
 Data File : BG032151.D
 Acq On : 19 Jan 2018 14:43
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 22:16:42 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 10:13:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.420	0.355	15.5	70	0.00
3	Pyridine	1.121	1.038	7.4	73	0.00
4	n-Nitrosodimethylamine	0.394	0.436	-10.7	86	-0.01
5 S	2-Fluorophenol	1.094	1.036	5.3	76	0.00
6	Aniline	1.737	1.690	2.7	77	0.00
7 S	Phenol-d6	1.377	1.388	-0.8	79	0.00
8	2-Chlorophenol	1.224	1.162	5.1	75	0.00
9	Benzaldehyde	0.798	0.752	5.8	73	0.00
10 C	Phenol	1.357	1.290	4.9	75	0.00
11	bis(2-Chloroethyl)ether	1.087	1.039	4.4	76	0.00
12	1,3-Dichlorobenzene	1.602	1.574	1.7	78	0.00
13 C	1,4-Dichlorobenzene	1.613	1.593	1.2	79	0.00
14	1,2-Dichlorobenzene	1.560	1.521	2.5	79	0.00
15	Benzyl Alcohol	1.001	1.087	-8.6	84	0.00
16	2,2'-oxybis(1-Chloropropane	1.105	1.070	3.2	78	0.00
17	2-Methylphenol	1.037	1.006	3.0	76	0.00
18	Hexachloroethane	0.496	0.527	-6.3	85	0.00
19 P	n-Nitroso-di-n-propylamine	0.855	0.885	-3.5	82	0.00
20	3+4-Methylphenols	1.436	1.423	0.9	79	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
22	Acetophenone	0.454	0.451	0.7	79	0.00
23 S	Nitrobenzene-d5	0.296	0.316	-6.8	83	0.00
24	Nitrobenzene	0.295	0.312	-5.8	83	0.00
25	Isophorone	0.550	0.561	-2.0	80	0.00
26 C	2-Nitrophenol	0.175	0.184	-5.1	80	0.00
27	2,4-Dimethylphenol	0.277	0.273	1.4	79	0.00
28	bis(2-Chloroethoxy)methane	0.332	0.325	2.1	77	0.00
29 C	2,4-Dichlorophenol	0.341	0.359	-5.3	82	0.00
30	1,2,4-Trichlorobenzene	0.441	0.446	-1.1	80	0.00
31	Naphthalene	0.991	0.981	1.0	78	0.00
32	Benzoic acid	0.204	0.190	6.9	69	0.02
33	4-Chloroaniline	0.425	0.436	-2.6	81	0.00
34 C	Hexachlorobutadiene	0.316	0.351	-11.1	88	0.00
35	Caprolactam	0.104	0.115	-10.6	84	0.00
36 C	4-Chloro-3-methylphenol	0.312	0.347	-11.2	89	0.00
37	2-Methylnaphthalene	0.787	0.801	-1.8	81	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
39	1,2,4,5-Tetrachlorobenzene	0.872	0.867	0.6	87	0.00
40 P	Hexachlorocyclopentadiene	0.491	0.509	-3.7	89	0.00
41 S	2,4,6-Tribromophenol	0.331	0.340	-2.7	87	0.00
42 C	2,4,6-Trichlorophenol	0.490	0.494	-0.8	87	0.00
43	2,4,5-Trichlorophenol	0.567	0.588	-3.7	91	0.00
44 S	2-Fluorobiphenyl	1.613	1.567	2.9	86	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG011918\
 Data File : BG032151.D
 Acq On : 19 Jan 2018 14:43
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 22:16:42 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 10:13:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.715	1.632	4.8	84	0.00
46	2-Chloronaphthalene	1.334	1.264	5.2	83	0.00
47	2-Nitroaniline	0.255	0.294	-15.3	96	0.00
48	Acenaphthylene	2.031	1.963	3.3	84	0.00
49	Dimethylphthalate	1.750	1.765	-0.9	88	0.00
50	2,6-Dinitrotoluene	0.363	0.382	-5.2	89	0.00
51 C	Acenaphthene	1.343	1.347	-0.3	87	0.00
52	3-Nitroaniline	0.328	0.354	-7.9	91	0.00
53 P	2,4-Dinitrophenol	0.153	0.166	-8.5	91	0.00
54	Dibenzofuran	2.004	1.985	0.9	87	0.00
55 P	4-Nitrophenol	0.239	0.266	-11.3	92	0.00
56	2,4-Dinitrotoluene	0.489	0.547	-11.9	91	0.00
57	Fluorene	1.643	1.648	-0.3	88	0.00
58	2,3,4,6-Tetrachlorophenol	0.525	0.585	-11.4	94	0.00
59	Diethylphthalate	1.697	1.744	-2.8	90	0.00
60	4-Chlorophenyl-phenylether	1.008	1.049	-4.1	92	0.00
61	4-Nitroaniline	0.315	0.373	-18.4	96	0.00
62	Azobenzene	1.062	1.106	-4.1	90	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.128	-7.6	100	0.00
65 c	n-Nitrosodiphenylamine	0.607	0.568	6.4	87	0.00
66	4-Bromophenyl-phenylether	0.285	0.279	2.1	93	0.00
67	Hexachlorobenzene	0.315	0.294	6.7	89	0.00
68	Atrazine	0.255	0.248	2.7	91	0.00
69 C	Pentachlorophenol	0.201	0.194	3.5	89	0.00
70	Phenanthrene	1.114	1.048	5.9	90	0.00
71	Anthracene	1.111	1.055	5.0	90	0.00
72	Carbazole	0.963	0.951	1.2	93	0.00
73	Di-n-butylphthalate	1.138	1.069	6.1	89	0.00
74 C	Fluoranthene	1.394	1.350	3.2	93	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
76	Benzidine	0.500	0.490	2.0	96	0.00
77	Pyrene	1.257	1.162	7.6	93	0.00
78 S	Terphenyl-d14	0.906	0.838	7.5	95	0.00
79	Butylbenzylphthalate	0.450	0.406	9.8	90	0.00
80	Benzo(a)anthracene	1.244	1.208	2.9	98	0.00
81	3,3'-Dichlorobenzidine	0.465	0.480	-3.2	100	0.00
82	Chrysene	1.195	1.160	2.9	97	0.00
83	Bis(2-ethylhexyl)phthalate	0.661	0.578	12.6	88	0.00
84 c	Di-n-octyl phthalate	1.049	0.923	12.0	87	0.00
85	Indeno(1,2,3-cd)pyrene	1.462	1.474	-0.8	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Benzo(b)fluoranthene	1.209	1.162	3.9	94	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG011918\
Data File : BG032151.D
Acq On : 19 Jan 2018 14:43
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 19 22:16:42 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 18 10:13:16 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.181	1.156	2.1	97	0.00
89 C	Benzo(a)pyrene	1.144	1.124	1.7	97	0.00
90	Dibenzo(a,h)anthracene	1.102	1.127	-2.3	101	0.00
91	Benzo(a,h,i)perylene	1.126	1.122	0.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0