


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG012020\
 Data File : BG044091.D
 Acq On : 20 Jan 2020 9:38
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG123019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 31 13:32:23 2019



AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.6	22904	PASS
68	69	0.00	2	0.7	154	PASS
69	198	0.00	100	35.5	22853	PASS
70	69	0.00	2	0.2	51	PASS
127	198	10	80	45.5	29290	PASS
197	198	0.00	2	0.7	433	PASS
198	198	100	100	100.0	64357	PASS
199	198	5	9	6.8	4360	PASS
275	198	10	60	24.6	15851	PASS
365	198	1	100	3.0	1899	PASS
441	198	0.01	100	11.1	7137	PASS
442	198	50	100	66.8	42994	PASS
443	442	15	24	19.8	8521	PASS