



2,4-Dichlorophenol

Response Ratio



$R = 6.804e-002 A^2 + 1.698e-001 A + 2.562e-002$
 Coef of Det (r^2) = 0.993951 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021


Benzoic acid



$R = 6.108e-002 A^2 + 1.093e-002 A + 3.407e-002$
 Coef of Det (r^2) = 0.991106 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

4-Chloro-3-methylphenol


Response Ratio



$R = 5.236e-002 A^2 + 2.624e-001 A + 2.660e-003$
 Coef of Det (r^2) = 0.994081 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2,4,6-Trichlorophenol


Response Ratio



$R = 1.083e-001 A^2 + 1.734e-001 A + 3.396e-002$
 Coef of Det (r^2) = 0.996075 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2,4,5-Trichlorophenol


Response Ratio



$R = 9.207e-002 A^2 + 2.350e-001 A + 2.325e-002$
 Coef of Det (r^2) = 0.995795 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2-Nitroaniline


Response Ratio



$R = 1.213e-001 A^2 + 6.901e-002 A + 3.470e-002$
 Coef of Det (r^2) = 0.993310 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2,6-Dinitrotoluene


Response Ratio



$R = 6.421e-002 A^2 + 1.543e-001 A + 1.773e-002$
 Coef of Det (r^2) = 0.996339 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2,4-Dinitrotoluene


Response Ratio



$R = 9.706e-002 A^2 + 2.067e-001 A + 2.003e-002$
Coef of Det (r^2) = 0.994001 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

2,3,4,6-Tetrachlorophenol


Response Ratio



$R = 1.069e-001 A^2 + 1.958e-001 A + 2.900e-002$
 Coef of Det (r^2) = 0.995656 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

Diethylphthalate


Response Ratio



$R = 1.768e-001 A^2 + 1.218e+000 A - 1.906e-002$
 Coef of Det (r^2) = 0.993988 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

Di-n-butylphthalate


Response Ratio



$R = 1.411e-001 A^2 + 7.480e-001 A - 2.507e-003$
 Coef of Det (r^2) = 0.994736 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

Butylbenzylphthalate


Response Ratio



$R = 1.196e-001 A^2 + 1.176e-001 A + 2.947e-002$
 Coef of Det (r^2) = 0.996432 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

Bis(2-ethylhexyl)phthalate


Response Ratio



$R = 1.417e-001 A^2 + 2.602e-001 A + 2.640e-002$
 Coef of Det (r^2) = 0.995107 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021

Di-n-octyl phthalate

Response Ratio



$R = 2.834e-001 A^2 + 2.446e-001 A + 1.015e-001$
 Coef of Det (r^2) = 0.995170 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012021.M
 Calibration Table Last Updated: Thu Jan 21 10:00:28 2021