

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012117\
 Data File : BG025574.D
 Acq On : 21 Jan 2017 15:24
 Operator : UM/SJ
 Sample : I1329-09
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S3-0-3FT-COMP

Manual Integrations
 APPROVED

mohammad
 1/23/2017 5:00:17 PM

Quant Time: Jan 23 02:46:08 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 11:26:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	77212	20.00	ng	-0.02
21) Naphthalene-d8	11.01	136	328706	20.00	ng	-0.02
38) Acenaphthene-d10	14.81	164	252558	20.00	ng	-0.02
63) Phenanthrene-d10	17.56	188	541431	20.00	ng	-0.02
75) Chrysene-d12	21.85	240	695604	20.00	ng	-0.02
86) Perylene-d12	25.23	264	752183	20.00	ng	-0.04

System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	484776	114.12	ng	-0.02
7) Phenol-d6	7.34	99	1031749	172.46	ng	-0.02
23) Nitrobenzene-d5	9.36	82	808035	108.60	ng	-0.02
41) 2,4,6-Tribromophenol	16.30	330	222973	54.84	ng	-0.02
44) 2-Fluorobiphenyl	13.43	172	1481898	94.99	ng	-0.02
78) Terphenyl-d14	20.14	244	2325430	88.64	ng	-0.02

Target Compounds					Qvalue	
10) Phenol	7.37	94	55629	8.39	ng	92
14) 1,2-Dichlorobenzene	8.53	146	17811	3.11	ng	# 79
30) 1,2,4-Trichlorobenzene	10.86	180	78184	11.79	ng	90
31) Naphthalene	11.06	128	101937	6.21	ng	98
37) 2-Methylnaphthalene	12.65	142	46190	3.80	ng	97
45) 1,1'-Biphenyl	13.64	154	37418	2.20	ng	98
49) Dimethylphthalate	14.25	163	192038	10.43	ng	97
51) Acenaphthene	14.87	154	63276	4.70	ng	91
54) Dibenzofuran	15.21	168	54654	2.57	ng	93
57) Fluorene	15.86	166	62848	3.53	ng	# 96
70) Phenanthrene	17.60	178	831472	29.80	ng	95
71) Anthracene	17.69	178	190663	6.73	ng	98
72) Carbazole	17.97	167	106090	4.18	ng	97
74) Fluoranthene	19.60	202	1261352	34.23	ng	95
77) Pyrene	19.96	202	1082631	29.78	ng	96
80) Benzo(a)anthracene	21.82	228	654561	16.87	ng	98
82) Chrysene	21.89	228	625247	16.80	ng	95
85) Indeno(1,2,3-cd)pyrene	29.13	276	390196	8.25	ng	# 97
87) Benzo(b)fluoranthene	24.15	252	750039	18.28	ng	99
88) Benzo(k)fluoranthene	24.21	252	288371m	7.28	ng	
89) Benzo(a)pyrene	25.07	252	556759	14.05	ng	98
90) Dibenz(a,h)anthracene	29.18	278	111030m	2.64	ng	
91) Benzo(g,h,i)perylene	30.35	276	375666	9.00	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA G\DATA\BG012117\
Data File : BG025574.D
Acq On : 21 Jan 2017 15:24
Operator : UM/SJ
Sample : I1329-09
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Manual Integrations
APPROVED

mohammad
1/23/2017 5:00:17 PM

Quant Time: Jan 23 02:46:08 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 10 11:26:22 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.18 min Scan# 909

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Instrument :

BNA_G





ClientSampleId :

S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.0	114.0	171.0
115	63.9	48.1	72.1

Manual Integrations
APPROVED

mohammad
1/23/2017 5:00:17 PM



#5

2-Fluorophenol

Concen: 114.12 ng




RT: 5.72 min Scan# 490


Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
112	100		
64	75.7	61.8	92.6
63	45.9	37.2	55.8





#7

Phenol-d6

Concen: 172.46 ng

RT: 7.34 min Scan# 766

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Instrument :

BNA_G

ClientSampleId :

S3-0-3FT-COMP

Tot Ion: 99 Resp: 1031749

Ion Ratio Lower Upper

99 100

42 25.1 20.5 30.7


71 50.7 37.6 56.4

Manual Integrations

APPROVED

mohammad




1/23/2017 5:00:17 PM



Abundance Ion 99.00 (98.70 to 99.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 71.00 (70.70 to 71.70): BGC



#10

Phenol

Concen: 8.39 ng

RT: 7.37 min Scan# 771

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24


Tgt Ion: 94 Resp: 55629

Ion Ratio Lower Upper

94 100

65 36.4 20.6 60.6



66 62.5 35.5 75.5




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC





#14

1,2-Dichlorobenzene

Concen: 3.11 ng

RT: 8.53 min Scan# 969

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Instrument :

BNA_G





ClientSampleId :

S3-0-3FT-COMP

Tot Ion:	146	Resp:	17811
Ion Ratio	Lower	Upper	
146	100		
148	49.9	54.6	81.8#
111	37.2	39.7	59.5#

Manual Integrations
APPROVED

mohammad
1/23/2017 5:00:17 PM



#21

Naphthalene-d8

Concen: 20.00 ng




RT: 11.01 min Scan# 1390


Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Tgt Ion:	136	Resp:	328706
Ion Ratio	Lower	Upper	
136	100		
137	11.6	8.9	13.3
54	11.1	9.3	13.9
68	5.6	5.1	7.7









#23
 Nitrobenzene-d5
 Concen: 108.60 ng
 RT: 9.36 min Scan# 1110
 Delta R.T. -0.02 min
 Lab File: BG025574.D
 Acq: 21 Jan 2017 15:24

Instrument :
 BNA_G
 ClientSampleId :
 S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
82	100		
128	33.2	26.4	39.6
54	62.6	49.4	74.2




Manual Integrations
 APPROVED


mohammad
 1/23/2017 5:00:17 PM



#30
 1,2,4-Trichlorobenzene
 Concen: 11.79 ng
 RT: 10.86 min Scan# 1366
 Delta R.T. -0.02 min
 Lab File: BG025574.D
 Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
180	100		
182	84.2	77.0	115.4
145	26.7	23.4	35.0









#31
Naphthalene
Concen: 6.21 ng
RT: 11.06 min Scan# 1399
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.7	9.3	13.9
127	13.6	11.4	17.0




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#37
2-Methylnaphthalene
Concen: 3.80 ng
RT: 12.65 min Scan# 1670
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.0	70.4	105.6
115	45.0	35.5	53.3









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.81 min Scan# 2037
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Resp	Lower	Upper
164	100	252558		
162	99.7	85.1	127.7	
160	43.7	34.7	52.1	




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#41
2,4,6-Tribromophenol
Concen: 54.84 ng
RT: 16.30 min Scan# 2291
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	222973		
332	98.2	77.3	115.9	
141	33.4	32.1	48.1	









#44
2-Fluorobiphenyl
Concen: 94.99 ng
RT: 13.43 min Scan# 1802
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
172	100		
171	39.2	32.2	48.2
170	24.6	21.8	32.6




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#45
1,1'-Biphenyl
Concen: 2.20 ng
RT: 13.64 min Scan# 1839
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
154	100		
153	42.0	23.0	63.0
76	14.0	0.0	33.5









#49
Dimethylphthalate
Concen: 10.43 ng
RT: 14.25 min Scan# 1942
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Resp	Lower	Upper
163	100	192038		
194	4.7		3.8	5.6
164	10.0		9.3	13.9




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#51
Acenaphthene
Concen: 4.70 ng
RT: 14.87 min Scan# 2048
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	63276		
153	97.0		87.8	131.6
152	50.2		42.2	63.4









#54
Dibenzenofuran
Concen: 2.57 ng
RT: 15.21 min Scan# 2105
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
168	100		
139	38.6	35.3	52.9
169	13.0	11.5	17.3




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#57
Fluorene
Concen: 3.53 ng
RT: 15.86 min Scan# 2216
Delta R.T. -0.01 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.9	78.6	117.8
167	20.2	11.6	17.4









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.56 min Scan# 2505
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Resp	Lower	Upper
188	100	541431		
94	7.0		5.8	8.8
80	8.1		7.5	11.3




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#70
Phenanthrene
Concen: 29.80 ng
RT: 17.60 min Scan# 2512
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	831472		
176	19.6		17.7	26.5
179	16.2		15.0	22.6





#71

Anthracene

Concen: 6.73 ng

RT: 17.69 min Scan# 2528

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Instrument :

BNA_G





ClientSampleId :

S3-0-3FT-COMP

Tot Ion:	178	Resp:	190663
Ion Ratio	Lower	Upper	
178	100		
176	20.1	16.4	24.6
179	15.9	13.8	20.8

Manual Integrations
APPROVED

mohammad
1/23/2017 5:00:17 PM



#72

Carbazole

Concen: 4.18 ng




RT: 17.97 min Scan# 2575


Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Tgt Ion:	167	Resp:	106090
Ion Ratio	Lower	Upper	
167	100		
166	26.4	20.2	30.2
139	17.8	12.8	19.2





#74

Fluoranthene

Concen: 34.23 ng

RT: 19.60 min Scan# 2852

Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Instrument :

BNA_G





ClientSampleId :

S3-0-3FT-COMP

Tot Ion:	202	Resp:	1261352
Ion Ratio	Lower	Upper	
202	100		
101	8.1	0.0	30.1
203	19.1	1.3	41.3

Manual Integrations
APPROVED

mohammad
1/23/2017 5:00:17 PM



#75

Chrysene-d12

Concen: 20.00 ng




RT: 21.85 min Scan# 3235


Delta R.T. -0.02 min

Lab File: BG025574.D

Acq: 21 Jan 2017 15:24

Tgt Ion:	240	Resp:	695604
Ion Ratio	Lower	Upper	
240	100		
120	5.7	5.7	8.5
236	26.2	22.2	33.4









#77
Pvrene
Concen: 29.78 ng
RT: 19.96 min Scan# 2914
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
202	100		
200	22.5	19.6	29.4
203	18.5	15.8	23.8




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#78
Terphenyl-d14
Concen: 88.64 ng
RT: 20.14 min Scan# 2944
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.8	10.0	15.0
122	8.9	8.6	12.8









#80
Benzo(a)anthracene
Concen: 16.87 ng
RT: 21.82 min Scan# 3231
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.4	24.9	37.3
229	22.1	18.2	27.2




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#82
Chrysene
Concen: 16.80 ng
RT: 21.89 min Scan# 3243
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.1	27.0	40.4
229	22.5	17.8	26.6









#85
Indeno(1,2,3-cd)pyrene
Concen: 8.25 ng
RT: 29.13 min Scan# 4474
Delta R.T. -0.07 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
276	100		
138	10.3	4.7	7.1#
277	26.1	20.6	31.0




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#86
Perylene-d12
Concen: 20.00 ng
RT: 25.23 min Scan# 3811
Delta R.T. -0.04 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.5	21.8	32.8
265	24.3	18.2	27.4









#87
Benzo(b)fluoranthene
Concen: 18.28 ng
RT: 24.15 min Scan# 3627
Delta R.T. -0.02 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.7	28.1
125	6.5	5.3	7.9




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#88
Benzo(k)fluoranthene
Concen: 7.28 ng m
RT: 24.21 min Scan# 3638
Delta R.T. -0.03 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.3	20.2	30.2
125	8.4	5.1	7.7#









#89
Benzo(a)pyrene
Concen: 14.05 ng
RT: 25.07 min Scan# 3784
Delta R.T. -0.03 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Instrument :
BNA_G
ClientSampleId :
S3-0-3FT-COMP

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.8	19.2	28.8
125	6.1	5.4	8.0




Manual Integrations
APPROVED


mohammad
1/23/2017 5:00:17 PM



#90
Dibenzo(a,h)anthracene
Concen: 2.64 ng m
RT: 29.18 min Scan# 4483
Delta R.T. -0.05 min
Lab File: BG025574.D
Acq: 21 Jan 2017 15:24

Tgt Ion	Ratio	Lower	Upper
278	100		
139	10.3	7.7	11.5
279	24.4	19.1	28.7





#91
 Benzo(a,h,i)perylene
 Concen: 9.00 ng
 RT: 30.35 min Scan# 4682
 Delta R.T. -0.08 min
 Lab File: BG025574.D
 Acq: 21 Jan 2017 15:24

Instrument :
 BNA_G
 ClientSampleId :
 S3-0-3FT-COMP

Tot Ion:	276	Resp:	375666
Ion Ratio	Lower	Upper	
276	100		
277	19.6	18.6	27.8
138	10.8	8.1	12.1

Manual Integrations
 APPROVED

mohammad
 1/23/2017 5:00:17 PM

