

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012121\
 Data File : BG047948.D
 Acq On : 21 Jan 2021 10:51
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC020

Quant Time: Jan 21 12:41:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 21 12:32:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.146	152	34723	20.00	ng	0.00
21) Naphthalene-d8	10.966	136	137370	20.00	ng	0.00
39) Acenaphthene-d10	14.767	164	102266	20.00	ng	0.00
64) Phenanthrene-d10	17.511	188	271979	20.00	ng	# 0.00
76) Chrysene-d12	21.794	240	303077	20.00	ng	# 0.00
86) Perylene-d12	25.096	264	302860	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.696	112	87352	45.76	ng	0.00
7) Phenol-d6	7.288	99	131079	44.92	ng	0.00
23) Nitrobenzene-d5	9.309	82	119710	42.76	ng	0.00
42) 2,4,6-Tribromophenol	16.248	330	65788	48.44	ng	0.00
45) 2-Fluorobiphenyl	13.398	172	313502	40.81	ng	0.00
79) Terphenyl-d14	20.114	244	700324	42.61	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.592	88	20825	20.12	ng	# 92
3) Pyridine	3.986	79	61074	20.08	ng	96
4) n-Nitrosodimethylamine	3.898	42	29084	19.02	ng	# 61
6) Aniline	7.464	93	77931	20.77	ng	94
8) 2-Chlorophenol	7.711	128	45111	23.60	ng	92
9) Benzaldehyde	7.282	77	35843	18.70	ng	# 80
10) Phenol	7.311	94	62606	22.48	ng	82
11) bis(2-Chloroethyl)ether	7.558	93	49080	20.75	ng	88
12) 1,3-Dichlorobenzene	8.040	146	58089	20.87	ng	# 90
13) 1,4-Dichlorobenzene	8.181	146	56891	20.70	ng	94
14) 1,2-Dichlorobenzene	8.181	146	56891	21.62	ng	93
15) Benzyl Alcohol	8.375	79	53867	23.71	ng	89
16) 2,2'-oxybis(1-Chloropr...	8.675	45	83333	20.20	ng	97
17) 2-Methylphenol	8.575	107	42147	23.23	ng	# 85
18) Hexachloroethane	9.239	117	20931	22.46	ng	95
19) n-Nitroso-di-n-propyla...	8.945	70	46193	21.58	ng	96
20) 3+4-Methylphenols	8.898	107	55774	22.93	ng	92
22) Acetophenone	8.968	105	79893	19.69	ng	# 88
24) Nitrobenzene	9.350	77	63026	21.75	ng	# 86
25) Isophorone	9.873	82	116490	20.72	ng	97
26) 2-Nitrophenol	10.067	139	22619	28.89	ng	# 85
27) 2,4-Dimethylphenol	10.114	122	40910	23.27	ng	96
28) bis(2-Chloroethoxy)met...	10.355	93	70210	20.43	ng	98
29) 2,4-Dichlorophenol	10.596	162	47953	23.16	ng	95
30) 1,2,4-Trichlorobenzene	10.819	180	63111	20.22	ng	94
31) Naphthalene	11.013	128	153796	20.03	ng	98
32) Benzoic acid	10.196	122	26258	26.29	ng	# 85
33) 4-Chloroaniline	11.107	127	67400	20.20	ng	# 86
34) Hexachlorobutadiene	11.301	225	46741	19.89	ng	95
35) Caprolactam	11.859	113	19287	25.19	ng	86
36) 4-Chloro-3-methylphenol	12.212	107	53990	23.08	ng	92
37) 2-Methylnaphthalene	12.611	142	116628	20.33	ng	# 92
38) 1-Methylnaphthalene	12.611	142	116628	21.62	ng	99
40) 1,2,4,5-Tetrachloroben...	12.970	216	78829	20.00	ng	98
41) Hexachlorocyclopentadiene	12.958	237	39242	23.99	ng	98
43) 2,4,6-Trichlorophenol	13.199	196	48784	25.22	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012121\
 Data File : BG047948.D
 Acq On : 21 Jan 2021 10:51
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC020


Quant Time: Jan 21 12:41:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 21 12:32:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	13.269	196	51254	24.98	ng	#	91
46) 1,1'-Biphenyl	13.610	154	154912	19.95	ng		96
47) 2-Chloronaphthalene	13.651	162	124019	19.47	ng		95
48) 2-Nitroaniline	13.839	65	41792	27.48	ng	#	79
49) Acenaphthylene	14.491	152	200558	20.64	ng		96
50) Dimethylphthalate	14.215	163	183979	24.13	ng		98
51) 2,6-Dinitrotoluene	14.327	165	34327	24.66	ng	#	75
52) Acenaphthene	14.832	154	129325	20.48	ng		95
53) 3-Nitroaniline	14.656	138	38268	23.57	ng	#	78
54) 2,4-Dinitrophenol	14.861	184	17594	24.71	ng	#	80
55) Dibenzofuran	15.167	168	198010	20.46	ng		96
56) 4-Nitrophenol	14.944	139	29322	23.53	ng	#	70
57) 2,4-Dinitrotoluene	15.114	165	48377	25.09	ng	#	76
58) Fluorene	15.813	166	168320	20.64	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.384	232	53835	26.83	ng	#	97
60) Diethylphthalate	15.578	149	180288	24.52	ng		96
61) 4-Chlorophenyl-phenyle...	15.807	204	99010	20.18	ng		94
62) 4-Nitroaniline	15.819	138	41493	22.43	ng	#	63
63) Azobenzene	16.095	77	175209	20.85	ng		87
65) 4,6-Dinitro-2-methylph...	15.878	198	23989	24.67	ng		92
66) n-Nitrosodiphenylamine	16.013	169	158055	20.27	ng		95
67) 4-Bromophenyl-phenylether	16.700	248	70360	19.89	ng		98
68) Hexachlorobenzene	16.818	284	75999	19.91	ng		100
69) Atrazine	16.959	200	61549	25.18	ng		97
70) Pentachlorophenol	17.153	266	42504	23.11	ng		94
71) Phenanthrene	17.558	178	298563	20.20	ng		95
72) Anthracene	17.646	178	297384	20.71	ng		97
73) Carbazole	17.911	167	273857	21.23	ng		100
74) Di-n-butylphthalate	18.469	149	334278	26.00	ng	#	98
75) Fluoranthene	19.556	202	418231	21.35	ng		94
77) Benzidine	19.732	184	154692	23.20	ng		99
78) Pyrene	19.914	202	418885	20.25	ng		96
80) Butylbenzylphthalate	20.802	149	145819	28.66	ng	#	85
81) Benzo(a)anthracene	21.777	228	426999	20.49	ng		97
82) 3,3'-Dichlorobenzidine	21.683	252	144608	22.85	ng		94
83) Chrysene	21.841	228	415294	20.69	ng		97
84) Bis(2-ethylhexyl)phtha...	21.689	149	208072	27.14	ng		95
85) Di-n-octyl phthalate	22.940	149	342360	28.14	ng	#	93
87) Indeno(1,2,3-cd)pyrene	28.874	276	476934	21.30	ng	#	90
88) Benzo(b)fluoranthene	24.051	252	419178	20.84	ng	#	97
89) Benzo(k)fluoranthene	24.121	252	410989	21.14	ng	#	98
90) Benzo(a)pyrene	24.950	252	392874	20.99	ng	#	97
91) Dibenzo(a,h)anthracene	28.951	278	388459	21.53	ng	#	95
92) Benzo(g,h,i)perylene	30.061	276	377279	20.75	ng	#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 21 12:41:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 21 12:32:30 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.146 min Scan# 818
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:152 Resp: 34723
Ion Ratio Lower Upper
152 100
150 144.5 125.0 187.6
115 56.5 41.8 62.6



#2
1,4-Dioxane
Concen: 20.12 ng
RT: 3.592 min Scan# 43
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 88 Resp: 20825
Ion Ratio Lower Upper
88 100
58 80.4 61.6 92.4
43 36.0 21.3 31.9#









#3
Pyridine
Concen: 20.08 ng
RT: 3.986 min Scan# 110
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion: 79	Resp: 61074
Ion Ratio	Lower Upper
79 100	
52 74.7	61.3 91.9
51 39.6	28.4 42.6



#4
n-Nitrosodimethylamine
Concen: 19.02 ng
RT: 3.898 min Scan# 95
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 42	Resp: 29084
Ion Ratio	Lower Upper
42 100	
74 97.9	120.4 180.6
44 6.8	6.2 9.2









#5
2-Fluorophenol
Concen: 45.76 ng
RT: 5.696 min Scan# 401
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 87352
Ion Ratio Lower Upper
112 100
64 76.7 46.7 70.1#
63 39.7 22.8 34.2#



#6
Aniline
Concen: 20.77 ng
RT: 7.464 min Scan# 702
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 93 Resp: 77931
Ion Ratio Lower Upper
93 100
66 41.7 29.5 44.3
65 20.4 15.5 23.3









#7
Phenol-d6
Concen: 44.92 ng
RT: 7.288 min Scan# 672
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDICC020


Tgt Ion: 99 Resp: 131079
Ion Ratio Lower Upper
99 100
42 22.1 12.8 19.2#
71 42.1 24.6 36.8#



#8
2-Chlorophenol
Concen: 23.60 ng
RT: 7.711 min Scan# 744
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 128 Resp: 45111
Ion Ratio Lower Upper
128 100
130 33.5 12.6 52.6
64 53.4 24.8 64.8









#9
Benzaldehyde
Concen: 18.70 ng
RT: 7.282 min Scan# 671
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 35843
Ion Ratio Lower Upper
77 100
105 86.8 85.9 125.9
106 79.6 80.9 120.9#



#10
Phenol
Concen: 22.48 ng
RT: 7.311 min Scan# 676
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 94 Resp: 62606
Ion Ratio Lower Upper
94 100
65 33.4 6.6 46.6
66 45.8 13.9 53.9









#11
bis(2-Chloroethyl)ether
Concen: 20.75 ng
RT: 7.558 min Scan# 718
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 49080
Ion Ratio Lower Upper
93 100
63 90.5 56.4 96.4
95 31.8 13.8 53.8



#12
1,3-Dichlorobenzene
Concen: 20.87 ng
RT: 8.040 min Scan# 800
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:146 Resp: 58089
Ion Ratio Lower Upper
146 100
148 66.0 50.1 75.1
75 38.0 20.1 30.1#









#13
1,4-Dichlorobenzene
Concen: 20.70 ng
RT: 8.181 min Scan# 824
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDICC020


Tgt Ion:146 Resp: 56891
Ion Ratio Lower Upper
146 100
148 61.6 51.1 76.7
111 43.9 29.7 44.5



#14
1,2-Dichlorobenzene
Concen: 21.62 ng
RT: 8.181 min Scan# 824
Delta R.T. -0.327 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:146 Resp: 56891
Ion Ratio Lower Upper
146 100
148 61.6 53.0 79.6
111 43.9 30.7 46.1









#15
Benzyl Alcohol
Concen: 23.71 ng
RT: 8.375 min Scan# 857
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:	79	Resp:	53867
Ion Ratio	Lower	Upper	
79	100		
108	65.9	65.5	98.3
77	63.4	50.1	75.1



#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.20 ng
RT: 8.675 min Scan# 908
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:	45	Resp:	83333
Ion Ratio	Lower	Upper	
45	100		
77	14.1	0.0	33.4
79	9.5	0.0	30.8









#17
2-Methylphenol
Concen: 23.23 ng
RT: 8.575 min Scan# 891
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	42147
Ion Ratio	Lower	Upper	
107	100		
108	98.4	87.2	130.8
77	53.7	33.8	50.8#
79	56.3	32.7	49.1#



#18
Hexachloroethane
Concen: 22.46 ng
RT: 9.239 min Scan# 1004
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:	117	Resp:	20931
Ion Ratio	Lower	Upper	
117	100		
119	96.1	77.5	116.3
201	122.5	104.9	157.3









#19

n-Nitroso-di-n-propylamine
Concen: 21.58 ng
RT: 8.945 min Scan# 954
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Instrument :
BNA_G
ClientSampleId :
SSTDIC020




Tgt Ion:	70	Resp:	46193
Ion	Ratio	Lower	Upper
70	100		
42	59.1	45.6	68.4
101	9.1	8.4	12.6
130	18.8	17.8	26.8




#20

3+4-Methylphenols
Concen: 22.93 ng
RT: 8.898 min Scan# 946
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:	107	Resp:	55774
Ion	Ratio	Lower	Upper
107	100		
108	87.9	70.6	110.6
77	38.7	10.2	50.2
79	35.2	6.8	46.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.966 min Scan# 1298
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:	136	Resp:	137370
Ion Ratio	Lower	Upper	
136	100		
137	10.9	9.1	13.7
54	8.9	6.2	9.4
68	5.1	5.0	7.4



#22
Acetophenone
Concen: 19.69 ng
RT: 8.968 min Scan# 958
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:	105	Resp:	79893
Ion Ratio	Lower	Upper	
105	100		
71	0.9	2.7	4.1#
51	35.0	21.0	31.6#
120	21.0	19.3	28.9









#23
Nitrobenzene-d5
Concen: 42.76 ng
RT: 9.309 min Scan# 1016
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 119710
Ion Ratio Lower Upper
82 100
128 38.9 37.0 55.6
54 56.5 39.0 58.4



#24
Nitrobenzene
Concen: 21.75 ng
RT: 9.350 min Scan# 1023
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51


Tgt Ion: 77 Resp: 63026
Ion Ratio Lower Upper
77 100
123 38.6 39.7 59.5#
65 15.6 10.4 15.6#








#25
Isophorone
Concen: 20.72 ng
RT: 9.873 min Scan# 1112
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Instrument :
BNA_G
ClientSampleId :
SSTDIC020







Tgt Ion: 82 Resp: 116490
Ion Ratio Lower Upper
82 100
95 8.1 5.6 8.4
138 18.2 15.8 23.6



#26
2-Nitrophenol
Concen: 28.89 ng
RT: 10.067 min Scan# 1145
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 139 Resp: 22619
Ion Ratio Lower Upper
139 100
109 30.4 20.7 31.1
65 57.6 36.1 54.1#









#27
2,4-Dimethylphenol
Concen: 23.27 ng
RT: 10.114 min Scan# 1153
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 40910
Ion Ratio Lower Upper
122 100
107 115.3 89.2 133.8
121 54.8 46.9 70.3



#28
bis(2-Chloroethoxy)methane
Concen: 20.43 ng
RT: 10.355 min Scan# 1194
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 93 Resp: 70210
Ion Ratio Lower Upper
93 100
95 32.1 26.5 39.7
123 11.0 9.4 14.0









#29
2,4-Dichlorophenol
Concen: 23.16 ng
RT: 10.596 min Scan# 1235
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 47953
Ion Ratio Lower Upper
162 100
164 67.6 45.6 85.6
98 36.0 11.0 51.0



#30
1,2,4-Trichlorobenzene
Concen: 20.22 ng
RT: 10.819 min Scan# 1273
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:180 Resp: 63111
Ion Ratio Lower Upper
180 100
182 89.7 76.9 115.3
145 24.7 21.4 32.0









#31
Naphthalene
Concen: 20.03 ng
RT: 11.013 min Scan# 1306
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampled :
SSTDIC020


Tgt Ion:128 Resp: 153796
Ion Ratio Lower Upper
128 100
129 10.1 9.0 13.6
127 13.1 11.0 16.4



#32
Benzoic acid
Concen: 26.29 ng
RT: 10.196 min Scan# 1167
Delta R.T. -0.033 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:122 Resp: 26258
Ion Ratio Lower Upper
122 100
105 130.2 103.7 143.7
77 107.9 63.0 103.0#









#33
4-Chloroaniline
Concen: 20.20 ng
RT: 11.107 min Scan# 1322
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:	127	Resp:	67400
Ion Ratio	Lower	Upper	
127	100		
129	27.9	26.8	40.2
65	40.0	23.8	35.6#
92	23.4	14.5	21.7#



#34
Hexachlorobutadiene
Concen: 19.89 ng
RT: 11.301 min Scan# 1355
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:	225	Resp:	46741
Ion Ratio	Lower	Upper	
225	100		
223	66.4	49.5	74.3
227	62.3	52.0	78.0









#35
Caprolactam
Concen: 25.19 ng
RT: 11.859 min Scan# 1450
Delta R.T. -0.016 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:113 Resp: 19287
Ion Ratio Lower Upper
113 100
55 188.9 149.2 189.2
56 142.9 107.2 147.2



#36
4-Chloro-3-methylphenol
Concen: 23.08 ng
RT: 12.212 min Scan# 1510
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:107 Resp: 53990
Ion Ratio Lower Upper
107 100
144 27.8 22.5 33.7
142 79.2 70.8 106.2









#37
2-Methylnaphthalene
Concen: 20.33 ng
RT: 12.611 min Scan# 1578
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 116628
Ion Ratio Lower Upper
142 100
141 93.7 71.6 107.4
115 42.6 25.4 38.2#



#38
1-Methylnaphthalene
Concen: 21.62 ng
RT: 12.611 min Scan# 1578
Delta R.T. -0.215 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:142 Resp: 116628
Ion Ratio Lower Upper
142 100
141 93.7 75.4 113.2
116 3.9 2.9 4.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.767 min Scan# 1945
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:164 Resp: 102266
Ion Ratio Lower Upper
164 100
162 107.4 82.3 123.5
160 42.2 35.3 52.9



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.00 ng
RT: 12.970 min Scan# 1639
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:216 Resp: 78829
Ion Ratio Lower Upper
216 100
214 78.1 61.9 92.9
179 17.6 14.8 22.2
108 14.8 10.2 15.4









#41
Hexachlorocyclopentadiene
Concen: 23.99 ng
RT: 12.958 min Scan# 1637
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:237 Resp: 39242
Ion Ratio Lower Upper
237 100
235 64.2 43.1 83.1
272 15.2 0.0 33.1



#42
2,4,6-Tribromophenol
Concen: 48.44 ng
RT: 16.248 min Scan# 2197
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:330 Resp: 65788
Ion Ratio Lower Upper
330 100
332 96.0 76.5 114.7
141 31.9 26.0 39.0









#43
2,4,6-Trichlorophenol
Concen: 25.22 ng
RT: 13.199 min Scan# 1678
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 48784
Ion Ratio Lower Upper
196 100
198 89.8 74.9 112.3
200 31.6 25.3 37.9



#44
2,4,5-Trichlorophenol
Concen: 24.98 ng
RT: 13.269 min Scan# 1690
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:196 Resp: 51254
Ion Ratio Lower Upper
196 100
198 91.7 77.6 116.4
97 48.0 28.6 43.0#
132 19.5 17.8 26.8









#45
2-Fluorobiphenyl
Concen: 40.81 ng
RT: 13.398 min Scan# 1712
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:172	Resp:	313502
Ion Ratio	Lower	Upper
172	100	
171	36.4	31.3 46.9
170	23.3	20.8 31.2



#46
1,1'-Biphenyl
Concen: 19.95 ng
RT: 13.610 min Scan# 1748
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:154	Resp:	154912
Ion Ratio	Lower	Upper
154	100	
153	40.6	22.6 62.6
76	9.9	0.0 32.6









#47
2-Chloronaphthalene
Concen: 19.47 ng
RT: 13.651 min Scan# 1755
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 124019
Ion Ratio Lower Upper
162 100
127 38.4 27.8 41.6
164 35.5 27.0 40.4



#48
2-Nitroaniline
Concen: 27.48 ng
RT: 13.839 min Scan# 1787
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion: 65 Resp: 41792
Ion Ratio Lower Upper
65 100
92 63.1 58.6 88.0
138 93.6 98.2 147.4#





#49

Acenaphthylene

Concen: 20.64 ng

RT: 14.491 min Scan# 1898

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51

Instrument :

BNA_G

ClientSampleId :

SSTDIC020





Tgt Ion:152 Resp: 200558

Ion Ratio Lower Upper

152 100

151 20.3 17.4 26.0

153 12.4 11.5 17.3



#50

Dimethylphthalate

Concen: 24.13 ng

RT: 14.215 min Scan# 1851

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51




Tgt Ion:163 Resp: 183979


Ion Ratio Lower Upper

163 100

194 4.1 3.8 5.8

164 10.4 8.8 13.2









#51
2,6-Dinitrotoluene
Concen: 24.66 ng
RT: 14.327 min Scan# 1870
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 34327
Ion Ratio Lower Upper
165 100
63 58.4 33.7 50.5#
89 62.0 36.6 54.8#



#52
Acenaphthene
Concen: 20.48 ng
RT: 14.832 min Scan# 1956
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:154 Resp: 129325
Ion Ratio Lower Upper
154 100
153 118.8 90.3 135.5
152 52.7 43.8 65.8









#53
3-Nitroaniline
Concen: 23.57 ng
RT: 14.656 min Scan# 1926
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 38268
Ion Ratio Lower Upper
138 100
108 11.5 7.4 11.0#
92 130.4 85.2 127.8#



#54
2,4-Dinitrophenol
Concen: 24.71 ng
RT: 14.861 min Scan# 1961
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:184 Resp: 17594
Ion Ratio Lower Upper
184 100
63 52.8 32.3 48.5#
154 64.7 40.4 60.6#









#55
Dibenzofuran
Concen: 20.46 ng
RT: 15.167 min Scan# 2013
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 198010
Ion Ratio Lower Upper
168 100
139 39.2 28.6 43.0
169 13.6 11.0 16.6



#56
4-Nitrophenol
Concen: 23.53 ng
RT: 14.944 min Scan# 1975
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:139 Resp: 29322
Ion Ratio Lower Upper
139 100
109 88.1 45.9 85.9#
65 114.2 65.4 105.4#









#57
2,4-Dinitrotoluene
Concen: 25.09 ng
RT: 15.114 min Scan# 2004
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:165	Resp:	48377
Ion Ratio	Lower	Upper
165	100	
63	47.0	26.4 39.6#
89	78.4	47.8 71.8#
182	2.2	2.6 4.0#



#58
Fluorene
Concen: 20.64 ng
RT: 15.813 min Scan# 2123
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:166	Resp:	168320
Ion Ratio	Lower	Upper
166	100	
165	95.3	77.2 115.8
167	14.4	11.4 17.2









#59
2,3,4,6-Tetrachlorophenol
Concen: 26.83 ng
RT: 15.384 min Scan# 2050
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:232	Resp:	53835
Ion Ratio	Lower	Upper
232	100	
131	35.6	26.6 40.0
130	2.0	1.4 2.0#
166	25.9	21.0 31.6



#60
Diethylphthalate
Concen: 24.52 ng
RT: 15.578 min Scan# 2083
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:149	Resp:	180288
Ion Ratio	Lower	Upper
149	100	
177	23.9	21.4 32.2
150	12.6	10.7 16.1









#61
4-Chlorophenyl-phenylether
Concen: 20.18 ng
RT: 15.807 min Scan# 2122
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 99010
Ion Ratio Lower Upper
204 100
206 35.5 27.8 41.8
141 58.0 41.5 62.3



#62
4-Nitroaniline
Concen: 22.43 ng
RT: 15.819 min Scan# 2124
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:138 Resp: 41493
Ion Ratio Lower Upper
138 100
92 61.9 26.9 66.9
108 107.7 49.4 89.4#









#63
Azobenzene
Concen: 20.85 ng
RT: 16.095 min Scan# 2171
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion: 77	Resp: 175209
Ion Ratio	Lower Upper
77 100	
182 25.1	14.0 54.0
105 17.5	3.0 43.0
51 33.9	8.1 48.1



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.511 min Scan# 2412
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:188	Resp: 271979
Ion Ratio	Lower Upper
188 100	
94 6.2	6.3 9.5#
80 7.7	7.0 10.6









#65
4,6-Dinitro-2-methylphenol
Concen: 24.67 ng
RT: 15.878 min Scan# 2134
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:198 Resp: 23989
Ion Ratio Lower Upper
198 100
51 38.8 10.4 50.4
105 32.2 12.9 52.9



#66
n-Nitrosodiphenylamine
Concen: 20.27 ng
RT: 16.013 min Scan# 2157
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:169 Resp: 158055
Ion Ratio Lower Upper
169 100
168 64.0 55.5 83.3
167 36.7 30.5 45.7









#67
4-Bromophenyl-phenylether
Concen: 19.89 ng
RT: 16.700 min Scan# 2274
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 70360
Ion Ratio Lower Upper
248 100
250 97.9 78.9 118.3
141 61.2 46.2 69.4



#68
Hexachlorobenzene
Concen: 19.91 ng
RT: 16.818 min Scan# 2294
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:284 Resp: 75999
Ion Ratio Lower Upper
284 100
142 26.8 21.1 31.7
249 32.5 25.9 38.9





#69

Atrazine

Concen: 25.18 ng

RT: 16.959 min Scan# 2318

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51

Tgt Ion:200 Resp: 61549

Ion Ratio Lower Upper

200 100

173 23.6 4.2 44.2





215 53.0 30.0 70.0

Instrument :

BNA_G

ClientSampleId :

SSTDIC020



#70

Pentachlorophenol

Concen: 23.11 ng

RT: 17.153 min Scan# 2351

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51




Tgt Ion:266 Resp: 42504


Ion Ratio Lower Upper

266 100

268 66.8 49.8 74.6

264 66.8 49.3 73.9









#71
Phenanthrene
Concen: 20.20 ng
RT: 17.558 min Scan# 2420
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 298563
Ion Ratio Lower Upper
178 100
176 18.8 16.7 25.1
179 15.5 14.0 21.0



#72
Anthracene
Concen: 20.71 ng
RT: 17.646 min Scan# 2435
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:178 Resp: 297384
Ion Ratio Lower Upper
178 100
176 19.3 16.4 24.6
179 15.5 13.8 20.6









#73
 Carbazole
 Concen: 21.23 ng
 RT: 17.911 min Scan# 2480
 Delta R.T. 0.002 min
 Lab File: BG047948.D
 Acq: 21 Jan 2021 10:51




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC020


Tgt Ion:167 Resp: 273857
 Ion Ratio Lower Upper
 167 100
 166 23.6 18.7 28.1
 139 11.8 9.6 14.4



#74
 Di-n-butylphthalate
 Concen: 26.00 ng
 RT: 18.469 min Scan# 2575
 Delta R.T. -0.004 min
 Lab File: BG047948.D
 Acq: 21 Jan 2021 10:51

Tgt Ion:149 Resp: 334278
 Ion Ratio Lower Upper
 149 100
 150 9.6 8.2 12.4
 104 6.1 3.8 5.8#









#75
Fluoranthene
Concen: 21.35 ng
RT: 19.556 min Scan# 2760
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:202 Resp: 418231
Ion Ratio Lower Upper
202 100
101 6.4 0.0 29.7
203 17.1 0.0 39.2



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.794 min Scan# 3141
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:240 Resp: 303077
Ion Ratio Lower Upper
240 100
120 4.9 6.2 9.4#
236 25.2 20.5 30.7









#77
Benzidine
Concen: 23.20 ng
RT: 19.732 min Scan# 2790
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:184 Resp: 154692
Ion Ratio Lower Upper
184 100
185 16.0 12.4 18.6
183 13.0 10.6 15.8



#78
Pyrene
Concen: 20.25 ng
RT: 19.914 min Scan# 2821
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:202 Resp: 418885
Ion Ratio Lower Upper
202 100
200 20.8 18.6 27.8
203 18.2 15.3 22.9





#79

Terphenyl-d14

Concen: 42.61 ng

RT: 20.114 min Scan# 2855

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51





Tgt Ion:244 Resp: 700324

Ion Ratio Lower Upper

244 100

212 7.0 6.7 10.1

122 5.8 7.8 11.6#



#80

Butylbenzylphthalate

Concen: 28.66 ng

RT: 20.802 min Scan# 2972

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51




Tgt Ion:149 Resp: 145819


Ion Ratio Lower Upper

149 100

91 79.9 51.5 77.3#

206 25.4 21.2 31.8









#81
Benzo(a)anthracene
Concen: 20.49 ng
RT: 21.777 min Scan# 3138
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 426999
Ion Ratio Lower Upper
228 100
226 27.6 23.3 34.9
229 19.9 17.4 26.0



#82
3,3'-Dichlorobenzidine
Concen: 22.85 ng
RT: 21.683 min Scan# 3122
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:252 Resp: 144608
Ion Ratio Lower Upper
252 100
254 61.5 53.2 79.8
126 7.0 6.8 10.2





#83

Chrysene

Concen: 20.69 ng

RT: 21.841 min Scan# 3149

Delta R.T. -0.004 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51





Instrument :

BNA_G

ClientSampleId :

SSTDIC020

Tgt Ion:228	Resp: 415294
Ion Ratio	Lower Upper
228 100	
226 30.0	25.0 37.4
229 19.8	17.3 25.9



#84

Bis(2-ethylhexyl)phthalate

Concen: 27.14 ng




RT: 21.689 min Scan# 3123


Delta R.T. 0.002 min

Lab File: BG047948.D

Acq: 21 Jan 2021 10:51

Tgt Ion:149	Resp: 208072
Ion Ratio	Lower Upper
149 100	
167 27.6	24.8 37.2
279 5.6	4.6 7.0









#85
Di-n-octyl phthalate
Concen: 28.14 ng
RT: 22.940 min Scan# 3336
Delta R.T. 0.002 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 342360
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.2
43 11.4 6.8 10.2#



#86
Perylene-d12
Concen: 20.00 ng
RT: 25.096 min Scan# 3703
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:264 Resp: 302860
Ion Ratio Lower Upper
264 100
260 20.8 18.8 28.2
265 20.0 17.8 26.8









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 21.30 ng
 RT: 28.874 min Scan# 4346
 Delta R.T. -0.021 min
 Lab File: BG047948.D
 Acq: 21 Jan 2021 10:51




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC020


Tgt Ion:276 Resp: 476934
 Ion Ratio Lower Upper
 276 100
 138 10.0 16.2 24.2#
 277 25.3 20.3 30.5



#88
 Benzo(b)fluoranthene
 Concen: 20.84 ng
 RT: 24.051 min Scan# 3525
 Delta R.T. -0.010 min
 Lab File: BG047948.D
 Acq: 21 Jan 2021 10:51

Tgt Ion:252 Resp: 419178
 Ion Ratio Lower Upper
 252 100
 253 22.8 18.7 28.1
 125 5.3 6.5 9.7#









#89
Benzo(k)fluoranthene
Concen: 21.14 ng
RT: 24.121 min Scan# 3537
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51




Instrument :
BNA_G
ClientSampleId :
SSTDIC020


Tgt Ion:252 Resp: 410989
Ion Ratio Lower Upper
252 100
253 22.9 18.6 27.8
125 6.0 6.8 10.2#



#90
Benzo(a)pyrene
Concen: 20.99 ng
RT: 24.950 min Scan# 3678
Delta R.T. -0.004 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:252 Resp: 392874
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 6.1 7.0 10.4#









#91
Dibenzo(a,h)anthracene
Concen: 21.53 ng
RT: 28.951 min Scan# 4359
Delta R.T. -0.016 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Instrument :
BNA_G
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 388459
Ion Ratio Lower Upper
278 100
139 7.9 10.5 15.7#
279 24.5 19.4 29.0



#92
Benzo(g,h,i)perylene
Concen: 20.75 ng
RT: 30.061 min Scan# 4548
Delta R.T. -0.010 min
Lab File: BG047948.D
Acq: 21 Jan 2021 10:51

Tgt Ion:276 Resp: 377279
Ion Ratio Lower Upper
276 100
277 24.0 19.6 29.4
138 11.6 13.4 20.2#

