


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG012218\
 Data File : BG032204.D
 Acq On : 22 Jan 2018 9:46
 Operator : SJ/JU
 Sample : DFTPP74
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP74

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG012218.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 29 06:15:33 2018



AutoFind: Scans 2584, 2585, 2586; Background Corrected with Scan 2578

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	19.8	19591	PASS
68	69	0.00	2	1.4	299	PASS
69	198	0.00	100	21.7	21445	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	31.8	31456	PASS
197	198	0.00	2	0.3	260	PASS
198	198	100	100	100.0	98904	PASS
199	198	5	9	6.7	6613	PASS
275	198	10	60	32.1	31760	PASS
365	198	1	100	3.7	3694	PASS
441	442	0.01	100	15.4	11422	PASS
442	198	50	100	75.1	74266	PASS
443	442	15	24	20.0	14871	PASS

m/z	Abundance
38.00	155.0
39.00	958.0
40.00	182.0
41.00	321.0
43.00	212.0
44.00	288.0
44.90	158.0
50.00	4593.0
51.00	16191.0
52.10	698.0
56.00	776.0
57.00	1036.0
61.00	195.0
63.00	783.0
63.90	174.0
69.00	17288.0
73.00	453.0
74.00	1847.0
75.00	2954.0
76.10	1104.0
77.10	20056.0
78.00	1484.0
79.00	1621.0
80.00	1584.0
81.00	1575.0
81.90	493.0
83.10	264.0
83.80	159.0
84.90	284.0
85.90	388.0
87.00	205.0
90.90	515.0
91.90	453.0
93.00	2931.0
96.00	150.0
98.00	2200.0
99.00	1857.0
100.10	297.0
101.00	1034.0
103.00	299.0
104.00	666.0
105.00	685.0
105.80	166.0
107.00	8010.0
108.00	1458.0
110.00	13085.0
111.00	2186.0
112.00	220.0
115.10	169.0
116.00	552.0
117.00	6323.0
118.00	484.0
120.00	173.0
122.00	566.0
123.00	800.0
124.00	438.0
125.10	358.0
127.00	28000.0
128.10	2196.0
129.00	12190.0
130.00	1306.0
131.80	172.0
133.00	169.0
134.00	526.0
135.00	954.0
136.10	613.0
136.90	414.0
137.90	294.0
140.10	253.0
141.00	1666.0
142.10	417.0
143.00	401.0
146.00	348.0
147.00	1111.0
148.00	1782.0
149.00	643.0
150.00	204.0
151.00	336.0
152.10	190.0
152.90	724.0
154.00	398.0
155.00	1030.0
156.00	1470.0
156.80	311.0
158.00	179.0
158.90	224.0
160.00	651.0
161.00	613.0
162.00	407.0
164.90	1161.0
166.00	638.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

167.00	4151.0
168.10	1777.0
168.90	538.0
170.10	180.0
171.10	188.0
171.90	315.0
173.10	716.0
174.00	622.0
175.10	1702.0
176.00	346.0
177.00	754.0
178.00	370.0
179.00	2341.0
180.00	2305.0
181.10	869.0
182.20	221.0
183.10	224.0
184.10	341.0
185.00	1492.0
186.10	11243.0
187.00	3236.0
188.00	486.0
189.00	726.0
190.00	178.0
191.00	452.0
192.10	945.0
193.00	1590.0
194.10	219.0
195.10	263.0
196.10	1906.0
198.00	91400.0
199.00	5958.0
199.90	549.0
201.80	340.0
203.00	974.0
204.00	3962.0
205.00	5348.0
206.10	25456.0
207.10	3562.0
208.10	1168.0
209.00	472.0
210.00	446.0
210.90	903.0
215.10	388.0
216.00	529.0
217.00	7117.0
218.00	657.0
221.00	3319.0
221.70	728.0
222.00	789.0
223.10	1638.0
224.10	14972.0
225.10	4422.0
226.10	555.0
227.00	6239.0
228.00	1047.0
229.00	985.0
230.00	244.0
231.00	554.0
232.10	223.0
234.10	400.0
235.00	356.0
236.00	415.0
236.80	375.0
237.20	413.0
239.10	197.0
240.10	175.0
241.00	418.0
242.10	781.0
243.10	928.0
244.10	11199.0
245.10	1725.0
246.00	2311.0
247.20	171.0
248.10	242.0
249.10	453.0
251.00	175.0
252.20	169.0
253.10	320.0
255.00	68864.0
256.00	9630.0
257.00	854.0
258.00	3956.0
259.00	568.0
265.00	1665.0
266.10	160.0
271.00	235.0
272.10	277.0
273.00	2456.0
274.00	5542.0
275.10	30952.0
276.10	3983.0
277.00	2957.0
278.00	523.0
282.00	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

283.00	389.0
283.40	209.0
284.20	267.0
285.00	772.0
289.00	390.0
293.10	578.0
296.00	9635.0
297.10	1179.0
298.10	166.0
301.00	229.0
303.10	1014.0
303.90	313.0
308.20	161.0
314.00	526.0
315.00	955.0
316.00	600.0
321.20	340.0
323.10	2616.0
324.10	462.0
327.00	615.0
332.00	269.0
333.20	317.0
334.00	1925.0
335.10	746.0
341.10	390.0
346.10	563.0
347.10	184.0
351.10	177.0
352.10	765.0
353.00	514.0
354.10	1112.0
355.10	168.0
358.10	150.0
365.10	3944.0
366.20	364.0
371.10	181.0
372.00	1491.0
373.10	307.0
383.10	327.0
402.10	587.0
403.00	539.0
404.00	242.0
420.90	559.0
423.10	4518.0
424.10	875.0
438.40	456.0
441.10	12362.0
442.10	87472.0
443.10	16896.0
444.20	1540.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

m/z	Abundance
38.00	263.0
39.10	1859.0
40.00	163.0
43.90	581.0
45.00	202.0
49.00	276.0
50.00	6366.0
51.00	23072.0
52.00	1115.0
55.00	153.0
56.00	755.0
57.00	1216.0
60.90	253.0
62.00	456.0
63.00	1172.0
65.00	379.0
68.00	899.0
69.00	25280.0
73.10	302.0
74.00	2942.0
75.00	4355.0
77.00	25984.0
78.00	1974.0
79.00	2297.0
80.00	1293.0
81.00	2455.0
82.00	678.0
83.10	267.0
85.00	726.0
86.00	690.0
87.10	231.0
87.90	361.0
91.00	728.0
92.00	741.0
93.00	3949.0
93.70	171.0
94.00	171.0
96.20	279.0
98.00	3331.0
99.00	2172.0
100.00	232.0
101.00	1421.0
102.90	396.0
104.00	904.0
105.00	808.0
106.00	334.0
107.00	11639.0
108.00	1786.0
108.90	386.0
110.00	18856.0
111.00	2798.0
114.30	186.0
115.10	205.0
116.00	938.0
117.00	9599.0
117.90	641.0
120.00	250.0
122.00	719.0
123.00	1071.0
124.00	333.0
125.00	913.0
127.00	36904.0
128.00	3227.0
129.00	15271.0
130.10	1043.0
131.00	164.0
132.10	254.0
133.90	417.0
135.00	1056.0
136.00	440.0
137.00	722.0
138.10	241.0
139.00	225.0
140.00	263.0
141.00	2155.0
142.10	757.0
143.10	490.0
145.90	703.0
147.00	1125.0
148.00	2317.0
149.10	490.0
150.00	210.0
150.90	288.0
151.30	192.0
153.00	948.0
154.00	675.0
155.00	1383.0
156.10	2123.0
157.10	465.0
158.00	623.0
159.10	494.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

160.10	843.0
161.00	952.0
162.00	388.0
165.00	1006.0
166.10	998.0
167.10	4324.0
168.00	2190.0
168.90	495.0
170.00	295.0
171.10	165.0
172.00	456.0
173.00	618.0
174.00	1108.0
175.00	2299.0
176.10	648.0
177.00	1136.0
178.00	266.0
179.00	3445.0
180.00	2452.0
181.00	1269.0
181.90	165.0
183.90	347.0
185.10	1794.0
186.10	16704.0
187.00	5135.0
188.00	281.0
189.00	876.0
190.00	248.0
191.00	662.0
192.00	1655.0
193.00	1410.0
194.00	270.0
195.00	210.0
196.00	3353.0
198.00	114552.0
199.00	7904.0
200.00	771.0
201.30	470.0
202.10	259.0
203.10	1013.0
204.00	4531.0
205.10	6627.0
206.10	30888.0
207.10	4431.0
207.90	1061.0
208.90	318.0
210.10	333.0
211.00	913.0
211.80	319.0
212.80	174.0
215.00	407.0
216.00	647.0
217.00	7942.0
218.00	1028.0
221.00	4096.0
222.00	964.0
223.00	2004.0
224.10	20360.0
225.10	4462.0
226.00	527.0
227.00	7695.0
228.00	1101.0
229.00	1355.0
230.00	178.0
231.00	489.0
231.90	229.0
233.20	315.0
234.00	548.0
235.10	322.0
235.90	374.0
237.00	702.0
239.00	326.0
240.10	413.0
241.00	374.0
242.00	952.0
243.10	1402.0
244.10	13471.0
245.10	1812.0
246.00	3297.0
247.00	733.0
249.00	710.0
251.00	305.0
251.90	323.0
252.50	365.0
253.00	298.0
255.00	82064.0
256.00	11324.0
257.00	1101.0
258.00	4106.0
259.10	687.0
261.00	237.0
264.00	198.0
265.00	2258.0
266.00	478.0
271.20	237.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

273.00	3148.0
274.00	5848.0
275.10	37224.0
276.00	5833.0
277.00	3527.0
278.00	652.0
283.00	406.0
284.40	152.0
285.00	460.0
286.00	154.0
289.00	197.0
291.20	200.0
292.20	360.0
293.10	464.0
293.90	312.0
296.00	9891.0
297.10	1287.0
301.00	173.0
302.10	205.0
303.10	1222.0
304.00	348.0
308.00	154.0
308.90	242.0
314.00	383.0
315.00	1130.0
316.10	645.0
321.10	494.0
323.10	3508.0
324.00	733.0
327.00	628.0
328.10	411.0
332.10	151.0
333.00	501.0
334.10	2522.0
335.10	628.0
341.10	518.0
346.10	733.0
352.10	1205.0
353.10	491.0
354.10	898.0
358.90	170.0
364.10	212.0
365.10	3999.0
366.10	660.0
371.20	295.0
372.10	1451.0
373.00	166.0
383.00	202.0
384.10	261.0
391.00	153.0
402.00	475.0
403.10	740.0
404.10	257.0
421.00	491.0
422.20	921.0
423.10	3810.0
424.10	961.0
429.60	151.0
437.20	241.0
438.20	231.0
438.90	350.0
441.10	13549.0
442.10	82640.0
443.10	16154.0
444.10	1577.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

m/z	Abundance
38.10	202.0
39.10	1672.0
41.00	170.0
43.90	571.0
49.00	316.0
50.10	5163.0
51.00	19512.0
52.00	1280.0
54.90	250.0
56.00	746.0
57.00	1335.0
61.00	205.0
62.10	234.0
63.00	824.0
64.90	330.0
69.00	21768.0
72.90	212.0
74.10	2442.0
75.00	3865.0
76.00	1764.0
77.00	25400.0
78.00	1530.0
79.00	1814.0
80.00	1432.0
81.00	2147.0
82.00	621.0
83.10	335.0
85.10	644.0
85.90	652.0
86.90	621.0
91.00	512.0
92.10	567.0
93.00	2909.0
94.00	411.0
98.00	2844.0
99.00	1871.0
100.00	226.0
101.00	1112.0
102.90	463.0
104.00	828.0
104.90	789.0
107.00	8402.0
108.10	1277.0
109.00	278.0
110.00	15182.0
111.00	2656.0
112.00	351.0
113.20	179.0
117.00	7436.0
118.00	523.0
122.00	663.0
123.00	1052.0
124.00	377.0
124.90	509.0
127.00	29464.0
128.00	2501.0
129.00	12648.0
129.90	805.0
131.10	294.0
132.50	158.0
134.00	396.0
135.00	852.0
136.10	205.0
137.00	499.0
137.50	274.0
140.20	171.0
140.90	1779.0
142.00	521.0
143.00	334.0
143.80	183.0
145.00	168.0
146.00	262.0
147.00	1242.0
148.00	2054.0
149.00	255.0
152.00	185.0
153.00	981.0
154.00	593.0
155.10	1431.0
156.00	1707.0
157.10	316.0
158.00	302.0
159.00	351.0
160.00	625.0
161.00	944.0
162.00	430.0
164.90	616.0
166.00	689.0
167.00	4058.0
168.00	1635.0
168.90	321.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

170.90	216.0
172.00	456.0
173.00	301.0
174.00	1004.0
175.10	1864.0
176.10	599.0
177.00	836.0
179.00	3047.0
180.00	2335.0
181.00	804.0
184.00	487.0
185.00	1408.0
186.00	12000.0
187.00	2937.0
188.10	389.0
189.00	579.0
191.00	478.0
192.00	1153.0
193.00	1121.0
194.00	438.0
195.00	189.0
196.00	2449.0
196.70	782.0
198.00	90760.0
199.00	5977.0
200.00	413.0
201.60	601.0
203.00	559.0
204.00	3853.0
205.10	5621.0
206.00	24152.0
207.00	3553.0
208.00	1116.0
208.90	432.0
210.00	455.0
211.10	1137.0
212.10	323.0
215.10	356.0
216.10	540.0
217.00	6272.0
218.00	831.0
221.10	3978.0
222.00	783.0
223.00	1264.0
224.10	13235.0
225.00	3134.0
226.00	463.0
227.00	6480.0
228.00	841.0
229.00	1142.0
229.80	253.0
231.00	514.0
232.90	175.0
234.00	351.0
235.10	229.0
236.00	364.0
237.10	562.0
237.90	168.0
240.00	419.0
241.00	375.0
242.00	797.0
243.20	869.0
244.10	10810.0
245.00	1913.0
246.00	2692.0
247.10	685.0
249.00	300.0
250.90	216.0
251.80	206.0
252.50	181.0
253.10	350.0
255.00	60912.0
256.00	8669.0
257.00	687.0
258.00	3055.0
258.90	614.0
260.90	196.0
265.10	1656.0
266.00	217.0
271.10	327.0
271.90	192.0
273.00	1715.0
274.00	4236.0
275.10	27104.0
276.00	3433.0
277.10	1907.0
278.10	489.0
283.00	398.0
284.00	316.0
285.10	426.0
289.10	169.0
293.00	589.0
294.00	286.0
296.10	7396.0
297.00	977.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

301.80	184.0
303.10	945.0
304.10	395.0
308.00	173.0
309.10	175.0
314.00	518.0
315.00	898.0
316.10	368.0
322.10	157.0
323.10	2784.0
324.10	310.0
327.00	452.0
333.00	346.0
334.10	1290.0
335.20	181.0
341.20	203.0
342.00	221.0
346.00	504.0
351.20	161.0
352.00	572.0
353.10	592.0
354.10	706.0
355.10	297.0
359.10	176.0
365.10	3140.0
366.00	298.0
371.10	212.0
372.10	983.0
373.00	314.0
402.10	213.0
403.00	460.0
403.90	156.0
421.10	442.0
422.30	426.0
423.10	2954.0
424.20	542.0
438.50	239.0
439.10	166.0
441.10	8357.0
442.10	52688.0
443.10	11565.0
444.10	1220.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

Scan 2578 (18.437 min): BG032204.D
DFTPP74

m/z	Abundance
44.00	403.0
45.00	191.0
57.00	224.0
73.10	161.0
207.10	359.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74