

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012224\
 Data File : BG060379.D
 Acq On : 22 Jan 2024 10:54
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 22 11:29:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	124	0.00
2	1,4-Dioxane	40.000	48.905	-22.3	133	0.00
3	Pyridine	40.000	44.260	-10.6	115	0.00
4	n-Nitrosodimethylamine	40.000	39.144	2.1	107	0.00
5 S	2-Fluorophenol	80.000	78.541	1.8	130	0.00
6	Aniline	40.000	41.379	-3.4	119	0.00
7 S	Phenol-d6	80.000	83.281	-4.1	111	0.00
8	2-Chlorophenol	40.000	41.427	-3.6	125	0.00
9	Benzaldehyde	40.000	37.141	7.1	106	0.00
10 C	Phenol	40.000	41.858	-4.6	111	0.00
11	bis(2-Chloroethyl)ether	40.000	40.270	-0.7	123	0.00
12	1,3-Dichlorobenzene	40.000	38.925	2.7	120	0.00
13 C	1,4-Dichlorobenzene	40.000	39.212	2.0	123	0.00
14	1,2-Dichlorobenzene	40.000	38.296	4.3	107	0.00
15	Benzyl Alcohol	40.000	42.139	-5.3	121	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.620	3.5	107	0.00
17	2-Methylphenol	40.000	40.523	-1.3	108	0.00
18	Hexachloroethane	40.000	39.260	1.9	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.415	-1.0	104	0.00
20	3+4-Methylphenols	40.000	41.963	-4.9	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	41.303	-3.3	111	0.00
23 S	Nitrobenzene-d5	80.000	84.250	-5.3	114	0.00
24	Nitrobenzene	40.000	39.606	1.0	106	0.00
25	Isophorone	40.000	39.539	1.2	106	0.00
26 C	2-Nitrophenol	40.000	45.171	-12.9	120	0.00
27	2,4-Dimethylphenol	40.000	41.891	-4.7	111	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.016	2.5	106	0.00
29 C	2,4-Dichlorophenol	40.000	42.066	-5.2	112	0.00
30	1,2,4-Trichlorobenzene	40.000	38.754	3.1	105	0.00
31	Naphthalene	40.000	39.433	1.4	109	0.00
32	Benzoic acid	40.000	44.013	-10.0	130	0.01
33	4-Chloroaniline	40.000	40.879	-2.2	113	0.00
34 C	Hexachlorobutadiene	40.000	38.739	3.2	113	0.00
35	Caprolactam	40.000	44.941	-12.4	123	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.909	-7.3	116	0.00
37	2-Methylnaphthalene	40.000	39.828	0.4	112	0.00
38	1-Methylnaphthalene	40.000	39.597	1.0	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.867	7.8	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.199	4.5	105	0.00
42 S	2,4,6-Tribromophenol	80.000	74.074	7.4	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.007	-0.0	115	0.00
44	2,4,5-Trichlorophenol	40.000	40.050	-0.1	115	0.00
45 S	2-Fluorobiphenyl	80.000	72.940	8.8	111	0.00
46	1,1'-Biphenyl	40.000	37.636	5.9	111	0.00
47	2-Chloronaphthalene	40.000	37.690	5.8	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012224\
 Data File : BG060379.D
 Acq On : 22 Jan 2024 10:54
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 22 11:29:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.837	-7.1	119	0.00
49	Acenaphthylene	40.000	39.104	2.2	112	0.00
50	Dimethylphthalate	40.000	38.956	2.6	113	0.00
51	2,6-Dinitrotoluene	40.000	43.410	-8.5	119	0.00
52 C	Acenaphthene	40.000	39.108	2.2	113	0.00
53	3-Nitroaniline	40.000	43.807	-9.5	122	0.00
54 P	2,4-Dinitrophenol	40.000	43.338	-8.3	133	0.00
55	Dibenzofuran	40.000	38.384	4.0	111	0.00
56 P	4-Nitrophenol	40.000	47.516	-18.8	123	0.00
57	2,4-Dinitrotoluene	40.000	44.055	-10.1	122	0.00
58	Fluorene	40.000	34.858	12.9	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.416	4.0	109	0.00
60	Diethylphthalate	40.000	35.761	10.6	102	0.00
61	4-Chlorophenyl-phenylether	40.000	34.380	14.0	102	0.00
62	4-Nitroaniline	40.000	38.242	4.4	106	0.00
63	Azobenzene	40.000	34.967	12.6	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.716	-11.8	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.412	-1.0	104	0.00
67	4-Bromophenyl-phenylether	40.000	40.108	-0.3	104	0.00
68	Hexachlorobenzene	40.000	38.858	2.9	102	0.00
69	Atrazine	40.000	37.245	6.9	97	0.00
70 C	Pentachlorophenol	40.000	46.731	-16.8	112	0.00
71	Phenanthrene	40.000	39.106	2.2	102	0.00
72	Anthracene	40.000	39.007	2.5	101	0.00
73	Carbazole	40.000	39.397	1.5	101	0.00
74	Di-n-butylphthalate	40.000	40.134	-0.3	102	0.00
75 C	Fluoranthene	40.000	36.796	8.0	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzidine	40.000	43.479	-8.7	94	0.00
78	Pyrene	40.000	41.537	-3.8	96	0.00
79 S	Terphenyl-d14	80.000	88.892	-11.1	98	0.00
80	Butylbenzylphthalate	40.000	48.162	-20.4	105	0.00
81	Benzo(a)anthracene	40.000	40.107	-0.3	91	0.00
82	3,3'-Dichlorobenzidine	40.000	40.672	-1.7	88	0.00
83	Chrysene	40.000	39.570	1.1	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.136	-12.8	98	0.00
85 c	Di-n-octyl phthalate	40.000	46.209	-15.5	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.028	-0.1	90	0.02
88	Benzo(b)fluoranthene	40.000	40.158	-0.4	89	0.00
89	Benzo(k)fluoranthene	40.000	39.296	1.8	87	0.00
90 C	Benzo(a)pyrene	40.000	39.846	0.4	89	0.00
91	Dibenzo(a,h)anthracene	40.000	39.910	0.2	88	0.01
92	Benzo(g,h,i)perylene	40.000	40.342	-0.9	91	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012224\
Data File : BG060379.D
Acq On : 22 Jan 2024 10:54
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 22 11:29:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0