

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012318\
 Data File : BG032225.D
 Acq On : 23 Jan 2018 10:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 09:39:20 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 10:13:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.420	0.341	18.8	64	0.00
3	Pyridine	1.121	1.005	10.3	68	0.00
4	n-Nitrosodimethylamine	0.394	0.472	-19.8	89	0.00
5 S	2-Fluorophenol	1.094	1.036	5.3	72	0.00
6	Aniline	1.737	1.676	3.5	73	0.00
7 S	Phenol-d6	1.377	1.407	-2.2	77	0.00
8	2-Chlorophenol	1.224	1.193	2.5	73	0.00
9	Benzaldehyde	0.798	0.746	6.5	69	0.00
10 C	Phenol	1.357	1.346	0.8	74	0.00
11	bis(2-Chloroethyl)ether	1.087	1.004	7.6	70	0.00
12	1,3-Dichlorobenzene	1.602	1.560	2.6	74	0.00
13 C	1,4-Dichlorobenzene	1.613	1.576	2.3	75	0.00
14	1,2-Dichlorobenzene	1.560	1.524	2.3	76	0.00
15	Benzyl Alcohol	1.001	1.113	-11.2	82	0.00
16	2,2'-oxybis(1-Chloropropane	1.105	1.018	7.9	71	0.00
17	2-Methylphenol	1.037	1.047	-1.0	75	0.00
18	Hexachloroethane	0.496	0.505	-1.8	77	0.00
19 P	n-Nitroso-di-n-propylamine	0.855	0.884	-3.4	78	0.00
20	3+4-Methylphenols	1.436	1.450	-1.0	77	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
22	Acetophenone	0.454	0.478	-5.3	79	0.01
23 S	Nitrobenzene-d5	0.296	0.325	-9.8	81	0.00
24	Nitrobenzene	0.295	0.330	-11.9	83	0.00
25	Isophorone	0.550	0.556	-1.1	75	0.00
26 C	2-Nitrophenol	0.175	0.198	-13.1	81	0.00
27	2,4-Dimethylphenol	0.277	0.281	-1.4	76	0.01
28	bis(2-Chloroethoxy)methane	0.332	0.332	0.0	74	0.01
29 C	2,4-Dichlorophenol	0.341	0.384	-12.6	83	0.00
30	1,2,4-Trichlorobenzene	0.441	0.482	-9.3	81	0.00
31	Naphthalene	0.991	1.016	-2.5	76	0.00
32	Benzoic acid	0.204	0.189	7.4	65	0.01
33	4-Chloroaniline	0.425	0.434	-2.1	76	0.00
34 C	Hexachlorobutadiene	0.316	0.377	-19.3	90	0.00
35	Caprolactam	0.104	0.103	1.0	72	0.01
36 C	4-Chloro-3-methylphenol	0.312	0.345	-10.6	83	0.00
37	2-Methylnaphthalene	0.787	0.829	-5.3	80	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
39	1,2,4,5-Tetrachlorobenzene	0.872	0.980	-12.4	88	0.00
40 P	Hexachlorocyclopentadiene	0.491	0.506	-3.1	79	0.00
41 S	2,4,6-Tribromophenol	0.331	0.342	-3.3	78	0.00
42 C	2,4,6-Trichlorophenol	0.490	0.536	-9.4	84	0.00
43	2,4,5-Trichlorophenol	0.567	0.621	-9.5	86	0.00
44 S	2-Fluorobiphenyl	1.613	1.705	-5.7	84	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012318\
 Data File : BG032225.D
 Acq On : 23 Jan 2018 10:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 09:39:20 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 10:13:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.715	1.746	-1.8	81	0.00
46	2-Chloronaphthalene	1.334	1.361	-2.0	80	0.00
47	2-Nitroaniline	0.255	0.295	-15.7	86	0.00
48	Acenaphthylene	2.031	2.057	-1.3	79	0.00
49	Dimethylphthalate	1.750	1.777	-1.5	79	0.00
50	2,6-Dinitrotoluene	0.363	0.389	-7.2	81	0.00
51 C	Acenaphthene	1.343	1.385	-3.1	80	0.00
52	3-Nitroaniline	0.328	0.332	-1.2	77	0.00
53 P	2,4-Dinitrophenol	0.153	0.170	-11.1	83	0.00
54	Dibenzofuran	2.004	2.043	-1.9	80	0.00
55 P	4-Nitrophenol	0.239	0.236	1.3	73	0.00
56	2,4-Dinitrotoluene	0.489	0.533	-9.0	80	0.00
57	Fluorene	1.643	1.648	-0.3	79	0.00
58	2,3,4,6-Tetrachlorophenol	0.525	0.553	-5.3	79	0.00
59	Diethylphthalate	1.697	1.684	0.8	77	0.00
60	4-Chlorophenyl-phenylether	1.008	1.070	-6.2	83	0.00
61	4-Nitroaniline	0.315	0.334	-6.0	77	0.00
62	Azobenzene	1.062	1.063	-0.1	77	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.141	-18.5	89	0.00
65 c	n-Nitrosodiphenylamine	0.607	0.616	-1.5	76	0.00
66	4-Bromophenyl-phenylether	0.285	0.305	-7.0	82	0.01
67	Hexachlorobenzene	0.315	0.328	-4.1	80	0.00
68	Atrazine	0.255	0.261	-2.4	77	0.00
69 C	Pentachlorophenol	0.201	0.200	0.5	74	0.00
70	Phenanthrene	1.114	1.118	-0.4	78	0.00
71	Anthracene	1.111	1.135	-2.2	78	0.00
72	Carbazole	0.963	0.984	-2.2	78	0.00
73	Di-n-butylphthalate	1.138	1.102	3.2	74	0.00
74 C	Fluoranthene	1.394	1.477	-6.0	82	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	85	0.01
76	Benzidine	0.500	0.325	35.0#	54	0.00
77	Pyrene	1.257	1.216	3.3	82	0.00
78 S	Terphenyl-d14	0.906	0.901	0.6	86	0.00
79	Butylbenzylphthalate	0.450	0.418	7.1	78	0.01
80	Benzo(a)anthracene	1.244	1.300	-4.5	89	0.00
81	3,3'-Dichlorobenzidine	0.465	0.486	-4.5	86	0.01
82	Chrysene	1.195	1.241	-3.8	87	0.00
83	Bis(2-ethylhexyl)phthalate	0.661	0.605	8.5	77	0.00
84 c	Di-n-octyl phthalate	1.049	0.975	7.1	78	0.02
85	Indeno(1,2,3-cd)pyrene	1.462	1.574	-7.7	90	0.02
86 I	Perylene-d12	1.000	1.000	0.0	87	0.02
87	Benzo(b)fluoranthene	1.209	1.208	0.1	84	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012318\
Data File : BG032225.D
Acq On : 23 Jan 2018 10:57
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 24 09:39:20 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG011218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 18 10:13:16 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.181	1.242	-5.2	90	0.01
89 C	Benzo(a)pyrene	1.144	1.197	-4.6	89	0.02
90	Dibenzo(a,h)anthracene	1.102	1.139	-3.4	88	0.03
91	Benzo(a,h,i)perylene	1.126	1.109	1.5	85	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0