

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012517\
 Data File : BG025644.D
 Acq On : 25 Jan 2017 21:38
 Operator : SJ/MA
 Sample : PB96324BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK24

Quant Time: Jan 26 12:47:00 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG011817.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 18:08:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	69041	20.00	ng/ul	0.00
18) Naphthalene-d8	11.00	136	296444	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.80	164	262959	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.55	188	614056	20.00	ng/ul	0.00
75) Chrysene-d12	21.85	240	749024	20.00	ng/ul	0.00
83) Perylene-d12	25.23	264	765553	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.50	96	8647	6.27	ng/uL	0.00
5) Phenol-d5	7.34	99	208774	35.28	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	148561	38.87	ng/ul	0.00
9) 2-Chlorophenol-d4	7.71	132	144403	34.60	ng/ul	0.00
13) 4-Methylphenol-d8	8.89	113	177182	35.23	ng/ul	0.00
19) Nitrobenzene-d5	9.36	128	84401	40.17	ng/ul	0.00
22) 2-Nitrophenol-d4	10.09	143	101077	40.31	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.63	165	178009	35.99	ng/ul	0.00
29) 4-Chloroaniline-d4	11.15	131	251376	45.46	ng/ul	0.00
43) Dimethylphthalate-d6	14.20	166	748820	40.33	ng/ul	0.00
46) Acenaphthylene-d8	14.50	160	845940	42.17	ng/ul	0.00
51) 4-Nitrophenol-d4	15.04	143	94695	37.22	ng/ul	0.00
57) Fluorene-d10	15.80	176	682792	39.01	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.94	200	132203	34.90	ng/ul	0.00
70) Anthracene-d10	17.65	188	1101426	42.13	ng/ul	0.00
76) Pyrene-d10	19.93	212	1333853	42.80	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.00	264	1350115	41.96	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG012517\
Data File : BG025644.D
Acq On : 25 Jan 2017 21:38
Operator : SJ/MA
Sample : PB96324BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK24

Quant Time: Jan 26 12:47:00 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG011817.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 25 18:08:33 2017
Response via : Initial Calibration

