


Data Path : Z:\HPCHEM1\BNA G\DATA\BG012518\
 Data File : BG032313.D
 Acq On : 26 Jan 2018 3:46
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG011218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 24 14:02:41 2018



AutoFind: Scans 2584, 2585, 2586; Background Corrected with Scan 2577

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	19.7	11203	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	21.2	12054	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	31.3	17817	PASS
197	198	0.00	2	0.3	174	PASS
198	198	100	100	100.0	56901	PASS
199	198	5	9	6.4	3639	PASS
275	198	10	60	33.4	18978	PASS
365	198	1	100	4.2	2386	PASS
441	442	0.01	24	15.7	7443	PASS
442	198	50	100	83.5	47488	PASS
443	442	15	24	18.3	8713	PASS