

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
 Data File : BG047997.D
 Acq On : 25 Jan 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012521

Quant Time: Jan 25 22:05:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.580	0.554	4.5	100	0.00
3	Pyridine	1.741	1.722	1.1	93	0.00
4	n-Nitrosodimethylamine	0.805	0.796	1.1	94	0.00
5 S	2-Fluorophenol	1.301	1.274	2.1	96	0.00
6	Aniline	2.390	2.443	-2.2	98	0.00
7 S	Phenol-d6	1.979	2.017	-1.9	98	0.00
8	2-Chlorophenol	1.362	1.363	-0.1	96	0.00
9	Benzaldehyde	1.015	1.048	-3.3	98	0.00
10 C	Phenol	1.897	1.936	-2.1	98	0.00
11	bis(2-Chloroethyl)ether	1.485	1.477	0.5	95	0.00
12	1,3-Dichlorobenzene	1.644	1.611	2.0	96	0.00
13 C	1,4-Dichlorobenzene	1.648	1.653	-0.3	97	0.00
14	1,2-Dichlorobenzene	1.573	1.572	0.1	96	0.00
15	Benzyl Alcohol	1.674	1.689	-0.9	95	0.00
16	2,2'-oxybis(1-Chloropropane	2.085	2.071	0.7	96	0.00
17	2-Methylphenol	1.282	1.274	0.6	97	0.00
18	Hexachloroethane	0.643	0.610	5.1	90	0.00
19 P	n-Nitroso-di-n-propylamine	1.444	1.441	0.2	96	0.00
20	3+4-Methylphenols	1.773	1.785	-0.7	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.616	0.625	-1.5	97	0.00
23 S	Nitrobenzene-d5	0.505	0.523	-3.6	99	0.00
24	Nitrobenzene	0.506	0.522	-3.2	98	0.00
25	Isophorone	0.907	0.926	-2.1	98	0.00
26 C	2-Nitrophenol	0.208	0.218	-4.8	101	0.00
27	2,4-Dimethylphenol	0.303	0.302	0.3	97	0.00
28	bis(2-Chloroethoxy)methane	0.546	0.552	-1.1	98	0.00
29 C	2,4-Dichlorophenol	0.389	0.417	-7.2	102	0.00
30	1,2,4-Trichlorobenzene	0.471	0.480	-1.9	96	0.00
31	Naphthalene	1.156	1.177	-1.8	97	0.00
32	Benzoic acid	0.271	0.299	-10.3	99	0.00
33	4-Chloroaniline	0.529	0.549	-3.8	99	0.00
34 C	Hexachlorobutadiene	0.334	0.343	-2.7	98	0.00
35	Caprolactam	0.148	0.155	-4.7	100	0.00
36 C	4-Chloro-3-methylphenol	0.437	0.451	-3.2	99	0.00
37	2-Methylnaphthalene	0.873	0.902	-3.3	99	0.00
38	1-Methylnaphthalene	0.833	0.854	-2.5	97	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.762	0.777	-2.0	102	0.00
41 P	Hexachlorocyclopentadiene	0.432	0.440	-1.9	99	0.00
42 S	2,4,6-Tribromophenol	0.333	0.357	-7.2	105	0.00
43 C	2,4,6-Trichlorophenol	0.532	0.549	-3.2	102	0.00
44	2,4,5-Trichlorophenol	0.551	0.575	-4.4	101	0.00
45 S	2-Fluorobiphenyl	1.547	1.559	-0.8	100	0.00
46	1,1'-Biphenyl	1.546	1.576	-1.9	101	0.00
47	2-Chloronaphthalene	1.346	1.347	-0.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
 Data File : BG047997.D
 Acq On : 25 Jan 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012521

Quant Time: Jan 25 22:05:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.488	0.516	-5.7	101	0.00
49	Acenaphthylene	1.986	1.993	-0.4	100	0.00
50	Dimethylphthalate	1.810	1.861	-2.8	102	0.00
51	2,6-Dinitrotoluene	0.376	0.384	-2.1	100	0.00
52 C	Acenaphthene	1.344	1.390	-3.4	102	0.00
53	3-Nitroaniline	0.413	0.435	-5.3	103	0.00
54 P	2,4-Dinitrophenol	0.235	0.277	-17.9	110	0.00
55	Dibenzofuran	2.039	2.082	-2.1	101	0.00
56 P	4-Nitrophenol	0.322	0.340	-5.6	102	0.00
57	2,4-Dinitrotoluene	0.543	0.575	-5.9	102	0.00
58	Fluorene	1.639	1.659	-1.2	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.551	0.588	-6.7	106	0.00
60	Diethylphthalate	1.861	1.913	-2.8	102	0.00
61	4-Chlorophenyl-phenylether	1.004	1.036	-3.2	102	0.00
62	4-Nitroaniline	0.450	0.470	-4.4	102	0.00
63	Azobenzene	1.752	1.752	0.0	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.139	-7.8	103	0.00
66 c	n-Nitrosodiphenylamine	0.601	0.617	-2.7	102	0.00
67	4-Bromophenyl-phenylether	0.267	0.283	-6.0	105	0.00
68	Hexachlorobenzene	0.291	0.299	-2.7	103	0.00
69	Atrazine	0.228	0.238	-4.4	102	0.00
70 C	Pentachlorophenol	0.176	0.196	-11.4	106	0.00
71	Phenanthrene	1.154	1.199	-3.9	103	0.00
72	Anthracene	1.135	1.165	-2.6	102	0.00
73	Carbazole	1.059	1.081	-2.1	101	0.00
74	Di-n-butylphthalate	1.293	1.316	-1.8	101	0.00
75 C	Fluoranthene	1.513	1.535	-1.5	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzydine	0.577	0.585	-1.4	93	0.00
78	Pyrene	1.465	1.511	-3.1	102	0.00
79 S	Terphenyl-d14	1.154	1.183	-2.5	102	0.00
80	Butylbenzylphthalate	0.559	0.578	-3.4	102	0.00
81	Benzo(a)anthracene	1.454	1.478	-1.7	102	0.00
82	3,3'-Dichlorobenzidine	0.490	0.517	-5.5	104	0.00
83	Chrysene	1.381	1.417	-2.6	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.766	0.780	-1.8	102	0.00
85 c	Di-n-octyl phthalate	1.282	1.322	-3.1	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.609	1.649	-2.5	102	-0.02
88	Benzo(b)fluoranthene	1.421	1.467	-3.2	103	0.00
89	Benzo(k)fluoranthene	1.381	1.414	-2.4	102	0.00
90 C	Benzo(a)pyrene	1.335	1.372	-2.8	102	0.00
91	Dibenzo(a,h)anthracene	1.328	1.352	-1.8	101	0.00
92	Benzo(g,h,i)perylene	1.279	1.319	-3.1	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
Data File : BG047997.D
Acq On : 25 Jan 2021 20:17
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG012521

Quant Time: Jan 25 22:05:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0