

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
 Data File : BG047997.D
 Acq On : 25 Jan 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012521

Quant Time: Jan 25 22:05:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	38.192	4.5	100	0.00
3	Pyridine	40.000	39.550	1.1	93	0.00
4	n-Nitrosodimethylamine	40.000	39.557	1.1	94	0.00
5 S	2-Fluorophenol	80.000	78.332	2.1	96	0.00
6	Aniline	40.000	40.887	-2.2	98	0.00
7 S	Phenol-d6	80.000	81.539	-1.9	98	0.00
8	2-Chlorophenol	40.000	40.020	-0.1	96	0.00
9	Benzaldehyde	40.000	41.312	-3.3	98	0.00
10 C	Phenol	40.000	40.841	-2.1	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.776	0.6	95	0.00
12	1,3-Dichlorobenzene	40.000	39.195	2.0	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.118	-0.3	97	0.00
14	1,2-Dichlorobenzene	40.000	39.956	0.1	96	0.00
15	Benzyl Alcohol	40.000	40.366	-0.9	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.742	0.6	96	0.00
17	2-Methylphenol	40.000	39.751	0.6	97	0.00
18	Hexachloroethane	40.000	37.947	5.1	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.905	0.2	96	0.00
20	3+4-Methylphenols	40.000	40.256	-0.6	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	40.578	-1.4	97	0.00
23 S	Nitrobenzene-d5	80.000	82.701	-3.4	99	0.00
24	Nitrobenzene	40.000	41.277	-3.2	98	0.00
25	Isophorone	40.000	40.826	-2.1	98	0.00
26 C	2-Nitrophenol	40.000	41.769	-4.4	101	0.00
27	2,4-Dimethylphenol	40.000	39.948	0.1	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.491	-1.2	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.803	-7.0	102	0.00
30	1,2,4-Trichlorobenzene	40.000	40.812	-2.0	96	0.00
31	Naphthalene	40.000	40.700	-1.8	97	0.00
32	Benzoic acid	40.000	44.127	-10.3	99	0.00
33	4-Chloroaniline	40.000	41.573	-3.9	99	0.00
34 C	Hexachlorobutadiene	40.000	40.975	-2.4	98	0.00
35	Caprolactam	40.000	41.950	-4.9	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.282	-3.2	99	0.00
37	2-Methylnaphthalene	40.000	41.289	-3.2	99	0.00
38	1-Methylnaphthalene	40.000	40.991	-2.5	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.770	-1.9	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.725	-1.8	99	0.00
42 S	2,4,6-Tribromophenol	80.000	85.779	-7.2	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.305	-3.3	102	0.00
44	2,4,5-Trichlorophenol	40.000	41.702	-4.3	101	0.00
45 S	2-Fluorobiphenyl	80.000	80.641	-0.8	100	0.00
46	1,1'-Biphenyl	40.000	40.762	-1.9	101	0.00
47	2-Chloronaphthalene	40.000	40.052	-0.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
 Data File : BG047997.D
 Acq On : 25 Jan 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG012521

Quant Time: Jan 25 22:05:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 25 19:39:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.237	-5.6	101	0.00
49	Acenaphthylene	40.000	40.148	-0.4	100	0.00
50	Dimethylphthalate	40.000	41.114	-2.8	102	0.00
51	2,6-Dinitrotoluene	40.000	40.847	-2.1	100	0.00
52 C	Acenaphthene	40.000	41.343	-3.4	102	0.00
53	3-Nitroaniline	40.000	42.152	-5.4	103	0.00
54 P	2,4-Dinitrophenol	40.000	47.077	-17.7	110	0.00
55	Dibenzofuran	40.000	40.837	-2.1	101	0.00
56 P	4-Nitrophenol	40.000	42.201	-5.5	102	0.00
57	2,4-Dinitrotoluene	40.000	42.291	-5.7	102	0.00
58	Fluorene	40.000	40.485	-1.2	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.695	-6.7	106	0.00
60	Diethylphthalate	40.000	41.121	-2.8	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.294	-3.2	102	0.00
62	4-Nitroaniline	40.000	41.814	-4.5	102	0.00
63	Azobenzene	40.000	40.008	-0.0	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.114	-7.8	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.057	-2.6	102	0.00
67	4-Bromophenyl-phenylether	40.000	42.409	-6.0	105	0.00
68	Hexachlorobenzene	40.000	41.101	-2.8	103	0.00
69	Atrazine	40.000	41.766	-4.4	102	0.00
70 C	Pentachlorophenol	40.000	44.456	-11.1	106	0.00
71	Phenanthrene	40.000	41.578	-3.9	103	0.00
72	Anthracene	40.000	41.057	-2.6	102	0.00
73	Carbazole	40.000	40.837	-2.1	101	0.00
74	Di-n-butylphthalate	40.000	40.715	-1.8	101	0.00
75 C	Fluoranthene	40.000	40.590	-1.5	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	40.573	-1.4	93	0.00
78	Pyrene	40.000	41.261	-3.2	102	0.00
79 S	Terphenyl-d14	80.000	82.065	-2.6	102	0.00
80	Butylbenzylphthalate	40.000	41.353	-3.4	102	0.00
81	Benzo(a)anthracene	40.000	40.668	-1.7	102	0.00
82	3,3'-Dichlorobenzidine	40.000	42.186	-5.5	104	0.00
83	Chrysene	40.000	41.036	-2.6	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.736	-1.8	102	0.00
85 c	Di-n-octyl phthalate	40.000	41.269	-3.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.998	-2.5	102	-0.02
88	Benzo(b)fluoranthene	40.000	41.291	-3.2	103	0.00
89	Benzo(k)fluoranthene	40.000	40.953	-2.4	102	0.00
90 C	Benzo(a)pyrene	40.000	41.111	-2.8	102	0.00
91	Dibenzo(a,h)anthracene	40.000	40.728	-1.8	101	0.00
92	Benzo(g,h,i)perylene	40.000	41.243	-3.1	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012521\
Data File : BG047997.D
Acq On : 25 Jan 2021 20:17
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG012521

Quant Time: Jan 25 22:05:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 25 19:39:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0