

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012524\
 Data File : BG060424.D
 Acq On : 25 Jan 2024 12:20
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 25 12:55:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	50.759	-26.9#	105	0.00
3	Pyridine	40.000	49.716	-24.3	98	0.00
4	n-Nitrosodimethylamine	40.000	43.453	-8.6	90	0.00
5 S	2-Fluorophenol	80.000	88.749	-10.9	111	0.00
6	Aniline	40.000	40.674	-1.7	89	0.00
7 S	Phenol-d6	80.000	81.840	-2.3	83	0.00
8	2-Chlorophenol	40.000	39.033	2.4	90	0.00
9	Benzaldehyde	40.000	34.517	13.7	75	0.00
10 C	Phenol	40.000	40.678	-1.7	82	0.00
11	bis(2-Chloroethyl)ether	40.000	38.465	3.8	89	0.00
12	1,3-Dichlorobenzene	40.000	39.849	0.4	93	0.00
13 C	1,4-Dichlorobenzene	40.000	39.509	1.2	94	0.00
14	1,2-Dichlorobenzene	40.000	38.556	3.6	82	0.00
15	Benzyl Alcohol	40.000	44.357	-10.9	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.220	-0.5	85	0.00
17	2-Methylphenol	40.000	41.563	-3.9	84	0.00
18	Hexachloroethane	40.000	39.235	1.9	83	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.588	-4.0	82	0.00
20	3+4-Methylphenols	40.000	43.076	-7.7	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22	Acetophenone	40.000	39.624	0.9	87	0.00
23 S	Nitrobenzene-d5	80.000	81.861	-2.3	90	0.00
24	Nitrobenzene	40.000	38.640	3.4	84	0.00
25	Isophorone	40.000	38.533	3.7	85	0.00
26 C	2-Nitrophenol	40.000	45.916	-14.8	99	0.00
27	2,4-Dimethylphenol	40.000	41.981	-5.0	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.123	2.2	87	0.00
29 C	2,4-Dichlorophenol	40.000	42.867	-7.2	93	0.00
30	1,2,4-Trichlorobenzene	40.000	39.228	1.9	87	0.00
31	Naphthalene	40.000	39.707	0.7	89	0.00
32	Benzoic acid	40.000	46.030	-15.1	113	0.00
33	4-Chloroaniline	40.000	41.790	-4.5	94	0.00
34 C	Hexachlorobutadiene	40.000	39.638	0.9	94	0.00
35	Caprolactam	40.000	43.360	-8.4	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.850	-7.1	95	0.00
37	2-Methylnaphthalene	40.000	40.031	-0.1	92	0.00
38	1-Methylnaphthalene	40.000	39.902	0.2	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.789	5.5	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.488	1.3	90	0.00
42 S	2,4,6-Tribromophenol	80.000	83.706	-4.6	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.417	-1.0	96	0.00
44	2,4,5-Trichlorophenol	40.000	41.551	-3.9	98	0.00
45 S	2-Fluorobiphenyl	80.000	76.542	4.3	96	0.00
46	1,1'-Biphenyl	40.000	38.326	4.2	94	0.00
47	2-Chloronaphthalene	40.000	38.364	4.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012524\
 Data File : BG060424.D
 Acq On : 25 Jan 2024 12:20
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 25 12:55:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.284	-5.7	97	0.00
49	Acenaphthylene	40.000	39.537	1.2	94	0.00
50	Dimethylphthalate	40.000	38.917	2.7	93	0.00
51	2,6-Dinitrotoluene	40.000	43.187	-8.0	98	0.00
52 C	Acenaphthene	40.000	39.670	0.8	95	0.00
53	3-Nitroaniline	40.000	43.268	-8.2	99	0.00
54 P	2,4-Dinitrophenol	40.000	43.961	-9.9	112	0.00
55	Dibenzofuran	40.000	39.218	2.0	94	0.00
56 P	4-Nitrophenol	40.000	44.759	-11.9	95	0.00
57	2,4-Dinitrotoluene	40.000	44.353	-10.9	102	0.00
58	Fluorene	40.000	38.577	3.6	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.538	-6.3	100	0.00
60	Diethylphthalate	40.000	39.482	1.3	94	0.00
61	4-Chlorophenyl-phenylether	40.000	38.743	3.1	95	0.00
62	4-Nitroaniline	40.000	41.674	-4.2	96	0.00
63	Azobenzene	40.000	37.439	6.4	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.009	-15.0	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.282	-0.7	95	0.00
67	4-Bromophenyl-phenylether	40.000	41.175	-2.9	97	0.00
68	Hexachlorobenzene	40.000	40.633	-1.6	97	0.00
69	Atrazine	40.000	36.371	9.1	87	0.00
70 C	Pentachlorophenol	40.000	47.329	-18.3	104	0.00
71	Phenanthrene	40.000	38.988	2.5	93	0.00
72	Anthracene	40.000	39.255	1.9	93	0.00
73	Carbazole	40.000	38.791	3.0	91	0.00
74	Di-n-butylphthalate	40.000	39.973	0.1	93	0.00
75 C	Fluoranthene	40.000	37.784	5.5	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzydine	40.000	35.092	12.3	75	0.00
78	Pyrene	40.000	39.750	0.6	91	0.00
79 S	Terphenyl-d14	80.000	85.966	-7.5	95	0.00
80	Butylbenzylphthalate	40.000	44.642	-11.6	96	0.00
81	Benzo(a)anthracene	40.000	40.160	-0.4	90	0.00
82	3,3'-Dichlorobenzidine	40.000	40.478	-1.2	87	0.00
83	Chrysene	40.000	39.437	1.4	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.778	-6.9	91	0.00
85 c	Di-n-octyl phthalate	40.000	44.366	-10.9	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.192	-0.5	89	0.01
88	Benzo(b)fluoranthene	40.000	39.592	1.0	86	0.00
89	Benzo(k)fluoranthene	40.000	39.699	0.8	86	0.00
90 C	Benzo(a)pyrene	40.000	40.270	-0.7	89	0.00
91	Dibenzo(a,h)anthracene	40.000	40.115	-0.3	87	0.01
92	Benzo(g,h,i)perylene	40.000	39.907	0.2	88	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012524\
Data File : BG060424.D
Acq On : 25 Jan 2024 12:20
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 25 12:55:38 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0