

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
 Data File : BG015934.D
 Acq On : 27 Jan 2015 6:51
 Operator : TP/IZ
 Sample : G1139-10
 Misc : S
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 121B2C

Quant Time: Jan 27 09:50:31 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 20 17:47:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	96494	20.00	ng	0.00
21) Naphthalene-d8	10.47	136	387160	20.00	ng	-0.01
38) Acenaphthene-d10	14.33	164	309812	20.00	ng	0.00
63) Phenanthrene-d10	17.08	188	872672	20.00	ng	0.00
75) Chrysene-d12	21.27	240	1239220	20.00	ng	0.00
86) Perylene-d12	23.52	264	1155998	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	272812	42.10	ng	0.00
7) Phenol-d6	6.86	99	293315	28.09	ng	0.00
23) Nitrobenzene-d5	8.84	82	895304	66.66	ng	0.00
41) 2,4,6-Tribromophenol	15.82	330	490071	86.41	ng	0.00
44) 2-Fluorobiphenyl	12.95	172	1593934	80.72	ng	0.00
78) Terphenyl-d14	19.71	244	3109511	65.98	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.52	128	646974	32.65	ng	97
37) 2-Methylnaphthalene	12.13	142	593313	37.70	ng	88
45) 1,1'-Biphenyl	13.16	154	74781	3.82	ng	91
51) Acenaphthene	14.39	154	39859	2.53	ng	91
57) Fluorene	15.38	166	79299	3.38	ng	# 93
70) Phenanthrene	17.12	178	173594	3.96	ng	93
72) Carbazole	17.49	167	121549	3.21	ng	96
83) Bis(2-ethylhexyl)phthalate	21.18	149	179429	6.18	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG012615\
Data File : BG015934.D
Acq On : 27 Jan 2015 6:51
Operator : TP/IZ
Sample : G1139-10
Misc : S
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
121B2C

Quant Time: Jan 27 09:50:31 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG012015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 20 17:47:40 2015
Response via : Initial Calibration

